



IL-4 High Sensitivity ELISA Kit (Human) (OKDB00045)

Instruction for Use

For the quantitative measurement of IL-4 in sera, plasmas, buffered solutions or cell culture media.

This product is intended for research use only.

Table of Contents

1. Background.....	2
2. Assay Summary	Error! Bookmark not defined.
3. Storage and Stability	3
4. Kit Components.....	3
5. Precautions	3
6. Required Materials Not Supplied	4
7. Technical Application Tips.....	4
8. Reagent Preparation	5
9. Sample Preparation.....	5
10. Assay Procedure.....	9
11. Calculation of Results.....	11
12. Typical Expected Data	11
13. Technical Resources.....	13

1. Background

Principle

Aviva Systems Biology IL-4 High Sensitivity ELISA Kit (Human) (OKDB00045) is designed for the highly sensitive in-vitro quantitative measurement of recombinant or natural IL-4 in human sera, plasmas, buffered solutions or cell culture media. The kit is based on standard sandwich enzyme-linked immuno-sorbent assay methods along with an additional detection process employing a polymerization step to amplify the number of detector bound HRP molecules. A monoclonal antibody specific for IL-4 has been pre-coated into wells of a 96-wellplate. Standards or test samples are added to the wells and incubated. After washing, a biotinylated detector antibody specific for IL-4 is added, incubated and followed by washing. Streptavidin-HRP conjugate is then added, incubated and unbound conjugate is washed away. Amplification is achieved by adding a Biotin-Tyramine reagent, incubating and washing, followed by a second incubation with Streptavidin-HRP conjugate. An enzymatic reaction is then produced through the addition of TMB substrate which is catalyzed by HRP to generating a blue color product that changes yellow after adding acidic stop solution. The density of yellow coloration read by absorbance at 450 nm and is quantitatively proportional to the amount of sample IL-4 captured in well.

Target Background

This gene encodes a soluble cytokine that is a member of the type II interferon class. The encoded protein is secreted by cells of both the innate and adaptive immune systems. The active protein is a homodimer that binds to the interferon gamma receptor which triggers a cellular response to viral and microbial infections. Mutations in this gene are associated with an increased susceptibility to viral, bacterial and parasitic infections and to several autoimmune diseases.

General Specifications

General Specifications	
Range	0.31 – 10 pg/mL
LOD	<0.31 pg/mL (Derived by linear regression of OD ₄₅₀ of the Mean Blank + 2xSD)
Specificity	Human IL-4 UniProt ID: P05112 GeneID: 3565 Target Alias: BSF1, IL-4, BCGF1, BSF-1, BCGF-1
Cross-Reactivity	Ten specificities were tested with concentrations higher than IL-4 curve concentrations. No cross reaction was observed for IL-2, IL-5, IL-6, IL-7, IL-8, IL-12, IL-13, TNF- α , IFN γ and FasL.

2. Storage and Stability

- Upon receipt store kit at 4°C until expiration date on packaging.

3. Kit Components

- The following reagents are the provided contents of the kit.

Reagents	Quantity			Handling Requirements
	1 x 48 Well Kit	1 x 96 Well Kit	2 x 96 Well Kit	
IL-4 Well Strip Microplate	6 x Wells	12 x 8 Wells	24 x 8 Wells	Ready to use (Pre-coated)
IL-4 Standard	1 Vial	2 Vials	4 Vials	Reconstitute as directed in Step 9.4
10X Standard Diluent (Buffer)	1 x 25 mL	1 x 25 mL	1 x 25 mL	Dilute as directed in Step 9.2
27.5X Biotinylated anti-IL-4 Detector Antibody	1 x 400 µL	1 x 400 µL	2 x 400 µL	Dilute as directed in Step 9.3
Biotinylated Detector Antibody Diluent	1 x 7 mL	1 x 7 mL	1 x 13 mL	Ready to use
Streptavidin-HRP	1 x 5 µL	2 x 5 µL	4 x 5 µL	Prepare as directed in 9.5
Amplification Diluent	1 x 25 mL	1 x 25 mL	1 x 25 mL	Ready to use
100X Amplifier	1 x 200 µL	1 x 200 µL	2 x 200 µL	Dilute as directed in 9.6
HRP Diluent	1 x 25 mL	1 x 25 mL	2 x 25 mL	Ready to use
200X Wash Buffer	1 x 10 mL	1 x 10 mL	2 x 10 mL	Dilute as directed in Step 9.1
TMB Substrate	1 x 11 mL	1 x 11 mL	2 x 24 mL	Ready to use
H ₂ SO ₄ Stop Reagent	1 x 11 mL	1 x 11 mL	2 x 11 mL	Ready to use

4. Precautions

- Read instructions fully prior to beginning use of the assay kit.
- Any deviations or modifications from the described method or use of other reagents could result in a reduction of performance.
- Reduce exposure to potentially harmful substances by wearing personal protective lab equipment including lab coats, gloves and glasses.
- For information on hazardous substances included in the kit please refer to the Material Safety Data Sheet (MSDS).
- Kit cannot be used beyond the expiration date on the label.

5. Required Materials Not Supplied

- Microplate reader capable of reading absorbance at 450 nm.
- Automated plate washer (optional).
- Pipettes capable of precisely dispensing 0.5 μ L through 1 mL volumes of aqueous solutions.
- Pipettes or volumetric glassware capable of precisely measuring 1 mL through 100 mL of aqueous solutions.
- New, clean tubes and/or micro-centrifuge tubes for the preparation of standards or samples.
- Absorbent paper or paper toweling.
- Distilled or deionized ultrapure water.
- 37°C Incubator (optional)

6. Technical Application Tips

- Do not mix or substitute components from other kits.
- To ensure the validity of experimental operation, it is recommended that pilot experiments using standards and a small selection of sample dilutions to ensure optimal dilution range for quantitation.
- Samples exhibiting OD measurements higher than the highest standard should be diluted further in the appropriate sample dilution buffers.
- Prior to using the kit, briefly spin component tubes to collect all reagents at the bottom.
- Replicate wells are recommended for standards and samples.
- Cover microplate while incubating to prevent evaporation.
- Do not allow the microplate wells dry at any point during the assay procedure.
- Do not reuse tips or tube to prevent cross contamination.
- Avoid causing bubbles or foaming when pipetting, mixing or reconstituting.
- Completely remove of all liquids when washing to prevent cross contamination.
- Prepare reagents immediately prior to use and do not store, with the exception of the top standard.
- Equilibrate all materials to ambient room temperature prior to use (standards exception).
- For optimal results for inter- and intra-assay consistency, equilibrate all materials to 37°C prior to performing assay (standards exception) and perform all incubations at 37°C.
- Pipetting less than 1 μ L is not recommended for optimal assay accuracy.
- Once the procedure has been started, all steps should be completed without interruption. Ensure that all reagents, materials and devices are ready at the appropriate time.
- Incubation times will affect results. All wells should be handled in the same sequential order and time intervals for optimal results.
- Samples containing bilirubin, precipitates or fibrin strands or are hemolytic or lipemic might cause inaccurate results due to interfering factors.
- TMB Substrate is easily contaminated and should be colorless or light blue until added to plate. Handle carefully and protect from light.

7. Sample Preparation

7.1 Sample Preparation and Storage

- Store samples to be assayed at 2-8°C for 24 hours prior being assayed.
- For long term storage, aliquot and freeze samples at -20°C. Avoid repeated freeze-thaw cycles.
- Samples not indicated in the manual must be tested to determine if the kit is valid.
- Prepare samples as follows:
 - **Serum** - Use a serum separator tube (SST) and allow samples to clot for 30 minutes at room temperature or overnight at 4°C before centrifugation for 15 minutes at 1,000 x g. Remove serum and assay immediately or aliquot and store samples at -20°C or -80°C. Avoid repeated freeze-thaw cycles.
 - **Plasma** - Collect plasma using EDTA, or heparin as an anticoagulant. Centrifuge for 15 minutes at 1,000 x g at 2-8°C within 30 minutes of collection. Assay immediately or aliquot and store samples at -20°C or -80°C. Avoid repeated freeze-thaw cycles.
 - **Cell culture supernatants** – Remove particulates and aggregates by spinning at approximately 1,000 x g for 10 minutes.

7.2 Sample Dilution

Target protein concentration must be estimated and appropriate sample dilution selected such that the final target protein concentration falls near the middle of the assay linear dynamic range. Samples exhibiting saturation should be further diluted. Note that a dilution of at least 1:2 will mitigate interference factors.

- Dilute samples using **1X Sample Diluent (Buffer)**
- Mix diluted samples gently and thoroughly.
- Pipetting less than 2 µL is not recommended for optimal assay accuracy.

8. Reagent Preparation

- Equilibrate all materials to room temperature prior to use and use prepare immediately prior to use.

8.1 1X Wash Buffer

- 8.1.1 Equilibrate **200X Wash Buffer** concentrate to room temperature and mix gently.
- 8.1.2 Add the entire 10 mL contents of the **200X Wash Buffer** bottle to 1990 mL of ultra-pure water to a clean > 2,000 mL bottle, graduated cylinder or other vessel.
- 8.1.3 Seal and mix gently by inversion. Avoid foaming or bubbles.
- 8.1.4 Store the **1X Wash Buffer** at room temperature until ready to use in the procedure. Store the prepared **1X Wash Buffer** at 4°C for no longer than 1 week. Do not freeze.

8.2 1X Standard Diluent (Buffer)

- 8.2.1 Equilibrate **10X Standard Diluent (Buffer)** concentrate to room temperature and mix gently.
- 8.2.2 Add the entire 25 mL contents of the **10X Standard Diluent (Buffer)** bottle to 225 mL of ultra-pure water to a clean > 500 mL bottle or other vessel.
- 8.2.3 Seal and mix gently by inversion. Avoid foaming or bubbles.
- 8.2.4 Store the **1X Standard Diluent (Buffer)** at room temperature until ready to use in the procedure. Store the prepared **1X Standard Diluent (Buffer)** at 4°C for no longer than 1 week. Do not freeze.

8.3 1X Biotinylated anti-IL-4 Detector Antibody


- 8.3.1 Prepare the **1X Biotinylated anti-IL-4 Detector Antibody** immediately prior to use by diluting the **27.5X Biotinylated anti-IL-4 Detector Antibody** 1:30 with **Biotinylated Detector Diluent** in volumes according to the table below.
- 8.3.2 Mix thoroughly and gently. Hold no longer than 2 hours prior to using in procedure. Do not store at 1X concentration for future use.

Number of wells required	Biotinylated IL-4 Detector Antibody (μL)	Biotinylated Antibody Diluent (μL)
16	40	1,060
32	80	2,120
48	120	3,180
96	240	6,360

8.4 Human IL-4 Assay Standards

- 8.4.1 Prepare the IL-4 standards no greater than 2 hours prior to performing experiment. Standards should be held on ice until use in the experiment.
- 8.4.2 Reconstitute one vial of the provided **IL-4 Standard** for each experiment according using the volume of **1X Standard Diluent** volume indicated on the vial label. This will produce a stock solution of 10 pg/mL **IL-4 Standard**. Prepare the standard as follows:
- 8.4.2.1 Gently spin or tap the vial at 6,000 – 10,000 rpm for 30 seconds to collect all material at the bottom.
- 8.4.2.2 Add the indicated volume of **1X Sample Diluent** to the vial.
- 8.4.2.3 Seal the vial then mix gently and thoroughly.
- 8.4.2.4 Leave the vial at ambient temperature for 15 minutes.
- 8.4.3 Prepare a set of seven serially diluted standards as follows:
- 8.4.3.1 Label tubes with numbers 2 – 8.
- 8.4.3.2 Use the undiluted 10 pg/mL **IL-4 Standard** as the high standard point (Tube #1).
- 8.4.3.3 Add 300 μ L of **Sample Diluent** to Tube #2 – 8.
- 8.4.3.4 Prepare **Standard #2** by adding 300 μ L of 10 pg/mL **IL-4 Standard**(Tube #1) to Tube #2. Mix gently and thoroughly.
- 8.4.3.5 Prepare **Standard #3** by adding 300 μ L of **Standard #2** from Tube #2 to Tube #3. Mix gently and thoroughly.
- 8.4.3.6 Prepare further serial dilutions through Tube #7. Reference the table below as a guide for serial dilution scheme.
- 8.4.3.7 Tube #8 is a blank standard (only **Sample Diluent**), which should be included with every experiment.

Standard Number (Tube)	Standard To Dilute	Volume Standard to Dilute (μ L)	Volume Sample Diluent Buffer (μ L)	Total Volume (μ L)	Final Concentration
1	10pg/mL Reconstituted IL-4 Standard	NA	NA	NA	10 pg/mL
2	10 pg/mL	300	300	600	5 pg/mL
3	5 pg/mL	300	300	600	2.5 pg/mL
4	2.5 pg/mL	300	300	600	1.25 pg/mL
5	1.25 pg/mL	300	300	600	0.62 pg/mL
6	0.62 pg/mL	300	300	600	0.31 pg/mL
7	NA	0	300	600	NA



8.5 Streptavidin-HRP Solutions

- 8.5.1 Prepare the **Streptavidin-HRP Stock Solution** immediately prior to use by adding 500 μL of **HRP Diluent**. Mix gently and thoroughly. Do not retain this vial for future experiments.
- 8.5.2 The stock solution will be used to prepare **Streptavidin-HRP Solution #1** and **Streptavidin-HRP Solution #2** for the assay detection steps. Dilute the re-suspended **Streptavidin-HRP Stock Solution** as indicated on the table below for the appropriate number of wells being run in the experiment. Prepare this solution immediately prior to use.
- 8.5.3 Mix thoroughly and gently. Do not store for future use.

Number of experimental wells	Streptavidin-HRP Solution #1 (For use in Step 10.11)		Streptavidin-HRP Solution #2 (For use in Step 10.19)	
	Streptavidin-HRP Stock Solution (μL)	Streptavidin-HRP Diluent (mL)	Streptavidin-HRP Stock Solution (μL)	Streptavidin-HRP Diluent (mL)
16	10	1.990	80	1.920
32	20	3.980	160	3.840
48	30	5.970	240	5.760
96	55	10.945	420	10.080

8.6 1X Amplifier

- 8.6.1 Dilute the **100X Amplifier** into **Amplifier Diluent** according to the following table.
- 8.6.2 Seal and mix gently by inversion. Avoid foaming or bubbles.
- 8.6.3 Use immediately prior to use and do not store.

Number of experimental wells	Amplifier (μL)	Amplifier Diluent (mL)
16	20	1.98
32	40	3.96
48	60	5.94
96	120	11.88

8.7 Microplate Preparation

- Micro-plates are provided ready to use and do not require rinsing or blocking.
- Unused well strips should be returned to the original packaging, sealed and stored at 4°C .
- Equilibrate microplates to ambient temperatures prior to opening to reduce potential condensation.

9. Assay Procedure

•Equilibrate all reagents and materials to ambient room temperature prior to use in the procedure.

- 9.1** Determine the required number of wells and return any remaining unused wells and desiccant to the pouch.
- 9.2** Add 100 µL of serially titrated standards, diluted samples or blank into wells of the **IL-4 Microplate**. At least two replicates of each standard, sample or blank is recommended.
- 9.3** Cover the plate with the well plate lid and incubate for 60 minutes with slow shaking at room temperature.
- 9.4** Remove the plate lid and discard the liquid in the wells by rigorously flicking into an acceptable waste receptacle or aspiration.
- 9.5** Wash plate 3 times with **1X Wash Buffer** as follows:
 - 9.5.1 Add 300 µL of **1X Wash Buffer** to each assay well.
 - 9.5.2 Incubate for 1 minute.
 - 9.5.3 Discard the liquid in the wells by rigorously flicking into an acceptable waste receptacle.
 - 9.5.4 Gently blot any remaining liquid from the wells by tapping inverted on the benchtop onto paper toweling. Do not allow the wells to completely dry at any time.
 - 9.5.5 Repeat steps 10.5.1 through 10.4.4 **two** more times.
- 9.6** Add 50 µL of prepared **1X Biotinylated anti-IL-4 Detector Antibody** to each well.
- 9.7** Cover with the well-plate lid and incubate for 60 minutes with slow shaking at room temperature.
- 9.8** Discard the liquid in the wells by rigorously flicking into an acceptable waste receptacle or aspiration.
- 9.9** Gently blot any remaining liquid from the wells by tapping inverted on the benchtop onto paper toweling. Do not allow the wells to completely dry at any time.
- 9.10** Wash plate as in Step 10.5.
- 9.11** Add 100 µL of prepared **Streptavidin-HRP Solution #1** into each well.
- 9.12** Cover with the well-plate lid and incubate for 20 minutes with slow shaking at room temperature.
- 9.13** Discard the liquid in the wells by rigorously flicking into an acceptable waste receptacle or aspiration.
- 9.14** Gently blot any remaining liquid from the wells by tapping inverted on the benchtop onto paper toweling. Do not allow the wells to completely dry at any time.
- 9.15** Wash the plate as in Step 10.5
- 9.16** Add 100 µL of **1X Amplifier** to all wells.
- 9.17** Cover with the well-plate lid and incubate for 15 minutes with slow shaking at room temperature.
- 9.18** Wash the plate as in Step 10.5
- 9.19** Add 100 µL of **Streptavidin-HRP Solution #2** to all wells.
- 9.20** Cover with the well-plate lid and incubate for 20 minutes with slow shaking at room temperature
- 9.21** Add 100 µL of the **TMB Substrate Solution** to all wells.
- 9.22** Incubate in the dark for 10 – 20 minutes. Avoid direct light exposure by covering plate with aluminum foil. Wells should change to gradations of blue. If the color is too deep, reduce the incubation time.
(NOTE: optimal incubation time must be determined by the user. Optimal development can be visualized by blue shading in the top four standard wells, while the remaining standards are still clear.)
- 9.23** Add 100 µL of **H₂SO₄ Stop Reagent** into all wells. Well color should change to yellow immediately. Add the **H₂SO₄ Stop Reagent** in the same well order as done for the **TMB Substrate Solution**.
- 9.24** Read the O.D. absorbance at 450 nm with a standard microplate reader within 5 minutes of stopping the reaction in step 10.16. If wavelength correction is available, set to 540 nm or 570 nm.

Assay Summary

Specific molecule detection steps	<p>Add 100 μL of sample or diluted standard</p> <p>↓</p> <p>Incubate 1 hour at room temperature with slow shaking</p> <p>↓</p> <p>Wash three times</p> <p>↓</p> <p>Add 50μL of diluted biotinylated detection antibody to all wells</p>
Streptavidin –HRP / Amplification steps	<p>↓</p> <p>Incubate 1 hour at room temperature with slow shaking</p> <p>↓</p> <p>Wash three times</p> <p>↓</p> <p>Add 100μL of Streptavidin-HRP Solution #1 to all wells</p> <p>↓</p> <p>Incubate 20 min at room temperature with slow shaking</p> <p>↓</p> <p>Wash three times</p> <p>↓</p> <p>Add 100 μL Amplifier to all wells</p> <p>↓</p> <p>Incubate 15 min at room temperature with slow shaking</p> <p>↓</p> <p>Wash three times</p> <p>↓</p> <p>Add 100 μL of Streptavidin-HRP Solution #2 to all wells</p>
Revelation and reading steps	<p>↓</p> <p>Incubate 20 min at room temperature with slow shaking</p> <p>↓</p> <p>Wash three times</p> <p>↓</p> <p>Add 100 μL of ready-to-use TMB-Substrate. Protect from light. Let the color develop for around 5 min.</p> <p>↓</p> <p>Add 100 μL H₂SO₄ Stop Reagent</p> <p>↓</p> <p>Read Absorbance at 450 nm</p>

10. Calculation of Results

For analysis of the assay results, calculate the **Relative OD₄₅₀** for each test or standard well as follows:

$$(\text{Relative OD}_{450}) = (\text{Well OD}_{450}) - (\text{Mean Blank Well OD}_{450})$$

The standard curve is generated by plotting the mean replicate **Relative OD₄₅₀** of each standard serial dilution point vs. the respective standard concentration. The IL-4 concentration contained in the samples can be interpolated by using linear regression of each mean sample **Relative OD₄₅₀** against the standard curve. This is best achieved using curve fitting software.

Note: if wavelength correction readings were available, subtract the readings at 540 nm or 570 nm from the readings at 450 nm. This may provide greater reading accuracy.


Note: if the samples measured were diluted, multiply the derived mean sample concentration by the dilution factor for a final sample concentration.

11. Typical Expected Data

11.1 Reproducibility

Two sera and two cell culture medium samples with various concentrations of IL-4 were tested for repeatability and reproducibility. Each assay was carried out with 8 replicates of each sample. Three independent assays were performed. The intra-assay and inter-assay coefficient of variation has been calculated to be 9.2% and 13.7% respectively.

11.2 Typical standard curve. This standard curve is for demonstration purposes only. An assay specific standard curve should be performed with each assay.



Standard	IL-4 (pg/mL)	Mean OD (450nm)	CV (%)
1	10	2.289	0.6
2	5	1.118	4.6
3	2.5	0.519	6.9
4	1.25	0.316	0.9
5	0.62	0.235	0.4
6	0.31	0.200	5.3
Zero	0	0.154	-

11.3 Linearity

Two pooled human sera and two cell culture medium samples containing different concentrations of IL-4 were serially diluted in standard buffer diluent. Linearity was evaluated on 4 dilutions. The linear regression of samples versus the expected concentration yielded a slope of 0.992. Recoveries ranged from 89% to 125% with an overall mean recovery of 101%.

11.4 Recovery

The spike recovery was evaluated by spiking three levels of IL-4 into two different pooled human sera and two cell culture medium. Recovery was evaluated with one test. The recovery in pooled human sera ranged from 61 to 75% with a mean of 69%. In cell culture medium, recovery was 89% and 127%, with a mean of 95%.

12. Technical Resources

Technical Support:

For optimal service please be prepared to supply the lot number of the kit used.

USA

Aviva Systems Biology, Corp.
5754 Pacific Center Blvd, Suite 201
San Diego, CA 92121

Phone: 858-552-6979
Toll Free: 888-880-0001
Fax: 858-552-6975

Technical support: techsupport@avivasysbio.com

China

Beijing AVIVA Systems Biology
6th Floor, B Building, Kaichi Tower
#A-2 Jinfu Road.
Daxing Industrial Development Zone
Beijing, 102600, CHINA

Phone: (86)10-60214720
Fax: (86)10-60214722
E-mail: support@avivasysbio.com.cn

中国地址: 北京大兴工业开发区金辅路甲 2 号凯驰大厦 B 座 6 层 (102600)
电话: 010-60214720/21
传真: 010-60214722

产品售前咨询及销售: sales@avivasysbio.com.cn
售后及技术支持: support@avivasysbio.com.cn