

Feline Serum Amyloid A Detection ELISA Kit

Catalog # 6028

For Research Use Only - Not Human or Therapeutic Use

PRODUCT SPECIFICATIONS

DESCRIPTION:	Assay kit to quantify feline SAA proteins
FORMAT:	Pre-coated 96-well ELISA Plate with removable strips
ASSAY TYPE:	Sandwich ELISA
ASSAY TIME:	4 hours
STANDARD RANGE:	2000 ng/ml to 31 ng/ml
NUMBER OF SAMPLES:	Up to 40 (duplicate) samples/plate
SAMPLE TYPES:	Serum and Plasma
RECOMMENDED SAMPLE DILUTIONS:	1:1 (at least)
CHROMOGEN:	TMB (read at 450 nm)
STORAGE:	-20°C
VALIDATION DATA:	Intra-Assay (5.7-8.1%)/Inter-Assay (5.1-9.6%)/Spiking Test (104-109%)
NOTES:	

© 2020 Chondrex, Inc. All Rights Reserved, 6028 2.0

Feline Serum Amyloid A Detection ELISA Kit

Catalog # 6028

For Research Use Only - Not Human or Therapeutic Use

INTRODUCTION

Acute phase proteins (APPs) are blood proteins that are an integral part of acute phase responses (APR). The APR is a part of the innate host defense system which is triggered by tissue damage and inflammation caused by infectious, immunologic, or neoplastic agents. One of the well-characterized APPs, serum amyloid A (SAA) is a 11.4 - 12.5 kDa protein consisting of 104 - 112 amino acids in different species. The protein is well-conserved in the evolution of eutherian mammals, indicating an important role in biological functions. The N-terminus of the molecule is hydrophobic and probably responsible for its lipid-binding properties which allows SAA to form complexes with high-density lipoproteins as well as apoproteins (1). In inflamed tissues, macrophages serve as a major source of SAA.

In fact, during APR, SAA levels in plasma increase 1000-fold in human and other species, suggesting it may be an important indicator of disease status. Elevated SAA levels are also seen in rheumatoid arthritis, atherosclerosis, Crohn's disease, and Type 2 diabetes. These findings suggest that SAA may play an active and/or a protective role in inflammatory disorders (2, 3).

In veterinary medicine, Tamamoto *et.al.* indicated that measuring feline SAA concentration, in addition to white blood cell counts, would be clinically valuable as a routine test to detect inflammation (4). Furthermore, SAA levels work as a significant and independent prognostic marker in cats with various diseases such as neoplastic diseases and inflammatory diseases (5).

Chondrex, Inc provides a Feline Serum SAA Detection ELISA Kit (Cat # 6028) which can be used to detect and monitor feline inflammation, as well as in studies investigating inflammation. Please contact Chondrex, Inc. at support@chondrex.com for more information.

KIT COMPONENTS

Item	Quantity	Amount	Storage
SAA Standard (60281)	1 vial	2000 ng, lyophilized	-20°C
Detection Antibody (60283)	1 vial	100 µl	-20°C
Solution B - Sample/Standard/Detection Antibody Dilution Buffer (67015)	1 bottle	50 ml	-20°C
Solution D - Streptavidin Peroxidase Dilution Buffer (9055)	1 bottle	20 ml	-20°C
Streptavidin Peroxidase (9029)	2 vials	50 µl	-20°C
TMB (90023)	2 vials	0.2 ml	-20°C
Chromogen Dilution Buffer (90022)	1 bottle	20 ml	-20°C
Stop Solution - 2N Sulfuric Acid (9016)	1 bottle	10 ml	-20°C
Wash Buffer, 20X (9005)	1 bottle	50 ml	-20°C
Capture Antibody Coated 96-Well ELISA Plate (Green)	1 each	8-well strips x 12	-20°C

© 2020 Chondrex, Inc. All Rights Reserved, 6028 2.0

ASSAY OUTLINE





PLATE LAYOUT



NOTES BEFORE USING ASSAY

NOTE 1: It is recommended that the standard and samples be run in duplicate.

NOTE 2: Warm up all buffers to room temperature before use.

NOTE 3: Crystals may form in Wash Buffer, 20X when stored at cold temperatures. If crystals have formed, warm the wash buffer by placing the bottle in warm water until crystals are completely dissolved.

NOTE 4: Measure exact volume of buffers using a serological pipet, as extra buffer is provided.


NOTE 5: Cover the plate with plastic wrap or a plate sealer after each step to prevent evaporation from the outside wells of the plate.

NOTE 6: For partial reagent use, please see the assay protocol's corresponding step for the appropriate dilution ratio. For example, if the protocol dilutes 50 μ l of a stock solution in 10 ml of buffer for 12 strips, then for 6 strips, dilute 25 μ l of the stock solution in 5 ml of buffer. Partially used stock reagents may be kept in their original vials and stored at -20°C for use in a future assay.

NOTE 7: This kit contains animal components from non-infectious animals and should be treated as potential biohazards in use and for disposal.

ASSAY PROCEDURE

1. **Prepare Standard Dilutions:** The recommended standard range is 31-2000 ng/ml. Dissolve one vial of standard in 1 ml of Sample/Standard/Detection Antibody Dilution Buffer (Solution B) for the 2000 ng/ml standard. Then serially dilute it with Solution B. For example, mix 250 μ l of the standard (2000 ng/ml) with an equal volume of Solution B to make a 1000 ng/ml solution, and then repeat it five more times for 500, 250, 125, 63, and 31 ng/ml solutions. The remaining 2000 ng/ml standard stock may be stored at -20°C for use in a second assay. Chondrex, Inc. recommends making fresh serial dilutions for each assay.



2. **Prepare Samples:** Samples must be diluted with Solution B to maintain optimal assay conditions.

Serum: Clot blood samples by incubating samples at room temperature for 2 hours

Plasma: Collect blood samples with the use of anticoagulants such as heparin. Then, collect plasma by centrifuging samples at 10,000 rpm for 5 minutes within 30 minutes of blood collection. When not in use, store the plasma supernatant samples at -20°C; avoid repeat freeze-thaw cycles.

NOTE: Using lipemic (cloudy) samples may affect assay results. The stored samples must be centrifuged at 10,000 rpm for an additional 5 minutes before the assay.

Dilution: Dilute samples at least 1:1 with Solution B depending on the estimated SAA level in the samples. Two to three different sample dilutions are recommended if the SAA levels in the samples are unknown.

3. **Add Standards and Samples:** Add 100 μ l of Solution B (blank), standards, and samples to designated wells in duplicate. Incubate at room temperature for 2 hours.
4. **Dilute Wash Buffer:** Dilute 50 ml of Wash Buffer, 20X in 950 ml of distilled water (1X wash buffer). Wash the plate with 1X wash buffer at least 3 times using a wash bottle with manifold or an automated plate washer. Empty the plate by inverting it and blotting on a paper towel to remove excess liquid. *Do not allow the plate to dry out.*
5. **Add Detection Antibody Solution:** Dilute one vial of detection antibody solution in 10 ml of Sample/Standard/Detection Antibody Dilution Buffer (Solution B). Add 100 μ l of detection antibody solution to each well and incubate at room temperature for 1 hour.

Strip #	Detection Antibody (μ l)	Solution B (ml)
2	17	1.7
4	33	3.3
6	50	5.0
8	66	6.6
10	82	8.2
12	100	10.0

6. **Wash:** Wash the plate with 1X wash buffer at least 3 times using a wash bottle with manifold or an automated plate washer. Empty the plate by inverting it and blotting on a paper towel to remove excess liquid. *Do not allow the plate to dry out.*
7. **Add Streptavidin Peroxidase Solution:** Dilute one vial of Streptavidin Peroxidase in 10 ml of Streptavidin Peroxidase Dilution Buffer. Add 100 μ l of streptavidin peroxidase solution to each well and incubate at room temperature for 30 minutes.

Strip #	Streptavidin Peroxidase (μ l)	Solution D (ml)
2	8	1.7
4	17	3.3
6	25	5.0
8	33	6.6
10	42	8.2
12	50	10.0

8. **Wash:** Wash the plate with 1X wash buffer at least 3 times using a wash bottle with manifold or an automated plate washer. Empty the plate by inverting it and blotting on a paper towel to remove excess liquid. *Do not allow the plate to dry out.*
9. **Add TMB Solution:** Use new tubes when preparing TMB solution. Just prior to use, prepare TMB solution with Chromogen Dilution Buffer as shown in the following table. Add 100 μ l of TMB solution to each well immediately after washing the plate and incubate for 25 minutes at room temperature.


Strip #	TMB (μ l)	Chromogen Dilution Buffer (ml)
2	34	1.7
4	66	3.3
6	100	5.0
8	132	6.6
10	164	8.2
12	200	10.0

10. **Stop:** Stop the reaction with 50 μ l of 2N Sulfuric Acid (Stop Solution) to each well.
11. **Read Plate:** Read the OD values at 450 nm. If the OD values of samples are greater than the OD values of the highest standard, re-assay the samples at a higher dilution. A 630 nm filter can be used as a reference.

CALCULATING RESULTS

1. Average the duplicate OD values for the blank, standards, and test samples.
2. Subtract the “blank” (B) values from the averaged OD values in step 1.
3. Plot the OD values of standards against the concentration of SAA (ng/ml). Using a log/log plot will linearize the data. Figure 1 shows a representative experiment where the standard range is 31 - 2000 ng/ml.
4. The ng/ml of SAA in test samples can be calculated using regression analysis.

Figure 1 - A Typical Standard Curve for the Feline SAA Detection ELISA Kit



VALIDATION DATA

Table 1 - Reproducibility Data for the Feline SAA Detection ELISA Kit

Test	63 ng/ml	250 ng/ml	1000 ng/ml
Intra-Assay CV (%)	5.7	6.6	8.1
Inter-Assay CV (%)	5.1	6.1	9.6
Spike Test*	109%	104%	109%

* Known amounts of SAA were added to feline serum and then diluted with Sample/Standard/Detection Antibody Dilution Buffer (Solution B).

TROUBLESHOOTING

For frequently asked questions about assays and ELISAs, please see Chondrex, Inc.’s [ELISA FAQ](#) for more information.

© 2020 Chondrex, Inc. All Rights Reserved, 6028 2.0

REFERENCES

1. G. Marhaug, S. Dowton, Serum Amyloid A: An Acute Phase Apolipoprotein and Precursor of AA Amyloid. *Baillieres Clin Rheumatol* **8**, 553-73 (1994).
2. R. D. Ye, L. Sun, Emerging functions of serum amyloid A in inflammation. *Journal of Leukocyte Biology*. **98**, 923–929 (2015).
3. K. Sasaki, Z. Ma, T. Khatlani, M. Okuda, H. Inokuma, T. Onishi, *et al.*, Evaluation of feline serum amyloid A (SAA) as an inflammatory marker. *J Vet Med Sci* **65**, 545-8 (2003).
4. T. Tamamoto, K. Ohno, A. Ohmi, Y. Goto-Koshino, H. Tsujimoto, Verification of Measurement of the Feline Serum Amyloid A (SAA) Concentration by Human SAA Turbidimetric Immunoassay and Its Clinical Application. *J Vet Med Sci* **70**, 1247-52 (2008).
5. T. Tamamoto, K. Ohno, M. Takahashi, K. Nakashima, Y. Fujino, H. Tsujimoto, *et al.*, Serum Amyloid A as a Prognostic Marker in Cats With Various Diseases. *J Vet Diagn Invest* **25**, 428-32 (2013).