ACE steri® エーステリ

滅菌ロールバッグ

蒸気滅菌、EO ガス滅菌、ホルマリン滅菌共用タイプ

EN868-1、EN-868-3、EN868-6 の規格適合品

機　能　特　長

- シールのムラや破損を見分けやすいブルーのフィルムを使用しています。
- ケミカルインジケーターの色の変化によって滅菌済であることが分かります。
- ケミカルインジケーターのインクはフィルムとの間に印刷されているため、従来品の様なインジケーターインクの滅菌物への付着は解消。
- 滅菌処理時の圧力に耐える高強度設定です。
- 蒸気、ガスの透過性を持った他孔質構造医用透析紙を使用しています。
- 滅菌後の細菌の侵入防ぎ、長期間無菌状態を保持できます。
- 3 種類の滅菌法（蒸気滅菌、EO ガス滅菌、ホルマリン滅菌）に対応します。

ACE steri®

滅菌後ケミカルインジケーター色調変化



材質

医用透析紙　：BILLERUD（SKARBLACKA, Sweden）
EN868-1、EN868-3、EN868-6 の規格適合
フィルム　：12μm PET（ポリエチレン）と 45μm CPP（無延伸ポリプロピレン）の二層構造

使用方法

必要な長さのエーステリをカットして、一端をヒートシールします。被滅菌物を挿入し、次いで他端をヒートシールして滅菌を行ってください。

製品価格リスト

品番	幅×長さ	単価	1c/s 入目	c/s 価格
H7100	100mm×200m	¥9,800	4 巻	¥39,200
H7150	150mm×200m	¥14,800	4 巻	¥59,200
H7200	200mm×200m	¥18,800	4 巻	¥75,200
H7250	250mm×200m	¥22,400	2 巻	¥44,800
H7300	300mm×200m	¥29,400	2 巻	¥58,800

※消費税別途

製造元　上海福沢医薬器材有限公司
発売元　福沢商事株式会社

人と科学のステキな未来へ
コスモ・バイオ株式会社

〒135-0016　東京都江東区東陽 2-2-20 東陽駅前ビル
URL：http://www.cosmobio.co.jp/
● 営業部（お問い合わせ）
TEL：(03) 5632-9610　FAX：(03) 5632-9619
TEL：(03) 5632-9620

Certificate of Registration

nqa global assurance

This is to certify that the Quality Management System of

Shanghai Fuze Medicine Equipment Co., Ltd.

No.3266 E, Jindu Road, Xinzhuang Industry Park, Shanghai, People's Republic of China

applicable to

Production and sales of biological indicators for sterilization

has been assessed and registered by NQA against the provisions of

ISO 9001:2000

This registration is subject to the company maintaining a quality management system, to the above standard, which will be monitored by NQA.
Please consult the website: www.snqa.com.cn

Certificate Number: **19922**

Date: 12 October 2005
Reissue Date: 12 October 2008
Valid Until: 12 October 2011
EAC Code: 13

Head of NQA

UKAS QUALITY MANAGEMENT
015

The use of the UKAS Accreditation Mark indicates accreditation in respect of those activities covered by the accreditation certificate number 015 held by NQA.
NQA is a trading division of NICEIC Group Ltd, Registration No. 02513162. Registered Office: Warwick House, Houghton Hall Park, Houghton Regis, Dunstable, Bedfordshire, LU5 5ZX.
This certificate is the property of NQA and must be returned on request.

Certificate of Registration

nqa global assurance

This is to certify that the Quality Management System of

Shanghai Fuze Medicine Equipment Co., Ltd.

No.3266 E, Jindu Road, Xinzhuang Industry Park, Shanghai, People's Republic of China

applicable to

Production and sales of biological indicators for sterilization

has been assessed and registered by NQA against the provisions of

ISO13485：2003

This registration is subject to the company maintaining a quality management system, to the above standard, which will be monitored by NQA.
Please consult the website: www.snqa.com.cn

Head of NQA

UKAS QUALITY MANAGEMENT
015

Certificate Number: **19924**

Date:	12 October 2005
Reissue Date:	12 October 2008
Valid Until:	12 October 2011
EAC Code:	13

The use of the UKAS Accreditation Mark indicates accreditation in respect of those activities covered by the accreditation certificate number 015 held by NQA.
NQA is a trading division of NICEIC Group Ltd, Registration No. 02513162. Registered Office: Warwick House, Houghton Hall Park, Houghton Regis, Dunstable, Bedfordshire, LU5 5ZX.
This certificate is the property of NQA and must be returned on request.