

RayBio[®] Human Sp1 Transcription Factor Activity Assay Kit

Catalog #: TFEH-SP1

User Manual


Jan 12, 2018



3607 Parkway Lane, Suite 200
Norcross, GA 30092

Tel: 1-888-494-8555 (Toll Free) or 770-729-2992, Fax: 770-206-2393

Web: www.RayBiotech.com, Email: info@raybiotech.com



RayBiotech, Inc.

RayBio® Human Sp1 TF Activity Assay Kit Protocol

TABLE OF CONTENTS

| | | |
|-------|------------------------------------|----|
| I. | Introduction..... | 2 |
| II. | How It Works..... | 4 |
| III. | Reagents..... | 5 |
| IV. | Storage..... | 5 |
| V. | Additional Materials Required..... | 6 |
| VI. | Reagent Preparation..... | 6 |
| VII. | Assay Procedure..... | 9 |
| VIII. | Assay Procedure Summary..... | 11 |
| IX. | Typical Data..... | 12 |
| X. | Troubleshooting Guide | 14 |

Please read the entire manual carefully before starting your experiment.

I. INTRODUCTION


Sp1 is a 785-amino-acid, 100- to 110-kDa nuclear transcription factor belonging to a specific protein family which has 8 (Sp1–Sp8) members. It is ubiquitously expressed and regulates the expression of thousands of genes implicated in the control of a diverse array of cellular processes, such as cell growth, differentiation, apoptosis, angiogenesis, and immune response, via multiple mechanisms in response to physiologic and pathological stimuli. Sp1 possesses C2H2-type zinc finger domain through which it binds GC-rich motifs (such as 5'-G/T-GGGCGG-G/A-G/A-C/T-3' or 5'-G/T-G/A-GGCG-G/T-G/A-G/A-C/T-3') of a DNA sequence in promoters of target genes with high affinity. It can also regulate the expression of genes containing and not containing TATA and TATA genes via protein-protein interactions or interplay with other transcription factors, such as Ets-1, c-myc, c-Jun, Stat1, and Egr-1, and/or components of the basal transcriptional machinery. Sp1 has been linked to chromatin remodeling through interactions with chromatin-modifying factors such as p300 and histone deacetylases (HDACs). It is expected that posttranslational modifications including phosphorylation, acetylation, sumoylation, ubiquitylation, and glycosylation can influence the transcriptional activity and stability of Sp1.

Accurate monitoring of the level of activated Sp1 in cells, tissues or animal models is required for both science research investigating signal transduction pathways and applications such as drug development, and simple, speedy and high-throughput methods are needed for this purpose.

Traditionally, western blot has been used to detect the expression or modification of Sp1, electrophoretic mobility shift assays (EMSA) has been used to detect the DNA binding capacity of Sp1, and transfection of reporter genes such as luciferase and β -galactosidase with Sp1 binding sites in culture cells have been used in evaluation of ER reactivity. However, these methods are time consuming, laborious, and sometimes require the use of radioactivity.

The RayBio® Sp1 TF Activity Assay kit is a non-radioactive transcription factor assay with an ELISA format. It offers an easy, quick, sensitive and high-throughput method to detect the activation of transcription factors. Double-stranded oligonucleotides containing the Sp1 binding sequence have been coated on 96-well plates. These oligonucleotides specifically capture the active Sp1 contained in whole cell lysate or nuclear extracts after a short incubation. Subsequently, the primary antibody against Sp1 recognizes the Sp1-DNA complex in each well, and a HRP-conjugated secondary antibody is then used for detection. After washing away any unbound antibody, signal can be obtained easily through a colorimetric assay with a spectrophotometric plate reader at 450 nm. The specificity of the reaction between active Sp1 and the DNA probe is additionally stringent because of the establishment of specific competitive DNA and non-specific competitive DNA probes in this reaction system.

II. HOW IT WORKS



III. STORAGE

Upon receipt, the positive control should be removed and stored at -20° or -80°C. The remainder of the kit can be stored for up to 6 months at 2-8°C from the date of shipment. Opened Microplate Wells or reagents may be stored for up to 1 month at 2-8°C. Return unused wells to the pouch containing desiccant pack, reseal along entire edge.

Note: The kit can be used within one year if the whole kit is stored at -20°C upon receipt. Avoid repeated freeze-thaw cycles.

IV. REAGENTS

| Component | Description | Size |
|-----------------------------------|--|-----------------|
| Sp1 DNA Probe Microplate | 96 wells (12 strips X 8 wells) coated with Sp1 probes | 1 plate |
| DNA Binding Buffer | 5X concentrated Buffer | 4 ml |
| Positive Control | Cell nuclear extracts | 1 vial (20 µl) |
| Specific Competitor DNA Probe | Free DNA probes that compete with the coated probes by binding with activated Sp1. | 1 vial |
| Non-specific Competitor DNA Probe | Free DNA probes with mutations of the coated DNA probe. Cannot bind activated Sp1. | 1 vial |
| Assay Reagent | 1X solution | 1 vial (200 µl) |
| DTT | 300 mM DTT | 1 vial (200 µl) |
| Wash Buffer Concentrate (20X) | 20X concentrated solution | 25 ml |
| Sp1 Primary Antibody | Anti- Sp1 antibody | 1 vial |
| HRP-conjugated Secondary Antibody | Anti-IgG HRP conjugated antibody | 1 vial |
| Antibody Diluent Buffer | Buffer solution for diluting primary and secondary antibodies | 25 ml |
| TMB One-Step Substrate Reagent | 3,3,5,5'-tetramethylbenzidine (TMB) in buffer solution | 12 ml |
| Stop Solution | 0.2 M sulfuric acid | 8 ml |

V. ADDITIONAL MATERIALS REQUIRED

- 1 Microplate reader capable of measuring absorbance at 450 nm.
- 2 Precision pipettes to deliver 1 μ l to 1 ml volumes.
- 3 Adjustable 1-25 ml pipettes for reagent preparation.
- 4 100 ml and 1 liter graduated cylinders.
- 5 Absorbent paper.
- 6 Distilled or deionized water.
- 7 Tubes to prepare positive or sample mixtures.

VI. REAGENT PREPARATION

1. Preparation of samples:
Prepare nuclear extraction or whole lysate containing targeted protein Sp1 from cell culture or tissue. We recommend using the RayBiotech Nuclear Extraction Kit (Cat#: NE-50) to isolate nuclear proteins for subsequent use in this transcription factor assay.
2. Preparation of transcription factor binding reaction system:
Bring all reagents to room temperature (18 - 25°C) before use. Thaw the positive control and samples and keep them on ice before adding into wells. Prepare 100 μ l transcription factor binding reaction system for each well with 5 x DNA Binding Buffer, Assay Reagent, DTT, Specific Competitor DNA Probe, Non-specific Competitor DNA Probe, and Positive Control or samples containing targeted proteins. Typical examples are shown in the table below.

Note:

Each reaction may be prepared in a labeled microfuge tube or directly in the coated plate well. If the reaction system is prepared directly in the coated plate wells, please add the reagents sequentially as shown in the table to get the best results.

| COMPONENT | REACTION | | | | |
|---|--|--|--|--|--|
| | Positive control | Sample | Specific competitor | Non-Specific competitor | Blank |
| 5x DNA Binding Buffer | 20 μ l | 20 μ l | 20 μ l | 20 μ l | 20 μ l |
| Assay Reagent | 1.5 μ l | 1.5 μ l | 1.5 μ l | 1.5 μ l | 1.5 μ l |
| DTT | 1 μ l | 1 μ l | 1 μ l | 1 μ l | 1 μ l |
| Specific Competitor DNA probe | - | - | 10 μ l | - | - |
| Non-specific Competitor DNA probe | - | - | - | 10 μ l | - |
| Positive Control/Sample containing proteins | 5 μ l | * μ l | * μ l | * μ l | - |
| Total volume | bring final volume to 100μl with deionized water | bring final volume to 100μl with deionized water | bring final volume to 100μl with deionized water | bring final volume to 100μl with deionized water | bring final volume to 100μl with deionized water |

* Please note that the amount of total protein containing the target protein to be used in this test can be optimized and must be determined by the investigator.

3. Preparation of primary antibody:

Briefly spin down the Sp1 Primary Antibody vial. Add 60 μ l of Antibody Diluent Buffer into the vial to prepare a primary antibody concentrate. Pipette up and down to mix gently (the concentrate can be stored at 4°C for 5 days). The primary antibody concentrate should then be diluted 100-fold with the Antibody Diluent Buffer and used in step 4 of Part VII Assay Procedure.

4. Preparation of secondary antibody:

Briefly spin down the HRP-conjugated Secondary Antibody vial before use. Add 100 μ l of Antibody Diluent Buffer into the vial to

prepare a detection antibody concentrate. Pipette up and down to mix gently (the concentrate can be stored at 4°C for 5 days). The detection antibody concentrate should then be diluted 100-fold with the Antibody Diluent Buffer and used in step 6 of Part VII Assay Procedure.

5. Preparation of 1x Wash Buffer:

Dilute 25 ml of the 20x Wash Buffer Concentrate into deionized or distilled water to yield 500 ml of 1x Wash Buffer. If the Wash Buffer Concentrate (20x) contains visible crystals, warm to room temperature and mix gently until dissolved.

Note:

All reagents containing protein (positive control, samples) should be kept on ice to maintain protein stability.

If the reaction system is prepared directly in the coated plate wells, please add the reagents sequentially as shown in the table to get the best results.

To observe the specificity of the DNA binding activity, the amount of protein used in wells of sample, specific competitor and non-specific competitor must be the same.

A positive control should be included every time to confirm correct operation of the experiment, however it is not necessary to run the specific competitor and non-specific competitor for each sample every time.

VII. ASSAY PROCEDURE:

1. Bring the 96-well plate to room temperature (18 - 25°C) before use. If the whole plate will not be used in this assay, place remaining wells back to 2 to 8°C or -20°C. It is recommended that all positive control and samples be run at least in duplicate.
2. Add 100 µl of each prepared transcription factor binding reaction system (see Reagent Preparation step 2) including positive control, specific competitor, non-specific competitor and sample into appropriate wells. Cover wells and incubate for 2 hours at room temperature or overnight at 4°C with gentle shaking.
3. Discard the solution and wash 4 times by filling each well with 300 µl of 1x Wash Buffer (Reagent Preparation step 5) using a multi-channel pipette or autowasher. Complete removal of liquid at each step is essential for good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
4. Add 100 µl of prepared Sp1 Primary Antibody (Reagent Preparation step 3) to each well. Incubate for 1 hour at room temperature with gentle shaking.
5. Discard the solution. Wash as directed in step 3.
6. Add 100 µl of prepared HRP-conjugated Secondary Antibody (see Reagent Preparation step 4) to each well. Incubate for 1 hour at room temperature with gentle shaking.
7. Discard the solution. Wash as directed in step 3.
8. Add 100 µl of TMB One-Step Substrate Reagent to each well. Incubate for 30 minutes at room temperature in the dark with gentle shaking.

9. Add 50 μ l of Stop Solution to each well. Read at 450 nm immediately.

VII. ASSAY PROCEDURE SUMMARY

1. Prepare all reagents, samples and standards as instructed.



2. Add 100 μ l sample to each well.
Incubate 2 hours at room temperature or overnight at 4°C.



3. Add 100 μ l prepared primary antibody to each well.
Incubate 1 hour at room temperature.



4. Add 100 μ l prepared secondary antibody.
Incubate 1 hour at room temperature.



5. Add 100 μ l TMB One-Step Substrate Reagent to each well.
Incubate 30 minutes at room temperature.



6. Add 50 μ l Stop Solution to each well.
Read at 450 nm immediately.

VIII. TYPICAL DATA




Figure 1: Transcription factor assay of Sp1 from nuclear extracts of HeLa cells and HeLa cells treated with PMA (50 ng/ml) for 3 hr with the RayBio® Activity Assay Kit (cat # TFEH- SP1).




Fig. 2: Transcription factor assay of Sp1 from nuclear extracts of HeLa cells and HeLa cells treated with PMA (50 ng/ml) for 3 hr with the specific competitor and non-specific competitor. The result shows specific binding of Sp1 to the conserved binding site detected by using the RayBio® Sp1 TF Activity Assay Kit (cat # TFEH- SP1).

TROUBLESHOOTING GUIDE

| Problem | Cause | Solution |
|--------------------|---|--|
| 1. Low signal | <ol style="list-style-type: none"> 1. Too brief incubation times 2. Missed key reagent, inadequate reagent volumes or improper dilution 3. Not enough targeted protein per well 4. Inadequate development in colorimetric assay | <ol style="list-style-type: none"> 1. Ensure sufficient incubation time; change incubation time in assay procedure step 2 to overnight 2. Check to ensure all reagents have been added to ensure correct preparation 3. Check positive control wells and increase the amount of sample. 4. Ensure correct developing buffer and enough time used |
| 2. Large CV | <ol style="list-style-type: none"> 1. Inaccurate pipetting 2. Wells cross contamination | <ol style="list-style-type: none"> 1. Check pipettes 2. Be careful when preparing samples between wells |
| 3. High background | <ol style="list-style-type: none"> 1. Plate is insufficiently washed 2. Contaminated wash Buffer 3. Incorrect antibody dilution | <ol style="list-style-type: none"> 1. Review the manual for proper wash. If using a plate washer, check that all ports are unobstructed 2. Make fresh wash buffer 3. Check antibody dilutions |

RayBio® TF Activity Assay kits:

Choose TF Activity assay kits with more targets for human, mouse, rat and a variety of other species. Visit www.raybiotech.com for the complete list.

This product is for research use only



©2018 RayBiotech, Inc.