

# **RayBio® Label-Based (L-Series) Human Antibody Array L-8000 Glass Slide Kit**

A combination of Human L-507, L-493, L-3, L-4, L-5, L-6, L-7, L-8, L-9, L-10, L-11, L-12, L-13, L-14, L-15, and L-16 arrays

## **Patent Pending Technology User Manual (October 24, 2022)**

For the simultaneous detection of the relative expression of 8000 human proteins in serum, plasma, cell culture supernatants, cell/tissue lysates or other body fluids.

**L-Series Human Antibody Array L-8000**  
**Cat# AAH-BLG-8000-4 (4 Sample Kit)**  
**Cat# AAH-BLG-8000-8 (8 Sample Kit)**

**Please read manual carefully  
before starting experiment**



**Your Provider of Excellent Protein Array Systems and Services**

---

**Tel: (Toll Free) 1-888-494-8555 or +1-770-729-2992; Fax: +1-770-206-2393; Website: [www.raybiotech.com](http://www.raybiotech.com) Email: [info@raybiotech.com](mailto:info@raybiotech.com)**

## TABLE OF CONTENTS


---

I.	Introduction and How It Works.....	2
II.	Materials Provided.....	3
	A. Storage Recommendations.....	3
	B. Additional Materials Required.....	3
III.	Overview and General Considerations.....	4
	A. Preparation and Storage of Samples.....	4
	B. Handling the Glass Slides.....	6
	C. Layout of Array Slide.....	7
	D. Incubation and Washes.....	7
IV.	Protocol.....	9
	A. Sample Purification.....	9
	B. Biotin Labeling of Sample.....	10
	C. Drying of the Glass Slide.....	11
	D. Blocking and Incubations.....	12
	E. Fluorescence Detection.....	14
V.	Antibody Array Map .....	15
VI.	Antibody Array Target Lists.....	17
VII.	Interpretation of Results.....	34
	A. Explanation of Controls Spots.....	34
	B. Typical Results.....	34
	C. Background Subtraction.....	35
	D. Normalization of Array Data.....	36
	E. Threshold of Significant Difference.....	36
	F. Pathway Analysis of the Array Proteins.....	37
VIII.	Troubleshooting Guide.....	38
IX.	Selected References.....	39

## I. Introduction

Combining direct antigen-labeling technology with our vast library of array-validated antibodies, RayBiotech has created the largest commercially available antibody array to date. With the L-Series high density array platform, researchers can now detect thousands of proteins simultaneously, obtaining a broad, panoramic view of protein expression. Our newly expanded panel includes a wide variety of metabolic enzymes, structural proteins, epigenetic markers, neuroregulatory factors, in addition to our popular list of cytokines, growth factors, receptors, adipokines, proteases, and signaling proteins. Available on both glass slide and membrane formats, this array is ideally suited for biomarker discovery studies and exploratory screens.

The first step in using the RayBio® L-Series Antibody Array is to biotinylate the primary amine groups of the proteins in your sample (sera or plasma, cell culture supernatants, cell lysates or tissue lysates). The glass slide arrays are then blocked, just like a western blot, and the biotin-labeled sample is added onto the glass slide, which is pre-printed with capture antibodies. The slide is incubated to allow binding of target proteins. Streptavidin-conjugated fluorescent dye (Cy3 equivalent) is then applied to the array. Finally, the glass slide is dried, and laser fluorescence scanning is used to visualize the signals.



## II. Materials Provided

### A. Storage Recommendations

Upon receipt, the kit should be stored at -20°C until needed. It is recommended to use the kit within 6 months of the date of shipment. After initial use, remaining reagents should be stored at 4°C and may be stored for up to 3 months. Labeling Reagent (Item B) should be prepared fresh each time before use. Unused glass slides should be kept at -20 °C and repeated freeze-thaw cycles should be avoided (slides may be stored for 6 months).

ITEM	DESCRIPTION	4 SAMPLE Kit	8 SAMPLE Kit
A	Spin Columns (2ml)	8 columns	16 columns
B	Labeling Reagent	1 vial	2 vials
D	Stop Solution	1 vial (50 µl)	2 vials (50 µl)
E	RayBio® L-Series Glass Slide*	1 slide each of Human L-507, L-493, L-3, L-4, L-5, L-6, L-7, L-8, L-9, L-10, L-11, L-12, L-13, L14, L-15, and L-16	2 slides each of Human L-507, L-493, L-3, L-4, L-5, L-6, L-7, L-8, L-9, L-10, L-11, L-12, L-13, L14, L-15, and L-16
F	Blocking Buffer	4 bottles (30 ml)	8 bottles (30 ml)
G	20X Wash Buffer I	4 bottles (30 ml)	8 bottles (30 ml)
H	20X Wash Buffer II	4 bottles (30 ml)	8 bottles (30 ml)
I	Cy3 equivalent-Conjugated Streptavidin	6 vials	12 vials
J	Adhesive Plastic Strips		
K	Labeling Buffer	2 bottles (30 ml)	4 bottles (30 ml)
n/a	2X Cell Lysis Buffer**	1 bottle (10 ml)	2 bottles (10 ml)
M	30 ml Centrifuge Tube	4 tubes	8 tubes

\*Each slide contains 4 identical subarrays

\*\*Only needed if testing cell or tissue lysates

### B. Additional Materials Required

- 1 ml tube, small plastic or glass containers
- Orbital shaker or oscillating rocker
- Pipettors, pipette tips and other common lab consumables
- Laser scanner for fluorescence detection

- Aluminum foil

### III. Overview and General Considerations

#### A. Preparation and Storage of Samples

##### 1) Preparation of Cell Culture Supernatants

1. Seed cells at a density of  $1 \times 10^6$  cells in 100 mm tissue culture dishes.\*
2. Culture cells in complete culture medium for ~24–48 hours.\*\*
3. Replenish with serum-free or low-serum medium such as 0.2% FCS/FBS serum, and then incubate cells again for ~48 hours.\*\*\*<sup>†</sup> The membrane-based array is recommended if high serum medium such as 10% FCS/FBS is used, as high background can occur on glass slide arrays with high serum containing media samples.
4. To collect supernatants, centrifuge at 1,000 x g for 10 minutes and store as  $\leq 1$  ml aliquots at -80°C until needed.
5. If you want to use cell mass for inter-sample normalization, measure the total wet weight of cultured cells in the pellet and/or culture dish. You may then normalize between arrays by dividing fluorescent signals by total cell mass (i.e., express results as the relative amount of protein expressed/mg total cell mass). Or you can normalize between arrays by determining cell lysate concentration using a total protein assay (BCA Protein Assay Kit, Pierce, Prod #: 23227).

\*The density of cells per dish used is dependent on the cell type. More or less cells may be required.

\*\*Optimal culture time may vary and will depend on the cell line, treatment conditions and other factors.

<sup>†</sup>Bovine serum proteins produce detectable signals on the RayBio® L-Series Array in media containing serum concentrations as low as 0.2%. When testing serum-containing media, we strongly

*recommend testing an uncultured media blank for comparison with sample results.*

## 2) Extracting Protein from Cells

### 1. Centrifuging Cells

#### a. Adherent Cells:

- i. Remove supernatant from cell culture and wash cells gently twice with cold 1X PBS taking care not to disturb cell layer.
- ii. Add enough cold 1X PBS to cover cell layer and use cell scraper to detach cells.

#### b. Cells in Suspension: Pellet the cells by centrifuging using a microcentrifuge at 1500 rpm for 10 minutes.

2. Make sure to remove any remaining PBS before adding 1X Cell Lysis Buffer (2X Cell Lysis Buffer should be diluted 2-fold with ddH<sub>2</sub>O). Solubilize the cells at 2x10<sup>7</sup> cells/ml in 1X Cell Lysis Buffer.
3. Pipette up and down to resuspend cells and rock the lysates gently at 2–8 °C for 30 minutes. Transfer extracts to microfuge tubes and centrifuge at 13,000 rpm for 10 minutes at 2-8 °C.

*Note: If the lysates appear to be cloudy, transfer the lysates to a clean tube, centrifuge again at 13,000 rpm for 20 minutes at 2-8°C. If the lysates are still not clear, store them at -20°C for 20 minutes. Remove from the freezer and immediately centrifuge at 13,000 rpm for 20 minutes at 2-8°C.*

4. Transfer lysates to a clean tube. Determining cell lysate concentrations using a total protein assay (BCA Protein Assay Kit, Pierce, Prod# 23227). Aliquot the lysates and store at -80°C.

## 3) Extracting Protein from Crude Tissue

1. Transfer approximate 100 mg crude tissue into a tube with 1 ml 1X Cell Lysis Buffer (2X Cell Lysis Buffer should be diluted 2-fold with ddH<sub>2</sub>O).
2. Homogenize the tissue according to homogenizer manufacturer instructions.
3. Transfer extracts to microcentrifuge tubes and centrifuge for 20 minutes at 13,000 rpm (4°C).

*Note: If the supernatant appears to be cloudy, transfer the supernatants to a clean tube, centrifuge again at 13,000 rpm for 20 minutes at 2-8°C. If the supernatant is still not clear, store the lysate at -20°C for 20 minutes. Remove from the freezer, immediately centrifuge at 13,000 rpm for 20 minutes at 2-8°C.*

4. Transfer supernatant to a clean tube and store at -80°C.
- 4) Determine the total protein concentration  
For optimal biotin labeling, it is necessary to determine the protein concentration in the cell/tissue lysate. We recommended using a BCA total protein assay (e.g., Pierce, Catalog # 23227).


## B. Handling the Glass Slides

- The microarray slides are delicate. Please do not touch the array surface with pipette tips, forceps or your fingers. Hold the slides by the edges only.
- Handle the slides with powder-free gloves and in a clean environment.
- Do not remove the glass slide from the chamber assembly until step 20, and take great care not to break the glass slide when doing so.
- Remove reagents/sample by gently applying suction with a pipette to corners of each chamber. Do not touch the printed area of the array, only the sides as seen in image below.



### C. Layout of Array Slide

Four identical sub-arrays on one slide



4 printed sub-arrays per glass chip

### D. Incubations and Washes


- Cover incubation chamber with a Plastic Adhesive Strip (Item J) to prevent evaporation during incubation or wash steps, particularly those steps lasting 2 hours or longer.
- During incubation and wash steps avoid foaming and remove all bubbles from the sub-array surface.

- Perform all incubation and wash steps under gentle rotation or rocking motion (~0.5 to 1 cycle/sec).
- Wash steps in Wash Buffer II and all incubation steps may be performed overnight at 4°C.
- Avoid cross-contamination of samples to neighboring wells. To remove Wash Buffers and other reagents from chamber wells, you may invert the Glass Slide Assembly to decant, and aspirate the remaining liquid.
- Unlike most Cy3 fluors, the streptavidin-conjugated fluor used in this kit is very stable at room temperature (RT) and resistant to photobleaching on the hybridized glass slides. However, please protect glass slides from direct, strong light and temperatures above RT.


## IV. Protocol

### Assay Diagram

#### 1. Cell/tissue lysates



#### 2. Serum, plasma, body fluid, or Cell culture supernatants



### A. Sample purification

*Note: This step removes low molecular weight amine derivatives or unwanted buffer from samples to ensure quality biotinylation in Steps 5–7.*

1. Twist to remove the bottom plug of the Spin Column and loosen the cap (do not remove).
2. Place the Spin column into a 15 ml conical collection tube, centrifuge at 1,000 x g for 3 minutes to remove the storage buffer. Discard the flow-through.
3. Wash the Spin Column three times with 1 ml Labeling Buffer each, centrifuge at 3,000 x g for 3 minute to remove the flow-through.

Discard the flow-through and blot the bottom of the column to remove excess liquid. Transfer the Spin Column to a new collection tube.

4. Apply sample on top of the resin within the next few minutes. Centrifuge at 1,000 x g for 3 minutes. Collect the flow-through that contains the sample. The recommended sample dilutions are as follows:

- *Culture Media: 600 µl neat supernatant*
- *Serum/Plasma: 10 µl serum/plasma in 600 µl labeling buffer*
- *Cell/tissue lysate: 100 µg lysate in 500 µl labeling buffer*
- *Other body fluid: 100 µg in 600 µl labeling buffer*

*Note: Each labelled sample volume is enough for at least 20 arrays following the protocol below.*

*Note: The maximal sample volume is 700 µl for each Spin Column. Do not load over 700 µl of sample into a Spin Column.*

## B. Biotin-Labeling the Sample

*Note: Amines (e.g., Tris, glycine) and azides quench the biotinylation reaction. Avoid contaminating samples with these chemicals prior to biotinylation.*

5. Immediately before use, prepare the Labeling Reagent. Briefly spin down the Labeling Reagent tube (Item B). Add 100 µl Labeling Buffer into the tube, then pipette up and down or vortex slightly to dissolve the lyophilized reagent.
6. Add Labeling Reagent to the sample tube. Incubate the reaction solution at RT with gentle rocking or shaking for 30 min. Mix the reaction solution by gently tapping the tube every 5 minutes.

- a. For labeling cell culture supernatants: Add 10 µl of Labeling Reagent into the sample tube (for 600 µl supernatant).
- b. For labeling serum or plasma: Add 10 µl of Labeling Reagent Solution into the sample tube (for 10 µl serum/plasma *in 600 µl labeling buffer*).
- c. For labeling cell or tissue lysates: Add 5 µl of Labeling Reagent Solution into the sample tube (for 100 µg lysate *in 500 µl labeling buffer*).
- d. For all other body fluid: Add 8 µl of Labeling Reagent Solution per 100 µg sample to be labelled.

*Note: The addition of Labeling Reagent volume is based upon the sample amount used in Step 4. If the amount of sample being labelled differs from the example in Step 6, adjust this volume proportionally.*

7. Add 5 µl Stop Solution (Item D) to each sample tube. Using a new spin column, repeat Steps 1-4 of section A. Sample Purification to remove the excess non-reacted biotin reagent from each sample.

*Note: Biotinylated samples can be stored at -20°C or -80°C until you are ready to proceed with the assay.*

### C. Drying the Glass Slide

8. Remove the package containing the Assembled Glass Slide (Item E) from the freezer. Place unopened package on the bench top for ~15 minutes, and allow the Assembled Glass Slide to equilibrate to RT.
9. Open package, and take the Assembled Glass Slide out of the sleeve. Do not disassemble the Glass Slide from the chamber assembly. Place glass slide assembly in laminar flow hood or similar clean environment for 1-2 hours at RT.

*Note: Protect the slide from dust or other contaminants.*

## D. Blocking and Incubations

*Note: Glass slide should be completely dry before adding Blocking Buffer to wells.*

10. Block sub-arrays by adding 400 µl of Blocking Buffer (Item F) into each well of Assembled Glass Slide and incubating at RT for 30 minutes. Ensure there are no bubbles on the array surfaces.
11. Dilute samples with Blocking Buffer. Recommended dilution of the biotin-labeled samples with Blocking Buffer is 20-fold for cell culture supernatants and serum/plasma, 100-fold for cell/tissue lysate. *Dilution for other body fluid needs to be determined by the end user. Generally, most samples can be 10-20x dilution, while tears and saliva samples may need 100x dilution.*

*Note: Optimal sample dilution factor will depend on the abundance of target proteins. If the background or antigen-specific antibody signals are too strong, the sample can be diluted further in subsequent experiments. If the signal is too weak, more concentrated samples can be used.*

12. Completely remove the Blocking Buffer from each well. Add 400 µl of diluted sample into appropriate wells. Remove any bubbles on array surfaces. Incubate arrays with gentle rocking or shaking for 2 hours at RT or overnight at 4°C.

*Note: Avoid the flow of sample into neighboring wells.*

13. Based on number of samples and remaining protocol, calculate the amount of 1X Wash Buffer I and 1X Wash Buffer II needed to complete the experiment. Separately dilute the required amounts of 20X Wash Buffer I Concentrate (Item G) 20-fold and 20X Wash Buffer II Concentrate (Item H) with ddH<sub>2</sub>O.

14. Decant the samples from each well and wash 3 times with 800 µl of 1X Wash Buffer I at RT with gentle rocking or shaking for 5 minutes per wash.
15. Obtain a clean container (e.g., pipette tip box or slide-staining jar), place the Assembled Glass Slide into the container with enough volume of 1X Wash Buffer I to completely cover the entire assembly, and remove any bubbles in wells. Wash 2 times at RT with gentle rocking or shaking for 10 minutes per wash.
16. Decant the Wash Buffer I from each well, place the Assembled Glass Slide into the container with enough volume of 1X Wash Buffer II to completely cover the entire assembly, and remove any bubbles in wells. Wash 2 times at RT with gentle rocking or shaking for 5 minutes per wash.
17. Prepare 1X Cy3-Conjugated Streptavidin:
  - a) Briefly spin down tube containing the Cy3-Conjugated Streptavidin (Item I) immediately before use.
  - b) Add 1000 µl of Blocking Buffer into the Cy3-Conjugated Streptavidin tube to prepare a concentrated Cy3-Conjugated Streptavidin stock solution. Pipette up and down to mix gently (do not store the stock solution for later use).
  - c) To prepare 1X Cy3-Conjugated Streptavidin, add 200 µl of the concentrated Cy3-Conjugated Streptavidin stock solution into a tube with 800 µl of Blocking Buffer. Mix gently.
18. Carefully remove Assembled Glass Slide from container. Remove all of Wash Buffer II from the wells. Add 400 µl of 1X Cy3-Conjugated Streptavidin to each sub-array. Cover the incubation chamber with the plastic adhesive strips.


*Note: Avoid exposure to light in Steps 19–25 by covering the Glass Slide Assembly with aluminum foil or incubate in a dark room.*

19. Incubate with 1X Cy3-Conjugated Streptavidin at RT for 1 hour with gentle rocking or shaking.

*Note: Incubation may be done overnight at 4°C.*

20. Decant the solution and disassemble the glass slide from the incubation frame and chamber. Disassemble the device by pushing clips outward from the side, as shown below. Carefully remove the glass slide from the gasket.

*Note: Be careful not to touch the printed surface of the glass slide, which is on the same side as the barcode.*



21. Gently place the glass slide into 30 ml Centrifuge Tube (Item M). Add enough 1X Wash Buffer I to cover the entire glass slide (about 30 ml). Wash with gentle rocking or shaking for 10 min. Remove the wash buffer. Repeat 2 times for a total of 3 washes.
22. Add enough 1X Wash Buffer II to cover the entire glass slide (about 30 ml). Wash with gentle rocking or shaking for 5 minutes. Remove the wash buffer. Repeat one time for a total of two washes for 5 minutes per wash.
23. Finally, wash the glass slide with 30 ml of ddH<sub>2</sub>O for 5 minutes. Remove glass slide and decant water from Centrifuge Tube.
24. Remove buffer droplets from the slide completely by one of the following ways:
  - Put the glass slide into the Slide Washer/Dryer, and dry the glass slide by centrifuge at 1,000 rpm for 3 minutes without cap.
  - Or dry the glass slide by a compressed N2 stream.
  - Or gently apply suction with a pipette to remove buffer droplets. Do not touch the array surface, only the sides.

*Note: Make sure the finished glass slide is completely dry before scanning or storage.*


## E. Fluorescence Detection

25. You may proceed immediately to scanning or you may store the slide at -20 °C in the Centrifuge Tube provided or at RT to scan at a later time.

**Note:** Please protect the finished glass slides from temperatures above RT and store them in the dark. Do not expose glass slide to strong light, such as sunlight or a UV lamp.

**Note:** If you need to repeat any of the incubation steps after finishing the experiment, you must first re-assemble the glass slide into the incubation chamber by following the steps as described below. To avoid breaking the printed glass slide, you may first want to practice assembling the device with a blank glass slide.

1. Apply slide to incubation chamber barcode facing upward (image A).
2. Gently snap one edge of a snap-on side (image B).
3. Gently press other of side against lab bench and push in lengthwise direction (image C).
4. Repeat with the other side (image D)








































## VII. Interpretation of Results:


### A. Explanation of Controls Spots

There are three Positive Controls (POS1, POS2, POS3) in each array. These are three levels of standardized biotinylated IgG. All other variables being equal, the Positive Control intensities will be the same for each sub-array. This allows for normalization based upon the relative fluorescence signal responses to a known control. Some arrays may have beta-actin and GAPDH as internal controls, much as “housekeeping” genes or proteins are used to normalize results in PCR or Western blots, respectively.

### B. Typical Results

The following figure shows the typical result of this array probed with sample(s). The images were captured using an Axon GenePix laser scanner. The Positive control signals in the upper left and lower right corners of each array can be used to identify the orientation and help normalize the results between arrays.





*Note: In the absence of an external standard curve for each protein detected, there is no means of assessing absolute or relative concentrations of different proteins in the same sample using immunoassays. If you wish to obtain quantitative data (i.e., concentrations of the various analytes in your samples), try using our Quantibody® Arrays as a targeted follow-up experiment.*

### C. Background Subtraction

Once you have obtained fluorescence intensity data, you should subtract the background and normalize to the Positive Control signals before proceeding to analysis.

Most laser fluorescence scanners' software has an option to automatically measure the local background around each spot. For best results, we recommend comparing signal intensities representing the MEAN signals minus local background. If your resulting fluorescence signal intensity reports do not include these values (e.g., a column labeled as "F532 Mean - B532"), you may need to subtract the background manually or change the default settings on your scanner's data report menu.

### D. Normalization of Array Data

To normalize signal intensity data, one sub-array is defined as "reference" to which the other arrays are normalized. This choice is arbitrary. For example, in our Analysis Tool Software (described below), the array

represented by data entered in the left-most column each worksheet is the default “reference array.”

You can calculate the normalized values as follows:

$$X(Ny) = X(y) * P1/P(y)$$

Where:

P1 = mean signal intensity of POS spots on reference array

P(y) = mean signal intensity of POS spots on Array "y"

X(y) = mean signal intensity for spot "X" on Array "y"

X(Ny) = normalized signal intensity for spot "X" on Array "y"

The RayBio® Analysis Tool software is freely available for use with data obtained using RayBio® Biotin Label-based Antibody Arrays. You can copy and paste your signal intensity data (with and without background) into the Analysis Tool, and it will automatically normalize signal intensities to the Positive Controls. Analysis Tool software can be downloaded from the product page on the RayBiotech website.

## E. Threshold of Significant Difference

After subtracting background signals and normalization to Positive Controls, comparison of signal intensities between and among array images can be used to determine relative differences in expression levels of each protein between samples or groups.

Any  $\geq 1.5$ -fold increase or  $\leq 0.65$ -fold decrease in signal intensity for a single analyte between samples or groups may be considered a measurable and significant difference in expression, provided that both sets of signals are well above background (Mean background + 2 standard deviations, accuracy  $\approx 95\%$ ).

## F. Pathway Analysis of the Array Proteins

Human antibody array L-8000 detects 8000 unique human proteins, including most analyzed cytokines, chemokines, adipokines, extracellular matrix proteins, growth factors, angiogenic factors, proteases, enzymes, soluble and transmembrane receptors and transport proteins, adhesion molecules and other proteins. All the array proteins are provided with their Uniprot number and GeneID, which are essential for further data mining. Raybiotech offers affordable biostatistics and bioinformatics service, including data clean-up, differential expression analysis, cluster analysis, biomarker selection, pathway analysis and experimental design. See more details on the website: <https://www.raybiotech.com/biostatistics-and-bioinformatics-services>

## VIII. Troubleshooting Guide

<b>Problem</b>	<b>Cause</b>	<b>Recommendation</b>
<b>Weak Signal</b>	Inadequate detection	Increase laser power and PMT parameters
	Inadequate reagent volumes or improper dilution	Check pipettes and ensure correct preparation
	Short incubation time	Ensure sufficient incubation time and change sample incubation step to overnight
	Too low protein concentration in sample	Dilute starting sample less or concentrate sample
	Improper storage of kit	Store kit as suggested temperature. Don't freeze/thaw the slide.
<b>Uneven signal</b>	Bubble formed during incubation	Handle and pipette solutions more gently; De-gas solutions prior to use
	Arrays are not completely covered by reagent	Prepare more reagent and completely cover arrays with solution
	Reagent evaporation	Cover the incubation chamber with adhesive film during incubation
<b>General</b>	Cross-contamination from neighboring wells	Avoid overflowing wash buffer between wells
	Comet tail formation	Air dry the slide for at least 1 hour before usage
	Inadequate detection	Increase laser power so the highest standard concentration for each cytokine receives the highest possible reading yet remains unsaturated
<b>High background</b>	Overexposure	Lower the laser power
	Dark spots	Completely remove wash buffer in each wash step
	Insufficient wash	Increase wash time and use more wash buffer
	Dust	Minimize dust in work environment before starting experiment
	Slide is allowed to dry out	Take additional precautions to prevent slides from drying out during experiment

## IX. Selected References

Christina Scheel et all., *Paracrine and Autocrine Signals Induce and Maintain Mesenchymal and Stem Cell States in the Breast*. Cell. 2011;145, 926–940

Lin Y, Huang R, Chen L, et al., *Profiling of cytokine expression by biotin-labeled-based protein arrays*. Proteomics. 2003, 3: 1750–1757.

Huang R, Jiang W, Yang J, et al., *A Biotin Label-based Antibody Array for High-content Profiling of Protein Expression*. Cancer Genomics Proteomics. 2010; 7(3):129–141.

Liu T, Xue R, Dong L, et al., *Rapid determination of serological cytokine biomarkers for hepatitis B-virus-related hepatocellulare carcinoma using antibody arrays*. Acta Biochim Biophys Sin. 2011; 43(1):45–51.

Cui J, Chen Y, Chou W-C, et al., *An integrated transcriptomic and computational analysis for biomarker identification in gastric cancer*. Nucl Acids Res. 2011; 39(4):1197–1207.

Jun Zhong et all., *Temporal Profiling of the Secretome during Adipogenesis in Humans*. Journal of Proteome Research. 2010, 9, 5228–5238

Chowdury UR, Madden BJ, Charlesworth MC, Fautsch MP., *Proteomic Analysis of Human Aqueous Humor*. Invest Ophthalmol Visual Sci. 2010; 51(10):4921–4931.

Wei Y, Cui C, Lainscak M, et al., *Type-specific dysregulation of matrix metalloproteinases and their tissue inhibitors in end-stage heart failure patients: relationship between MMP-10 and LV remodeling*. J Cell Mol Med. 2011; 15(4):773–782.

Kuranda K, Berthon C, Lepêtre F, et al., *Expression of CD34 in hematopoietic cancer cell lines reflects tightly regulated stem/progenitor-like state*. J Cell Biochem. 2011; 112(5):1277–1285.

Toh HC, Wang W-W, Chia WK, et al., *Clinical Benefit of Allogenic Melanoma Cell Lysate-Pulsed Autologous Dendritic Cell Vaccine in MAGE-Positive Colorectal Cancer Patients*. Clin Chem Res. 2009; 15:7726–7736.

Zhen Hou, *Cytokine array analysis of peritoneal fluid between women with endometriosis of different stages and those without endometriosis*. Biomarkers. 2009;14(8): 604-618.

Yao Liang Tang, et al., *Hypoxic Preconditioning Enhances the Benefit of Cardiac Progenitor Cell Therapy for Treatment of Myocardial Infarction by Inducing CXCR4*. Circ Res. 2009;109:197723

RayBio® L-series Antibody Arrays are patent-pending technology developed by RayBiotech.

This product is intended for research only and is not to be used for clinical diagnosis. Our products may not be resold, modified for resale, or used to manufacture commercial products without written approval by RayBiotech, Inc.


Under no circumstances shall RayBiotech be liable for any damages arising out of the use of the materials.

Products are guaranteed for six months from the date of shipment when handled and stored properly. In the event of any defect in quality or merchantability, RayBiotech's liability to buyer for any claim relating to products shall be limited to replacement or refund of the purchase price.

RayBio® is a registered trademark of RayBiotech, Inc.

GenePix® is a registered trademark of Molecular Devices, Inc.

**This product is for research use only.**



©2021 RayBiotech, Inc.