

## Mouse VEGF ELISA Kit

**Catalog Number:** ELM-VEGF

User Manual

Last Revised: January 22, 2024



Protocol Video

*Please read entire manual carefully before starting experiment.*

### Introduction

The RayBio® Mouse VEGF (or VEGF-A) ELISA kit is an in vitro enzyme-linked immunosorbent assay for the quantitative measurement of mouse VEGF in serum, plasma and cell culture supernatants. This assay employs an antibody specific for mouse VEGF coated on a 96-well plate. Standards and samples are pipetted into the wells and VEGF present in a sample is bound to the wells by the immobilized antibody. The wells are washed and biotinylated anti-mouse VEGF antibody is added. After washing away unbound biotinylated antibody, HRP-conjugated streptavidin is pipetted to the wells. The wells are again washed, a TMB substrate solution is added to the wells and color develops in proportion to the amount of VEGF bound. The Stop Solution changes the color from blue to yellow, and the intensity of the color is measured at 450 nm.

**Short on sample, or need higher sensitivity?** Check out the [IQELISA® Immuno-PCR assay platform](#) and our [Simoa® Single Molecule Protein Detection Services](#).

### Storage

The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C. For prepared reagent storage, see table below.

## Reagents

| Component                      | Size / Description                                                                         | Storage / Stability After Preparation |
|--------------------------------|--------------------------------------------------------------------------------------------|---------------------------------------|
| Mouse VEGF Microplate          | 96 wells (12 strips x 8 wells) coated with anti-Mouse VEGF.                                | 1 month at 4°C*                       |
| Mouse VEGF Standard Protein    | 2 vials of Mouse VEGF. 1 vial is enough to run each standard in duplicate.                 | 1 week at -80°C                       |
| Mouse VEGF Detection Antibody  | 2 vials of biotinylated anti-Mouse VEGF. Each vial is enough to assay half the microplate. | 5 days at 4°C                         |
| Wash Buffer                    | 25 ml of 20X concentrated solution.                                                        | 1 month at 4°C                        |
| HRP-Streptavidin               | 200 µl 160X concentrated HRP-conjugated streptavidin.                                      | Do not store and reuse.               |
| TMB One-Step Substrate Reagent | 12 ml of 3,3,5,5'-tetramethylbenzidine (TMB) in buffer solution.                           | N/A                                   |
| Stop Solution                  | 8 ml of 0.2 M sulfuric acid.                                                               | N/A                                   |
| Assay Diluent A                | 30 ml of diluent buffer, 0.09% sodium azide as preservative.                               | N/A                                   |
| Assay Diluent B                | 15 ml of 5X concentrated buffer.                                                           | 1 month at 4°C                        |

\*Return unused wells to the pouch containing desiccant pack, reseal along entire edge.

## Additional Materials Required


1. Microplate reader capable of measuring absorbance at 450 nm.
2. Precision pipettes to deliver 2 µl to 1 ml volumes.
3. Adjustable 1-25 ml pipettes for reagent preparation.
4. 100 ml and 1 liter graduated cylinders.
5. Absorbent paper.
6. Distilled or deionized water.
7. Log-log graph paper or computer and software for ELISA data analysis.
8. Tubes to prepare standard or sample dilutions.

## Reagent Preparation

1. Bring all reagents and samples to room temperature (18 - 25°C) before use.
2. Assay Diluent B should be diluted 5-fold with deionized or distilled water before use.
3. Sample dilution: Assay Diluent A should be used for dilution of serum and plasma samples. 1X Assay Diluent B should be used for dilution of cell culture supernatant samples. The suggested dilution for normal serum/plasma is 3 fold.

**Note:** Levels of VEGF may vary between different samples. Optimal dilution factors for each sample must be determined by the investigator.

4. Preparation of standard: Briefly spin a vial of standard protein. Add 400 µl Assay Diluent A (for serum/plasma samples) or 1X Assay Diluent B (for cell culture medium) into the Standard Protein vial to prepare a 25 ng/ml standard. Dissolve the powder thoroughly by a gentle mix. Add 40 µl VEGF standard from the vial of Standard Protein, into a tube with 960 µl Assay Diluent A or 1X Assay Diluent B to prepare a 1,000 pg/ml stock standard solution. Pipette 300 µl Assay Diluent A or 1X Assay Diluent B into each tube. Use the stock standard solution to produce a dilution series (shown below). Mix each tube thoroughly before the next transfer. Assay Diluent A or 1X Assay Diluent B serves as the zero standard (0 pg/ml).



|                       |                   | Std1       | Std2      | Std3      | Std4     | Std5       | Std6       | Std7      | Zero Standard |
|-----------------------|-------------------|------------|-----------|-----------|----------|------------|------------|-----------|---------------|
| <b>Diluent volume</b> | Standard + 400 µl | 960 µl     | 300 µl    | 300 µl    | 300 µl   | 300 µl     | 300 µl     | 300 µl    | 300 µl        |
| <b>Conc.</b>          | 25 ng/ml          | 1000 pg/ml | 400 pg/ml | 160 pg/ml | 64 pg/ml | 25.6 pg/ml | 10.2 pg/ml | 4.1 pg/ml | 0 pg/ml       |

5. If the Wash Buffer (20X) contains visible crystals, warm to room temperature and mix gently until dissolved. Dilute 20 ml of Wash Buffer into deionized or distilled water to yield 400 ml of 1X Wash Buffer.
6. Briefly spin the Detection Antibody vial before use. Add 100 µl of 1X Assay Diluent B into the vial to prepare a detection antibody concentrate. Pipette up and down to mix gently (the concentrate can be stored at 4°C for 5 days). The detection antibody concentrate should be diluted 80-fold with 1X Assay Diluent B and used in step 5 of Part VI Assay Procedure.

7. Briefly spin the HRP-Streptavidin concentrate vial and pipette up and down to mix gently before use, as precipitates may form during storage. HRP-Streptavidin concentrate should be diluted 160-fold with 1X Assay Diluent B.

For example: Briefly spin the HRP-Streptavidin vial and pipette up and down to mix gently. Add 50 µl of HRP-Streptavidin concentrate into a tube with 8 ml 1X Assay Diluent B to prepare a 160-fold diluted HRP-Streptavidin solution (don't store the diluted solution for next day use). Mix well.

## Assay Procedure

1. Bring all reagents and samples to room temperature (18 - 25°C) before use. It is recommended that all standards and samples be run at least in duplicate.
2. Label removable 8-well strips as appropriate for your experiment.
3. Add 100 µl of each standard (see Reagent Preparation step 4) and sample into appropriate wells. Cover wells and incubate for 2.5 hours at room temperature with gentle shaking.
4. Discard the solution and wash 4 times with 1X Wash Solution. Wash by filling each well with Wash Buffer (300 µl) using a multi-channel Pipette or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
5. Add 100 µl of 1X prepared biotinylated antibody (Reagent Preparation step 6) to each well. Incubate for 1 hour at room temperature with gentle shaking.
6. Discard the solution. Repeat the wash as in step 4.
7. Add 100 µl of prepared Streptavidin solution (see Reagent Preparation step 7) to each well. Incubate for 45 minutes at room temperature with gentle shaking.
8. Discard the solution. Repeat the wash as in step 4.
9. Add 100 µl of TMB One-Step Substrate Reagent to each well. Incubate for 30 minutes at room temperature in the dark with gentle shaking.
10. Add 50 µl of Stop Solution to each well. Read at 450 nm immediately.

## Assay Procedure Summary


1. Prepare all reagents, samples and standards as instructed.
2. Add 100 µl standard or sample to each well. Incubate 2.5 hours at room temperature.
3. Add 100 µl prepared biotin antibody to each well. Incubate 1 hour at room temperature.
4. Add 100 µl prepared Streptavidin solution. Incubate 45 minutes at room temperature.
5. Add 100 µl TMB One-Step Substrate Reagent to each well. Incubate 30 minutes at room temperature.
6. Add 50 µl Stop Solution to each well. Read at 450 nm immediately.

## Calculation of Results

Calculate the mean absorbance for each set of duplicate standards, controls and samples, and subtract the average zero standard optical density. Plot the standard curve on log-log graph paper or using Sigma plot software, with standard concentration on the x-axis and absorbance on the y-axis. Draw the best-fit straight line through the standard points.

### A. Typical Data

These standard curves are for demonstration only. A standard curve must be run with each assay.



### B. Sensitivity

The minimum detectable dose of Mouse VEGF was determined to be 2 pg/ml.

Minimum detectable dose is defined as the analyte concentration resulting in an absorbance that is 2 standard deviations higher than that of the blank (diluent buffer).

### C. Spiking & Recovery

Recovery was determined by spiking various levels of Mouse VEGF into the sample types listed below. Mean recoveries are as follows:

| Sample Type        | Average % Recovery | Range (%) |
|--------------------|--------------------|-----------|
| Serum              | 87.67              | 79-102    |
| Plasma             | 89.32              | 81-103    |
| Cell culture media | 94.29              | 83-105    |

## D. Linearity

| Sample Type |                                 | Serum        | Plasma       | Cell Culture Media |
|-------------|---------------------------------|--------------|--------------|--------------------|
| 1:2         | Average % of Expected Range (%) | 96<br>83-103 | 97<br>84-103 | 95<br>82-102       |
| 1:4         | Average % of Expected Range (%) | 94<br>82-102 | 98<br>84-104 | 97<br>83-103       |

## E. Reproducibility

Intra-Assay CV%: <10%

Inter-Assay CV%: <12%

## Specificity

This ELISA kit shows no cross-reactivity with any of the cytokines tested: Mouse 6Ckine, CTACK, Eotaxin, GCSF, GM-CSF, IL-2, IL-3, IL-4, IL-5, IL-6, IL-9, IL-10, IL-12p40, IL-12p70, IL-13, IL-17, IFN-gamma, KC, Leptin, MCP-5, MIP-1 alpha, MIP-2, MIP-3 beta, RANTES, SCF, sTNF $\alpha$ , TARC, TIMP-1, TNF-alpha, Tpo.

## Troubleshooting Guide

| Problem             | Cause                                                                                                                                                                                                              | Solution                                                                                                                                                                                                                                                                                                                                                              |
|---------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Poor standard curve | <ul style="list-style-type: none"> <li>• Inaccurate pipetting</li> <li>• Improper standard dilution</li> </ul>                                                                                                     | <ul style="list-style-type: none"> <li>• Check pipettes</li> <li>• Briefly centrifuge the standard protein and dissolve the powder thoroughly by gently mixing</li> </ul>                                                                                                                                                                                             |
| Low signal          | <ul style="list-style-type: none"> <li>• Improper preparation of standard and/or biotinylated antibody</li> <li>• Too brief incubation times</li> <li>• Inadequate reagent volumes or improper dilution</li> </ul> | <ul style="list-style-type: none"> <li>• Briefly spin down vials before opening. Dissolve the powder thoroughly.</li> <li>• Ensure sufficient incubation time. Assay procedure step 3 may be done overnight at 4°C with gentle shaking (note: may increase overall signals including background).</li> <li>• Check pipettes and ensure correct preparation</li> </ul> |
| Large CV            | <ul style="list-style-type: none"> <li>• Inaccurate pipetting</li> <li>• Air bubbles in wells</li> </ul>                                                                                                           | <ul style="list-style-type: none"> <li>• Check pipettes</li> <li>• Remove bubbles in wells</li> </ul>                                                                                                                                                                                                                                                                 |
| High background     | <ul style="list-style-type: none"> <li>• Plate is insufficiently washed</li> <li>• Contaminated wash buffer</li> </ul>                                                                                             | <ul style="list-style-type: none"> <li>• Review the manual for proper wash. If using a plate washer, ensure that all ports are unobstructed.</li> <li>• Make fresh wash buffer</li> </ul>                                                                                                                                                                             |
| Low sensitivity     | <ul style="list-style-type: none"> <li>• Improper storage of the ELISA kit</li> <li>• Stop solution</li> </ul>                                                                                                     | <ul style="list-style-type: none"> <li>• Store your standard at &lt;-70°C after reconstitution, others at 4°C. Keep substrate solution protected from light.</li> <li>• Add stop solution to each well before reading plate</li> </ul>                                                                                                                                |