

Abbexa Ltd, Innovation Centre, Cambridge Science Park, Cambridge, CB4 0EY, UK
Telephone: +44 (0) 1223 755950 - Fax: +44 (0) 1223 755951 - E-Mail: info@abbexa.com

Keratin 19 (KRT19) Antibody Pair

Catalogue No.: abx130434

Keratin 19 (KRT19) Antibody Pair for use in Sandwich ELISA assay development. The 10 × 96 tests size contains:

- 480 µg Keratin 19 (KRT19) rabbit polyclonal capture antibody,
- 50 µg Keratin 19 (KRT19) biotin-conjugated rabbit polyclonal detection antibody,
- 1 µg Keratin (KRT19) standard.

It is recommended to use this antibody pair with [abx098958 Antibody Pair Support Kit \(Sandwich Method\)](#).

Target:	Keratin 19 (KRT19)
Reactivity:	Human
Tested Applications:	ELISA
Recommended dilutions:	Dilute the Capture Antibody 1000-fold with Coating Buffer. Dilute the biotin-bonjugated Detection Antibody 1000-fold with Detection Antibody Diluent. Optimal dilutions/concentrations should be determined by the end user.
Form:	Liquid (Capture Antibody and Detection Antibody)
Reconstitution:	Reconstitute the standard with Standard Diluent. The volume, and therefore standard concentration, should be determined by the end user.
Test Range:	0.312 pg/ml - 20 pg/ml
Storage:	Store at 2 to 8 °C for up to one month. Aliquot and store at -80 °C for up to one year. Avoid repeated freeze/thaw cycles. All solutions should be made fresh before the experiment.
Concentration:	480 µg/0.12 ml (Capture Antibody), 50 µg/0.12 ml (Detection Antibody), 1 µg/vial (Standard)
Standard Form:	Lyophilized
ELISA Type:	Sandwich
Capture Antibody Host:	Rabbit
Detection Antibody Host:	Rabbit
Capture Antibody Clonality:	Polyclonal
Detection Antibody Clonality:	Polyclonal
Capture Antibody Conjugation:	Unconjugated

Abbexa Ltd, Innovation Centre, Cambridge Science Park, Cambridge, CB4 0EY, UK
Telephone: +44 (0) 1223 755950 - Fax: +44 (0) 1223 755951 - E-Mail: info@abbexa.com

Detection Antibody Conjugation: Biotin

Buffer: The capture and detection antibody both contain 0.1% sodium azide.

Directions for use: Bring all components to room temperature (18-25°C) and briefly spin or centrifuge the vials before use. Working solutions should be prepared and used immediately.

Recommended Procedure:

1. Dilute the Capture Antibody to working concentration using Coating Buffer. Immediately coat the 96-well plate with diluted Capture Antibody (100 µl per well). Seal the plate and incubate at 4 °C overnight or at 37 °C for 2 hours
2. Aspirate the wells and wash with Wash Buffer (350 µl per well) and allow to soak for 1-2 min. Remove the liquid by inverting and tapping the plate on to absorbent paper.
3. Block the plate with Blocking Buffer (200 µl per well) at 37 °C for 1.5 hours.
4. Repeat the aspiration/wash process in Step 2.
5. Add 100 µl of standards or sample into the appropriate wells. Cover with a plate sealer and incubate at 37 °C for 1 hour.
6. Repeat the aspiration/wash process in Step 2.
7. Add appropriately diluted biotin-conjugated Detection Antibody (100 µl per well). Cover the plate with a new plate sealer and incubate at 37 °C for 1 hour.
8. Repeat the aspiration/wash process in Step 2.
9. Add appropriately diluted Streptavidin HRP (100 µl per well). Cover the plate with a new plate sealer and incubate at 37 °C for 30 min.
10. Repeat the aspiration/wash process in Step 2.
11. Add Substrate Solution (90 µl per well). Cover the plate with a new plate sealer and incubate at 37 °C for 10-20 min. Keep the plate in the dark and avoid exposure to light.
12. Add Stop Solution (50 µl per well). Tap the side of the plate to ensure thorough mixing.
13. Measure the absorbance immediately using a microplate reader set at 450 nm.

Note: This product is for research use only.