



Abbexa Ltd, Innovation Centre, Cambridge Science Park, Cambridge, CB4 0EY, UK Telephone: +44 (0) 1223 755950 - Fax: +44 (0) 1223 755951 - E-Mail: info@abbexa.com

## Human Insulin Like Growth Factor Binding Protein 7 (IGFBP7) ELISA Kit

Catalogue No.:abx151917



Human Insulin Like Growth Factor Binding Protein 7 (IGFBP7) ELISA Kit is a sandwich ELISA kit for use with Serum, plasma and other biological fluids. This assay has high sensitivity and excellent specificity for detection of Insulin Like Growth Factor Binding Protein 7 (IGFBP7)

No significant cross-reactivity or interference between Insulin Like Growth Factor Binding Protein 7 (IGFBP7) and analogues was observed.

Please note that this kit is also available as a CLIA Kit abx190343.

Target: Insulin Like Growth Factor Binding Protein 7 (IGFBP7)

Reactivity: Human

Tested Applications: ELISA

Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.

Test Range: 0.156 ng/ml - 10 ng/ml

**Sensitivity:** < 0.057 ng/ml

**Validity:** The validity for this kit is 6 months.

**Storage:** Store at 2°C to 8°C upon receipt.

**Stability:** The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within

the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested

that the whole assay is performed by the same user throughout.

Standard Form: Lyophilized



## **DATASHEET**

Abbexa Ltd, Innovation Centre, Cambridge Science Park, Cambridge, CB4 0EY, UK Telephone: +44 (0) 1223 755950 - Fax: +44 (0) 1223 755951 - E-Mail: info@abbexa.com

**ELISA Detection:** Colorimetric

ELISA Type: Sandwich

**ELISA Data:** Quantitative

**Sample Type:** Serum, plasma and other biological fluids.

**Note:** This product is for research use only. The range and sensitivity is subject to change. Please

contact us for the latest product information. For accurate results, sample concentrations must be diluted to mid-range of the kit. If you require a specific range, please contact us in advance or write

your request in your order comments.