



Abbexa Ltd, Innovation Centre, Cambridge Science Park, Cambridge, CB4 0EY, UK Telephone: +44 (0) 1223 755950 - Fax: +44 (0) 1223 755951 - E-Mail: info@abbexa.com

## Human Mucosal Addressin Cell Adhesion Molecule 1 (MAdCAM1) ELISA Kit

Catalogue No.:abx152255



Human MAdCAM1 ELISA Kit is a sandwich ELISA kit for use with Serum, plasma, tissue homogenates, urine and other biological fluids. This assay has high sensitivity and excellent specificity for detection of Mucosal Addressin Cell Adhesion Molecule 1 (MAdCAM1)

No significant cross-reactivity or interference between Mucosal Addressin Cell Adhesion Molecule 1 (MAdCAM1) and analogues was observed.

Please note that this kit is also available as a CLIA Kit abx492835.

Target: Mucosal Addressin Cell Adhesion Molecule 1 (MAdCAM1)

Reactivity: Human

Tested Applications: ELISA

Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.

**Test Range:** 78.12 pg/ml - 5000 pg/ml

Sensitivity: < 37 pg/ml

**Validity:** The validity for this kit is 6 months.

**Storage:** Store at 2°C to 8°C upon receipt.

**Stability:** The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within

the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested

that the whole assay is performed by the same user throughout.

Standard Form: Lyophilized



## **DATASHEET**

Abbexa Ltd, Innovation Centre, Cambridge Science Park, Cambridge, CB4 0EY, UK Telephone: +44 (0) 1223 755950 - Fax: +44 (0) 1223 755951 - E-Mail: info@abbexa.com

**ELISA Detection:** Colorimetric

ELISA Type: Sandwich

**ELISA Data:** Quantitative

**Sample Type:** Serum, plasma, tissue homogenates, urine and other biological fluids.

**Note:** This product is for research use only. The range and sensitivity is subject to change. Please

contact us for the latest product information. For accurate results, sample concentrations must be diluted to mid-range of the kit.If you require a specific range, please contact us in advance or write

your request in your order comments.