

## **Technical Manual**

## **ChromaDazzle Heme Assay Kit**

**Catalogue Code: BA0044**

**Pack Size: 250 assays**

**Research Use Only**

## DESCRIPTION

Heme is one important member of the porphyrin family. It is synthesized in both mitochondria and cytoplasm, and is a key prosthetic group for various essential proteins such as hemoglobin, cytochromes, catalases and peroxidases. Heme determination is widely practiced by researchers of various blood diseases.

Simple, direct and automation-ready procedures for measuring heme concentration are becoming popular in Research and Drug Discovery. The Assay Genie ChromaDazzle Heme Assay Kit is based on an improved aqueous alkaline solution method, in which the heme is converted into a uniform colored form. The intensity of color, measured at 400 nm, is directly proportional to the heme concentration in the sample. The optimized formulation substantially reduces interference by substances in the raw samples and exhibits high sensitivity.

## APPLICATIONS

**Direct Assays:** total heme in blood, serum, plasma, urine, heme-carrying enzymes.

**Pharmacology:** effects of drugs on heme metabolism.

**Drug Discovery:** HTS for drugs that modulate heme levels.

## KEY FEATURES

**Sensitive and accurate.** Linear detection range 0.6 – 125  $\mu$ M heme in 96-well plate assay.

**Simple and high-throughput.** The “mix-and-read” procedure involves addition of a single working reagent and reading the optical density. Can be readily automated as a high-throughput assay in 96-well plates for thousands of samples per day.

**Safety.** Reagents are non-toxic.

**Versatility.** Assays can be executed in 96-well plate or cuvette.

## KIT CONTENTS (250 tests in 96-well plates)

**Reagent:** 50 mL

**Calibrator:** 10 mL (equivalent to 62.5  $\mu$ M heme)

**Storage conditions.** The kit is shipped room temperature. Store reagent and Calibrator at 4°C. Shelf life: 12 months after receipt.

**Precautions:** reagents are for research use only. Normal precautions for laboratory reagents should be exercised while using the reagents. Please refer to Material Safety Data Sheet for detailed information.

## PROCEDURES

### Procedure using 96-well plate:

1. *Blank and Calibrator.* Pipette 50  $\mu$ L water (Blank) and 50  $\mu$ L Calibrator into wells of a clear bottom 96-well plate. Transfer 200  $\mu$ L water into the blank and Calibrator wells. The diluted calibrator is equivalent to 62.5  $\mu$ M heme.

2. *Samples.* Serum and plasma samples can be assayed directly ( $n = 1$ ). Blood samples should be diluted 100-fold in distilled water ( $n = 100$ ).

Transfer 50  $\mu$ L samples into wells (*important:* avoid bubble formation during the pipetting steps). Add 200  $\mu$ L Reagent to sample wells and tap plate lightly to mix.

3. Incubate 5 min at room temperature. Read OD at 380-420nm (peak 400nm).

### Procedure using cuvette:

1. Transfer 100  $\mu$ L sample and 1000  $\mu$ L Reagent into a cuvet and tap lightly to mix. Read OD at 380-420nm (peak 400 nm) against water.

2. Transfer 100  $\mu$ L Calibrator and 1000 $\mu$ L water to cuvet. Read OD at 400nm against water.

### CALCULATION

Subtract blank OD (water) from the Calibrator and Sample OD values. The total heme concentration of Sample is calculated as

$$= \frac{OD_{\text{SAMPLE}} - OD_{\text{BLANK}}}{OD_{\text{CALIBRATOR}} - OD_{\text{BLANK}}} \times 62.5 \times n \text{ (}\mu\text{M)}$$

OD<sub>SAMPLE</sub>, OD<sub>CALIBRATOR</sub> and OD<sub>BLANK</sub> are OD values of the sample, the Calibrator and water. *n* is the dilution factor (100 for blood samples).

**Conversions:** 1mg/dL heme equals 15.3  $\mu$ M, 0.001% or 10 ppm.

### MATERIALS REQUIRED, BUT NOT PROVIDED

Pipeting devices and accessories.

#### Procedure using 96-well plate:


Clear-bottom 96-well plates (e.g. Corning Costar) and plate reader.

#### Procedure using cuvette:

Cuvets and spectrophotometer.

### EXAMPLES

Heme was determined using the 96-well plate protocol. The values were  $27.3 \pm 0.2 \mu\text{M}$  for rat serum,  $7.8 \pm 0.4 \mu\text{M}$  for human plasma and  $11.2 \pm 0.2 \text{ mM}$  for a mouse whole blood sample.



**Standard Curve with Freshly Prepared Heme**  
in 96-well plate assay

### PUBLICATIONS

1. Mingone, C.J. et al (2006). Protoporphyrin IX generation from delta-aminolevulinic acid elicits pulmonary artery relaxation and soluble guanylate cyclase activation. *Am J Physiol Lung Cell Mol Physiol* 291: L337–L344.
2. Pamplona, A. et al (2007). Heme oxygenase-1 and carbon monoxide suppress the pathogenesis of experimental cerebral malaria. *Nature Med.* 13(6): 703-710.
3. Raghuram, S. et al (2007). Identification of heme as the ligand for the orphan nuclear receptors REV-ERB $\alpha$  and REV-ERB $\beta$ . *Nat Struct Mol Biol.* 14(12):1207-13.

**Contact Details**

**Dublin, Ireland**

**Email:** [info@assaygenie.com](mailto:info@assaygenie.com)

**Web:** [www.assaygenie.com](http://www.assaygenie.com)

**Technical Support:** [Techsupport@assaygenie.com](mailto:Techsupport@assaygenie.com)