

REFERENCES

1. Emlen W; O'Neill L Clinical significance of antinuclear antibodies: comparison of detection with immunofluorescence and enzyme-linked immunosorbent assays. *Arthritis Rheum* 1997;40(9):1612-8.
2. Gonz'alez C; Martin T; Arroyo T; Garc'ia-Isidoro M; Navajo JA; Gonz'alez-Buitrago JM. Comparison and variation of different methodologies for the detection of autoantibodies to nuclear antigens (ANA). *J Clin Lab Anal* 1997;11(6):388-92.
3. Parveen S; Morshed SA; Nishioka M. High prevalence of antibodies to recombinant CENP-B in primary biliary cirrhosis: nuclear immunofluorescence patterns and ELISA reactivities. *J Gastroenterol Hepatol* 1995;10(4):438-45.
4. Welin Henriksson E; Hansson H; Karlsson-Parra A; Pettersson I. Autoantibody profiles in canine ANA-positive sera investigated by immunoblot and ELISA. *Vet Immunol Immunopathol* 1998;61(2-4):157-70.
5. Koh WH; Dunphy J; Whyte J; Dixey J; McHugh NJ. Characterisation of anticytoplasmic antibodies and their clinical associations [see comments]. *Ann Rheum Dis* 1995;54(4):269-73.
6. Spronk PE; Bootsma H; Horst G; Huitema MG; Limburg PC; Cohen Tervaert JW; Kallenberg CG. Antineutrophil cytoplasmic antibodies in systemic lupus erythematosus. *Br J Rheumatol* 1996;35(7):625-31.

2017-04-13

Cat#: AN033G (96 Tests)
 For Order and Inquiries, please contact
 Calbiotech Inc.,
 1935 Cordell Ct., El Cajon, CA 92020
 Tel (619) 660-6162, Fax (619) 660-6970,
www.calbiotech.com



ANA Screen ELISA

Catalog No.: AN033G (96 Tests)

INTENDED USE

The Calbiotech Inc. (CBI) ANA Screen ELISA test system is an enzyme-linked immunosorbent assay (ELISA) for the detection of IgG class antibodies to ANA in human serum or plasma.

SUMMARY AND EXPLANATION

Antinuclear antibodies (ANA) are frequently present in patients with systemic lupus erythematosus (SLE) and, less commonly, in other autoimmune diseases Rheumatoid arthritis, Collagen vascular diseases, chronic liver diseases and systemic sclerosis (scleroderma). ANA bind to several nuclear antigens including DsDNA, SSDNA, RNP, Sm, SSA and SSB. ANA frequency increases with age in apparently healthy people, especially women after the age of 45 years. ANA ELISA is widely used as a screening procedure for different autoimmune diseases.

PRINCIPLE OF THE TEST

Diluted patient serum is added to wells coated with purified nuclear antigens. ANA IgG specific antibody, if present, binds to the antigen. All unbound materials are washed away and the enzyme conjugate is added to bind to the antibody-antigen complex, if present. Excess enzyme conjugate is washed off and substrate is added. The plate is incubated to allow the hydrolysis of the substrate by the enzyme. The intensity of the color generated is proportional to the amount of IgG specific antibody in the sample.

MATERIALS PROVIDED	96 Tests
1. Microwells coated with nuclear antigens	12x8x1
2. Sample Diluent: 1 bottle (ready to use)	22 ml
3. Calibrator 1 Vial (ready to use)	1ml
4. Positive Control 1 vial (ready to use)	1ml
5. Negative Control 1 vial (ready to use)	1ml
6. Enzyme conjugate: 1 bottle (ready to use)	12ml
7. TMB Substrate: 1 bottle (ready to use)	12ml
8. Stop Solution: 1 bottle (ready to use)	12ml
9. Wash concentrate 20X: 1 bottle	25ml

MATERIALS NOT PROVIDED

1. Distilled or deionized water.
2. Precision pipettes.
3. Disposable pipette tips.
4. ELISA reader capable of reading absorbance at 450 nm.
5. Absorbance paper or paper towel.

STORAGE AND STABILITY

1. Store the kit at 2 – 8° C.
2. Keep microwells sealed in a dry bag with desiccants.
3. The reagents are stable until expiration of the kit.
4. Do not expose test reagents to heat, sun or strong light.

WARNINGS AND PRECAUTIONS

1. For Research Use Only. Not for use in diagnostic procedures.
2. For Laboratory use.
3. Not for Internal or External Use in Humans or Animals.
4. There should be no eating or drinking within work area.
5. Always wear gloves and a protective lab coat.
6. No pipetting should be done by mouth. Handle all specimens and reagents as potentially infectious and biohazardous.
7. Do not add sodium azide to samples as preservative.
8. Do not use external controls containing sodium azide.
9. Use disposable pipette tips to avoid contaminating chromogenic substrate reagent. Discard reagent if it turns blue.
10. Do not pour chromogenic substrate back into container after use.
11. Do not freeze reagents.
12. Do not mix reagents from different kit lot numbers.
13. Keep reagents out of direct sunlight.
14. Handle stop reagent with care, since it is corrosive.
15. Bring all reagents to room temperature.
16. Viscous forensic samples should always be diluted in phosphate buffered saline or distilled water prior to pipetting.
17. Ensure the bag containing the micro-plate strips and desiccant is sealed well, if only a few strips are used.

SPECIMEN COLLECTION AND HANDLING

1. Collect blood specimens and separate the serum.
2. Typically, specimens may be refrigerated at 2–8 °C for up to seven days or frozen for up to six months. Avoid repetitive freezing and thawing.

REAGENT PREPARATION

Prepare 1X Wash buffer by adding the contents of the bottle (25 ml, 20X) to 475 ml of distilled or deionized water. Store at room temperature (20-25°C).

PREPARATION FOR ASSAY

Bring all specimens and kit reagents to room temperature (20-25°C) and gently mix.

ASSAY PROCEDURE

1. Place the desired number of coated strips into the holder.
2. Negative control, positive control, and calibrator are ready to use. Prepare 1:21 dilution of test samples, by adding 10 µl of the sample to 200 µl of sample diluent. Mix well.

3. Dispense 100 µl of diluted sera, calibrator and controls into the appropriate wells. For the reagent blank, dispense 100µl sample diluent in 1A well position. Tap the holder to remove air bubbles from the liquid and mix well. Incubate for 20 minutes at room temperature.
4. Remove liquid from all wells. Wash wells three times with 300 µL of 1X wash buffer. Blot on absorbance paper or paper towel.
5. Dispense 100 µl of enzyme conjugate to each well and incubate for 20 minutes at room temperature.
6. Remove enzyme conjugate from all wells. Wash wells three times with 300 µl of 1X wash buffer. Blot on absorbance paper or paper towel
7. Dispense 100 µl of TMB substrate and incubate for 10 minutes at room temperature.
8. Add 100 µl of stop solution.
9. Read O.D. at 450 nm using ELISA reader within 15 min. A dual wavelength is recommended with reference filter of 600-650 nm.

CALCULATION OF RESULTS

1. Check Calibrator Factor (CF) value on the calibrator bottle. This value might vary from lot to lot. Make sure you check the value on every kit.
2. Calculate the cut-off value: Calibrator OD x Calibrator Factor (CF).
3. Calculate the Ab (Antibody) Index of each determination by dividing the O.D. value of each sample by cut-off value.