

## Human IL-12 p70 High Sensitivity ELISA KIT

**Catalog No:** CDK039A  
CDK039B

**Quantity:** 1 x 96 tests  
2 x 96 tests

**SPECIFICITY :** Recognizes both natural and recombinant human IL-12 specific to p70 heterodimer

**RANGE :** 0.78 pg / ml - 25 pg / ml

**SENSITIVITY :** < 0.75 pg / ml

**INCUBATION :** 4 h 00

**SAMPLE TYPES :**  
Serum  
Plasma  
Cell culture supernatant

**SAMPLE SIZE :** 100 µl

**CROSS REACTION :** No cross reactivity with other human cytokines

**KIT CONTENT :** Pre-coated 12 strip plates, Biotinylated Secondary Antibody, Amplifier, Standard, Buffers, Streptavidin-HRP, TMB, Stop Reagent

### 1. INTENDED USE

The Human high sensitivity IL-12 ELISA is to be used for the *in-vitro* quantitative determination interleukin-12 (IL-12) p70 heterodimer in human serum, plasma, buffered solutions or cell culture medium. The assay will recognize both natural and recombinant human IL-12. **This kit has been configured for research use only.**

### 2. PRINCIPAL OF THE METHOD

The IL-12 Kit is a solid phase sandwich Enzyme Linked-Immuno-Sorbent Assay (ELISA). A monoclonal antibody specific for IL-12 has been coated onto the wells of the microtiter strips provided. Samples, including standards of known IL-12 concentrations and unknowns are pipetted into these wells.

During the first incubation, the IL-12 antigen and a biotinylated monoclonal antibody specific for IL-12 are simultaneously incubated.

After washing, the enzyme (streptavidin-peroxydase) is added. The entire unbound enzyme is removed by washing and the first amplification step is performed by adding the **Biotin-Tyramine Reagent**. Under the action of HRP, a **biotin polymerization reaction** occurs in the region of the HRP linked to the detection antibody. After washing, the second amplification step is performed and the polymerized biotin is revealed by a new streptavidin-HRP step. Finally, after washing, the substrate is added. The intensity of this colored product is directly proportional to the concentration of IL-12 present in the samples.



**Cell Sciences, Inc.**  
65 Parker Street  
Unit 11  
Newburyport, MA 01950

Toll Free: 888-769-1246  
Phone: 978-572-1070  
Fax: 978-992-0298

E-mail: [info@cellsciences.com](mailto:info@cellsciences.com)  
Web Site: [www.cellsciences.com](http://www.cellsciences.com)

### 3. REAGENTS PROVIDED AND RECONSTITUTION

Reagents (Store @2-8°C)	Quantity 1 x 96 well kit	Reconstitution
P: 96 well microtiter strip plate	1	Ready to use (Pre-coated)
A: Plastic plate covers	2	n/a
B: Standard: 25 pg/ml	2	Reconstitute as directed on the vial (see Assay preparation, section 8).
C: Standard Diluent (Buffer)	1 (25 ml)	10x Concentrate, dilute in distilled water (see reagent preparation, section 8).
D: Standard Diluent: Human Serum	1 (7 ml)	Ready to use.
E: Biotinylated anti-IL-12	1 (0.4 ml)	Dilute in biotinylated antibody diluent (see Assay preparation, section 8).
F: Biotinylated Antibody diluents	1 (7.5 ml)	Ready to use.
G: Streptavidin-HRP	2 (5 µl)	Add 0.5 ml of HRP diluent prior to use (see Assay preparation, section 8).
H: Amplification Diluent	1 (25 ml)	Ready to use.
I: Amplifier*	1 (80 µl)	Dilute in Amplification Buffer.
J: HRP Diluent	1 (25 ml)	Ready to use.
K: Wash Buffer	1 (10 ml)	200x Concentrate dilute in distilled water (see Assay preparation, section 8).
L: TMB Substrate	1 (11 ml)	Ready to use.
M: H <sub>2</sub> SO <sub>4</sub> stop reagent	1 (11 ml)	Ready to use.

\*Reagent contains ethyl alcohol

### 4. MATERIALS REQUIRED BUT NOT PROVIDED

- Microtiter plate reader fitted with appropriate filters (450 nm required with optional 630 nm reference filter)
- Microplate washer or wash bottle
- 10, 50, 100, 200 and 1,000 µl adjustable single channel micropipettes with disposable tips
- 50-300 µl multi-channel micropipette with disposable tips
- Multichannel micropipette reagent reservoirs
- Distilled water
- Vortex mixer
- Orbital shaker
- Miscellaneous laboratory plastic and/or glass, if possible sterile

### 5. STORAGE INSTRUCTIONS

Store the kit reagents between 2 and 8°C. Immediately after use, remaining reagents should be returned to cold storage (2-8°C). Expiry of the kit and reagents is stated on box front labels. The expiry of the kit components can only be guaranteed if the components are stored properly, and if, in case of repeated use of one component, the reagent is not contaminated by the first handling.

**Wash Buffer:** Once prepared, store at 2-8°C for up to 1 week.

**Standard Diluent Buffer:** Once prepared, store at 2-8°C for up to 1 week.



Cell Sciences, Inc.  
65 Parker Street  
Unit 11  
Newburyport, MA 01950

Toll Free: 888-769-1246  
Phone: 978-572-1070  
Fax: 978-992-0298

E-mail: [info@cellsciences.com](mailto:info@cellsciences.com)  
Web Site: [www.cellsciences.com](http://www.cellsciences.com)

**Standards:** Once prepared, use immediately and do not store.

**Biotinylated Secondary Antibody:** Once prepared, use immediately and do not store.

**Streptavidin-HRP:** Once prepared, use immediately and do not store.

## 6. SPECIMEN COLLECTION, PROCESSING AND STORAGE

Cell culture supernatants, serum, plasma or other biological samples will be suitable for use in the assay. Remove serum from the clot or red cells, respectively, as soon as possible after clotting and separation.

**Cell culture supernatants:** Remove particulates and aggregates by spinning at approximately 1000 x g for 10 min.

**Serum:** Avoid any unintentional stimulation of the cells by the procedure. Use pyrogen/endotoxin free collecting tubes. Serum should be removed rapidly and carefully from the red cells after clotting. For that, after clotting, centrifuge at approximately 1000 x g for 10 min and remove serum.

**Plasma:** EDTA, citrate and heparin plasma can be assayed. Spin samples at 1000 x g for 30 min to remove particulates. Harvest plasma.

**Storage:** If not analyzed shortly after collection, samples should be aliquoted (250-500 µl) to avoid repeated freeze-thaw cycles and stored frozen at -80°C. Avoid multiple freeze-thaw cycles of frozen specimens.

**Recommendation:** Do not thaw by heating at 37°C or 56°C. Thaw at room temperature and make sure that sample is completely thawed and homogeneous before use. When possible, avoid use of badly hemolyzed or lipemic sera. If large amounts of particles are present, these should be removed prior to use by centrifugation or filtration.

## 7. SAFETY AND PRECAUTIONS FOR USE

- Handling of reagents, serum or plasma specimens should be in accordance with local safety procedures, e.g. CDC/NIH Health manual: "Biosafety in Microbiological and Biomedical Laboratories" 1984.
- The human serum included in this kit have been tested and found non-reactive for HBsAg, anti HIV1 & 2 and anti VHC. Nevertheless, no known method can offer complete assurance that human blood derivatives will not transmit hepatitis, AIDS or other infections. Therefore handling of reagents, serum or plasma specimens should be in accordance with local safety procedures.
- Laboratory gloves should be worn at all times.
- Avoid any skin contact with H<sub>2</sub>SO<sub>4</sub> and TMB. In case of contact, wash thoroughly with water.
- Do not eat, drink, smoke or apply cosmetics where kit reagents are used.
- Do not pipette by mouth.
- When not in use, kit components should be stored refrigerated or frozen as indicated on vials or bottles labels
- All reagents should be warmed to room temperature before use. Lyophilized standards should be discarded after use.
- Once the desired number of strips has been removed, immediately reseal the bag to protect the remaining strips from deterioration.
- Cover or cap all reagents when not in use.



- Do not mix or interchange reagents between different lots.
- Do not use reagents beyond the expiration date of the kit.
- Use a clean disposable plastic pipette tip for each reagent, standard, or specimen addition in order to avoid cross contamination, for the dispensing of H<sub>2</sub>SO<sub>4</sub> and substrate solution, avoid pipettes with metal parts.
- Use a clean plastic container to prepare the washing solution.
- Thoroughly mix the reagents and samples before use by agitation or swirling.
- All residual washing liquid must be drained from the wells by efficient aspiration or by decantation followed by tapping the plate forcefully on absorbent paper. Never insert absorbent paper directly into the wells
- The TMB solution is light sensitive. Avoid prolonged exposure to light. Also, avoid contact of the TMB solution with metal to prevent color development. Warning TMB is toxic avoid direct contact with hands. Dispose of properly.
- If a dark blue color develops within a few minutes after preparation, this indicates that the TMB solution has been contaminated and must be discarded. Read absorbances within 1 hour after completion of the assay.
- When pipetting reagents, maintain a consistent order of addition from well-to-well. This will ensure equal incubation times for all wells.
- Follow incubation times described in the assay procedure.
- Dispense the TMB solution within 15 min of the washing of the microtiter plate.

## 8. ASSAY PREPARATION

**Bring all reagents to room temperature before use.**

### 8.1. ASSAY DESIGN

Determine the number of microwell strips required to test the desired number of samples, plus appropriate number of wells needed for running zeros and standards. Each sample, standard and zero should be tested **in duplicate**. Remove sufficient Microwell Strips for testing from the pouch immediately prior to use. Return any wells not required for this assay with desiccant to the pouch. Seal tightly and return to 2-8°C storage.

**Example plate layout** (example shown for a 6 point standard curve)

Standards (pg/mL)		Sample Wells											
		1	2	3	4	5	6	7	8	9	10	11	12
A	25	25											
B	12.5	12.5											
C	6.25	6.25											
D	3.12	3.12											
E	1.56	1.56											
F	0.78	0.78											
G	zero	zero											
H													

*All remaining empty wells can be used to test samples in duplicate.*



## 8.2. Preparation of Wash Buffer

Dilute the (200x) wash buffer concentrate 200 fold with distilled water to give a 1x working solution. Pour entire contents (10 ml) of the Washing Buffer Concentrate into a clean 2,000 ml graduated cylinder. Bring final volume to 2,000 ml with glass-distilled or deionized water. Mix gently to avoid foaming. Transfer to a clean wash bottle and store at 2°-8°C for up to 1 week.

## 8.3. Preparation of Standard Diluent Buffer

Add the contents of the vial (10x concentrate) to 225 ml of distilled water before use.

This solution can be stored at 2-8°C for up to 1 week.

## 8.4. Preparation of Standard

Depending on the type of samples you are assaying, the kit includes two standard diluents. Because biological fluids might contain proteases or cytokine-binding proteins that could modify the recognition of the cytokine you want to measure, you should reconstitute standard vials with the most appropriate standard diluent.

For serum and plasma samples, use Standard Diluent Human Serum and for cells culture supernatants use Standard Diluent Buffer.

Standard vials must be reconstituted with the volume of Standard Diluent shown on the vial immediately prior to use. This reconstitution gives a stock solution of 25 pg/ml of IL-12. **Mix the reconstituted standard gently by inversion only.** Serial dilutions of the standard are made directly in the assay plate to provide the concentration range from 25 to 0.78 pg/ml. A fresh standard curve should be produced for each new assay.

- Immediately after reconstitution, add 200 µl of the reconstituted standard to wells A1 and A2, which provides the highest concentration standard at 25 pg/ml.
- Add 100 µl of appropriate Standard Diluent to the remaining standard wells B1 and B2 to F1 and F2.
- Transfer 100 µl from wells A1 and A2 to B1 and B2. Mix the well contents by repeated aspirations and ejections taking care not to scratch the inner surface of the wells.
- Continue this 1:1 dilution using 100 µl from wells B1 and B2 through to wells F1 and F2 providing a serial diluted standard curve ranging from 25 pg/ml to 0.78 pg/ml.
- Discard 100 µl from the final wells of the standard curve (F1 and F2).

Alternatively these dilutions can be performed in separate clean tubes and immediately transferred directly into the relevant wells.



Cell Sciences, Inc.  
65 Parker Street  
Unit 11  
Newburyport, MA 01950

Toll Free: 888-769-1246  
Phone: 978-572-1070  
Fax: 978-992-0298

E-mail: [info@cellsciences.com](mailto:info@cellsciences.com)  
Web Site: [www.cellsciences.com](http://www.cellsciences.com)

## 8.5. Preparation of Biotinylated anti-IL-12

It is recommended this reagent is prepared immediately before use. Dilute the biotinylated anti-IL-10 with the biotinylated antibody diluent in an appropriate clean glass vials. Use volumes appropriate to the number of required wells. Please see example volumes below:

Number of wells required	Biotinylated Antibody (µl)	Biotinylated Antibody Diluent (µl)
16	40	1060
32	80	2120
48	120	3180
96	240	6360

## 8.6. Preparation of Streptavidin-HRP Solution

It is recommended to centrifuge the vial for a few seconds in a microcentrifuge to collect all the volume at the bottom.

Dilute the 5 µl vial with 0.5 ml of HRP diluent immediately before use. This pre-dilution will be used for Step 5 and step 11. Do-not keep this diluted vial for future experiments. Further dilute the HRP solution to volumes appropriate for the number of required wells in a clean glass vial. Please see example volumes below:

Number of wells required	Streptavidin-HRP solution	
	Streptavidin-HRP (µl)	Streptavidin-HRP Diluent (ml)
16	13	1.95
24	25	3.75
32	26	3.9
48	50	7.5
96	80	12

## 8.7. Preparation of Amplifier

It is recommended this reagent is prepared immediately before use. Dilute the Amplifier with the Amplification diluent in an appropriate clean glass vial using volumes appropriate to the number of required wells. Please see example volumes below:

Number of wells required	Amplifier (µl)	Amplification Diluent (ml)
16 to 32	10	4
48	14	5.6
96	30	12



## 9. Method

We strongly recommend that every vial is mixed thoroughly without foaming prior to use except the standard vial which must be mixed gently by inversion only.

Prepare all reagents as shown in section 8.

Note: Final preparation of Biotinylated anti-IL-12 (section 8.5) and Streptavidin-HRP (section 8.6) should occur immediately before use.

Assay Step		Details
1.	Addition	Add 100 $\mu$ l of each <b>standard, sample and zero</b> (appropriate Standard Diluent buffer) in duplicate to appropriate number of wells.
2.	Addition	Add 50 $\mu$ l of diluted <b>biotinylated anti-IL-12</b> to all wells.
3.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) with slow shaking for <b>3 hours</b> .
4.	Wash	Remove the cover and wash the plate as follows: a) Aspirate the liquid from each well. b) Dispense 0.3 ml of <b>1x washing solution</b> into each well. c) Aspirate the contents of each well. d) Repeat step b and c another two times.
5.	Addition	Add 100 $\mu$ l of <b>Streptavidin-HRP solution</b> into all wells.
6.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) with slow shaking for <b>20 min</b> .
7.	Wash	Repeat wash step 4.
8.	Addition	Add 100 $\mu$ l of diluted <b>Amplifier</b> to all wells.
9.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) with slow shaking for <b>15 min</b> .
10.	Wash	Repeat wash step 4.
11.	Addition	Add 100 $\mu$ l of <b>Streptavidin-HRP solution</b> into all wells.
12.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) with slow shaking for <b>20 min</b> .
13.	Wash	Repeat wash step 4.
14.	Addition	Add 100 $\mu$ l of ready-to-use <b>TMB Substrate Solution</b> into all wells.



15.	Incubation	Incubate in the dark for <b>10-20 minutes*</b> at room temperature. Avoid direct exposure to light by wrapping the plate in aluminium foil.
16.	Addition	Add 100 $\mu$ l of <b>H<sub>2</sub>SO<sub>4</sub>: Stop Reagent</b> into all wells.

**Read the absorbance** value of each well (immediately after step 11.) on a spectrophotometer using 450 nm as the primary wavelength and optionally 630 nm as the reference wave length (610 nm to 650 nm is acceptable).

*\*Incubation time of the substrate solution is usually determined by the ELISA reader performance. Many ELISA readers only record absorbance up to 2.0 O.D. Therefore the color development within individual microwells must be observed by the analyst, and the substrate reaction stopped before positive wells are no longer within recordable range.*

**Note:** In case of incubation without shaking the O.D values may be lower than with shaking; in this case let the color develop longer in order to obtain correct OD values.

## 10. Data Analysis


Calculate the average absorbance values for each set of duplicate standards and samples. Ideally duplicates should be within 20% of the mean.

Generate a linear standard curve by plotting the average absorbance of each standard on the vertical axis versus the corresponding Human IL-12 standard concentration on the horizontal axis.

The amount of IL-12 in each sample is determined by extrapolating OD values against IL-12 standard concentrations using the standard curve.

Standard	IL-12 Conc pg/ml	OD (450nm) Mean	CV (%)
1	25	1.628	4.3
2	12.5	0.939	1.7
3	6.25	0.465	8.8
4	3.12	0.292	9.2
5	1.56	0.187	1.9
6	0.78	0.135	0.0
Zero	0	0.076	-





**Note:** curve shown above should not be used to determine results. Every laboratory must produce a standard curve for each set of microwell strips assayed.

## 11. Assay limitations

Do not extrapolate the standard curve beyond the maximum standard curve point. The dose-response is non-linear in this region and good accuracy is difficult to obtain. Concentrated samples above the maximum standard concentration must be diluted with Standard diluent or with your own sample buffer to produce an OD value within the range of the standard curve. Following analysis of such samples, always multiply results by the appropriate dilution factor to produce actual final concentration.

The influence of various drugs on end results has not been investigated. Bacterial or fungal contamination and laboratory cross-contamination may also cause irregular results.

Improper or insufficient washing at any stage of the procedure will result in either false positive or false negative results. Completely empty wells before dispensing fresh Washing Buffer, fill with Washing Buffer as indicated for each wash cycle and do not allow wells to sit uncovered or dry for extended periods.

Disposable pipette tips, flasks or glassware are preferred, reusable glassware must be washed and thoroughly rinsed of all detergents before use.

As with most biological assays conditions may vary from assay to assay therefore **a fresh standard curve must be prepared and run for every assay.**



## 12. Performance Characteristics

### 12.1 Sensitivity

The minimum detectable dose of IL-12 is less than **0.75 pg/ml**.

This has been determined by adding 3 standard deviations to the mean optical density obtained when the zero standard was assayed 32 times.

### 12.2 Precision

#### Intra-Assay

Sample	n	Mean (pg/mL)	SD	CV %	Sample	n	Mean (pg/mL)	SD	CV %
A	8	16.1	1.25	7.8	A	20	16	1.2	7.9
B	8	6.1	0.4	6.6	B	20	6	0.6	9.5

### 12.3 Dilution Parallelism

A pooled human serum containing 25 pg/ml of measured IL-12 was serially diluted in standard buffer diluent over the range of the assay. Linear regression of samples versus the expected concentration yielded a quote slope of 0.99.

### 12.4 Spike Recovery

Recovery of IL-12 added to pooled normal serum was 89% for IL-12 concentration ranging from 25 to 1.56 pg/ml.



## 13. Assay Summary

Total procedure length: 4 h 00 min

<b>Specific molecule detection steps</b>	Add 100 $\mu$ l of sample or diluted standard ↓ Add 50 $\mu$ l of diluted biotinylated detection antibody to all wells ↓ Incubate 3 hours at room temperature with slow shaking ↓
<b>Streptavidin -HRP / Amplification steps</b>	Wash three times ↓ Add 100 $\mu$ l of streptavidin-HRP to all wells ↓ Incubate 20 min at room temperature with slow shaking ↓ Wash three times ↓ Add 100 $\mu$ l Amplifier to all wells ↓ Incubate 15 min at room temperature with slow shaking ↓ Wash three times ↓ Add 100 $\mu$ l of streptavidin-HRP to all wells ↓ Incubate 20 min at room temperature with slow shaking ↓
<b>Revelation and reading steps</b>	Wash three times ↓ Add 100 $\mu$ l of ready-to-use TMB Protect from light. Let the color develop for around 10-20 min. ↓ Add 100 $\mu$ l $H_2SO_4$ ↓ Read Absorbance at 450 nm

NOT FOR HUMAN USE. FOR RESEARCH ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.



Cell Sciences, Inc.  
 65 Parker Street  
 Unit 11  
 Newburyport, MA 01950

Toll Free: 888-769-1246  
 Phone: 978-572-1070  
 Fax: 978-992-0298

E-mail: [info@cellsciences.com](mailto:info@cellsciences.com)  
 Web Site: [www.cellsciences.com](http://www.cellsciences.com)