

Human IL-13 ELISPOT Kit, Pre-coated Plates

Catalog No.: CDK155A
CDK155B

Quantity: 1 x 96-well plates
5 x 96-well plates

Lot No.: TBD

Exp. Date: TBD

NOTE: This sample protocol is subject to variation by Lot Number. Refer to the protocol inserted in your package for the current lot number specifications and expiration date or contact our technical support at tech@cellsciences.com

1. INTENDED USE

The Cell Sciences® ELISPOT assay is a highly specific immunoassay for the analysis of cytokine and other soluble molecule production and secretion from T-cells at a single cell level in conditions closely comparable to the in-vivo environment with minimal cell manipulation. Cell secreted cytokines or soluble molecules are captured by coated antibodies avoiding diffusion in supernatant, protease degradation or binding on soluble membrane receptors. After cell removal, the captured cytokines are revealed by tracer antibodies and appropriate conjugates. This technique is designed to determine the frequency of cytokine producing cells under a given stimulation and the comparison of such frequency against a specific treatment or pathological state. The ELISPOT assay constitutes an ideal tool in the investigation of Th1 / Th2 responses, vaccine development, viral infection monitoring and treatment, oncology, infectious disease, autoimmune diseases and transplantation.

Utilizing sandwich immuno-enzyme technology, ELISPOT assays can detect both secreted cytokines and single cells that simultaneously produce multiple cytokines. Cell secreted cytokines or soluble molecules are captured by coated antibodies avoiding diffusion in supernatant, protease degradation or binding on soluble receptors. After cell removal, the captured cytokines are revealed by detection antibodies and appropriate conjugates.

Interleukin-13 (IL-13) is a Th2-type cytokine, secreted from CD4(+) T cells, mast cells, basophils and eosinophils. The IL-13 gene generates a cluster with other Th2-type cytokines such as IL-4 and IL-5. Although the homology between IL-13 and IL-4 at the amino acid level is low, the IL-13 structure determined by NMR is very similar to that of IL-4. Both cytokines share their receptors and signal pathways, giving these two cytokines similar biological properties.

IL-13 is an immunoregulatory cytokine. It has become evident that IL-13 is a key mediator in the pathogenesis of allergic inflammation. IL-13 mediates its effects by interacting with a complex receptor system comprised of IL-4R alpha and two IL-13 binding proteins, IL-13R alpha-1 and IL-13R alpha-2.

2. REAGENTS PROVIDED

Part No.	Quantity		Preparation
CDK155-P. Pre-coated 96-well PVDF-bottom plates	2	5	Rehydrate with 100 µl of 1X PBS
CDK155-B. Biotinylated Detection Antibody (liquid or lyophilized)	100µl	1	Reconstitute with 0.55 ml water prior to use per instructions
CDK155-C. Streptavidin-Alkaline Phosphatase Conjugate	10 µl	50 µl	Dilute prior to use per instructions
CDK155-D. Bovine Serum Albumin (BSA) – 2 g	1	1	Dissolve to prepare Dilution Buffer per instructions
CDK155-F. BCIP/NBT Substrate, 25 ml bottle	1	2	Ready to use

Human IL-13 ELISPOT Kit, Pre-coated Plates

3. MATERIALS & REAGENTS REQUIRED BUT NOT PROVIDED

- Miscellaneous laboratory plastic and/or glass, if possible sterile
- Cell culture reagents (e.g. RPMI-1640, L-glutamine, FCS)
- Cell stimulation reagents (PMA, Ionomycin)
- CO₂ incubator
- Tween 20
- Phosphate Buffered Saline (PBS)

4. STORAGE INSTRUCTIONS

Store the kit reagents between 2-8 °C. Immediately after use, remaining reagents should be returned to cold storage 2-8 °C. The expiration date of the kit and reagents is stated on the kit box label. The expiration date of the kit components can only be guaranteed if the components are stored properly, and if, in case of repeated use of one component, the reagent is not contaminated during handling.

5. SAFETY AND PRECAUTIONS FOR USE

- Handling of reagents, serum, or plasma specimens should be in accordance with local safety procedures (e.g., CDC/NIH Health manual: "Biosafety in Microbiological and Biomedical Laboratories" 2009.)
- Do not eat, drink, smoke or apply cosmetics where kit reagents are used.
- Do not pipette by mouth.
- When not in use, kit components should be stored refrigerated or frozen, as indicated on the vial or bottle labels.
- All reagents should be warmed to room temperature before use.
- Cover or cap all reagents when not in use.
- Do not mix or interchange reagents between different lots.
- Do not use reagents beyond the expiration date of the kit.
- Use a clean, disposable, plastic pipette tip for each reagent, standard, or specimen addition in order to avoid cross contamination.
- Use clean plastic containers to prepare the washing solution.
- Thoroughly mix the reagents and samples before use by agitation or swirling.
- All residual washing liquid must be drained from the wells by efficient aspiration, or by decantation, followed by tapping the plate forcefully on absorbent paper. Never insert absorbent paper directly into the wells.
- When pipetting reagents, maintain a consistent order of addition from well-to-well. This will ensure equal incubation times for all wells.
- **BCIP/NBT buffer is potentially carcinogenic and should be disposed of appropriately. Caution should be taken when handling this reagent, always wear gloves.**
- Follow incubation times described in the assay procedure.




Human IL-13 ELISPOT Kit, Pre-coated Plates


6. PRINCIPLE OF THE METHOD

A capture antibody highly specific for the analyte of interest is coated to the wells of a PVDF bottomed 96 well microtiter plate during kit manufacture. The plate is then blocked to minimize any non-antibody dependent nonspecific binding and washed. Cell suspension and stimulant are added and the plate is incubated, allowing the specific antibodies to bind any analytes produced. Cells are then removed by washing prior to the addition of biotinylated detection antibodies which bind to the previously captured analyte. Enzyme conjugated streptavidin is added, binding to the detection antibodies. Following incubation and washing, substrate is applied to the wells resulting in colored spots which can be quantified using appropriate analysis software or manually using a microscope.


1. 96-PVDF bottomed-well plates are first treated with 35% ethanol and then coated with capture antibody.




2. Cells are incubated in the presence of the stimulating agent. Upon stimulation they release cytokines which bind to the capture antibodies.



3. Cell removal by washing. Incubation with biotinylated detection antibody.




4. Any excess unbound detection antibodies is removed by washing. Incubation with streptavidin - alkaline phosphatase conjugate.



5. Any excess unbound Strep-AP is removed by washing. Incubation with BCIP/NBT.

Finally BCIP/NBT reduction by alkaline phosphatase give a precipitated product which give blue/purple spots.

One spot correspond to one single producing cell.



 Capture antibody

● Cytokines

 Cell

 Biotinylated detection antibody

 Streptavidin - alkaline phosphatase conjugated

 Substrate product



Human IL-13 ELISPOT Kit, Pre-coated Plates

7. REAGENT PREPARATION

7.1. Phosphate Buffered Saline (10X concentrate solution)

For 1 liter of 10X PBS, weigh out:

80.0 g	NaCl
2.0 g	KH ₂ PO ₄
14.4 g	Na ₂ HPO ₄ ·2H ₂ O

Add distilled water to 1 liter. Check that pH is 7.4 +/- 0.1

NOTE: This is a 10X stock solution. This solution should be diluted to 1X before use.

7.2. Wash Buffer (PBST) – 0.05% Tween in PBS

For one plate dissolve 50 µL of Tween-20 in 100 mL of 1X PBS.

7.3. Dilution Buffer – 1% BSA (CDK155-D) in PBS

For one plate, dissolve 0.2 g of BSA in 20 mL of 1X PBS.

7.4. CDK155-B - Detection Antibody

Reconstitute the lyophilized antibody with 0.55 ml of distilled water. Gently mix the solution and wait until all the lyophilized material has dissolved. To avoid nonspecific background, it is recommended to filter the working solution using a disposable syringe and a 0.2 µm filter disc.

For one plate, dilute 100 µL of capture antibody in 10 mL of Dilution Buffer and mix well.

If not used within a short period of time, reconstituted Detection Antibody should be aliquoted and stored at -20 °C. Under these conditions, the reagent is stable for one year. For optimal performance, prepare the reconstituted antibody dilution immediately prior to use.

7.5. CDK155-C - Streptavidin-Alkaline Phosphatase (AP) Conjugate

Centrifuge vial for a few seconds to collect material in bottom of vial. For 1 plate, dilute 10 µL Streptavidin AP Conjugate into 10 mL Dilution Buffer and mix well. To avoid nonspecific background, it is recommended to filter the working solution using a disposable syringe and a 0.2 µm filter disc.

For optimal performance, dilute the streptavidin-alkaline phosphatase immediately prior to use. DO NOT KEEP THE DILUTIONS FOR FURTHER EXPERIMENTS

7.6. CDK155-F - BCIP/NBT

The reagent is ready-to-use. It should be clear to pale yellow. If precipitates appear, filter using a disposable syringe and a 0.2 µm filter disc.



Human IL-13 ELISPOT Kit, Pre-coated Plates

8. SAMPLE AND CONTROL PREPARATION

8.1. Cell Stimulation

Cells can either be stimulated in the antibody-coated wells (direct stimulation) or can be first stimulated in separate plates or in flasks, harvested, and then plated into the coated wells (indirect stimulation).

The method used is dependent on 1) The type of cell assayed 2) The expected cell frequency. When a low number of cytokine-producing cells are expected, it is suggested to stimulate with the direct method. When the expected number is particularly high, it is better to use the indirect ELISpot method.

All the method steps following stimulation of the cells are the same whatever the method of stimulation chosen (direct or indirect).

8.2. Positive Assay Control, IL-13 production

Suggested polyclonal activation as a positive control in your assay:

Dilute T cells in culture medium (e.g. RPMI 1640 supplemented with 2mM L-glutamine and 10% heat inactivated fetal calf serum) containing 1 ng/ml PMA and 500 ng/ml ionomycin (Sigma, Saint Louis, MO). Distribute 1×10^5 to 2.5×10^5 cells per 100 μ l in required wells of an antibody coated 96-well PVDF plate and incubate for 15-20 hours in an incubator.

For other stimulators incubation times may vary, depending on the frequency of cytokine producing cells, and should be optimized in each situation.

8.3. Negative Assay Control

Dilute PBMC in culture medium to give an appropriate cell number (same number of unstimulated cells as stimulated sample cells) per 100 μ L with no stimulation.

8.4. Samples

Dilute PBMC in culture medium and chosen stimulator (e.g., Sample, Vaccine, Peptide pool or infected cells) to give an appropriate cell number per 100 μ L.

Optimal assay performances are observed between 1×10^5 and 2.5×10^5 cells per 100 μ L.

Stimulators and incubation times can be varied depending on the frequency of cytokine producing cells and should be optimized by the testing laboratory.



Human IL-13 ELISPOT Kit, Pre-coated Plates

9. METHOD

Prepare reagents and cells as shown in sections 7 & 8.

Note: For optimal performance dilute the Streptavidin-AP immediately prior to use.

Assay Step		Details
1.	Wash	Wash the plate 3x with 100 μ L of 1X PBS per well. Empty the wells by flicking the plate over a sink & gently tapping on absorbent paper.
2.	Addition/ Incubation	Add 100 μ L of blocking buffer to every well. Incubate plate at room temperature (RT) for 2 hours.
3.	Wash	Empty the wells by flicking the plate over a sink & gently tapping on absorbent paper. (Do not wash.)
4.	Addition	Add 100 μ L of sample, positive and negative controls cell suspension to appropriate wells providing the required concentration of cells and stimulant (cells may have been previously stimulated see section 8.)
5.	Incubation	Cover the plate and incubate at 37 $^{\circ}$ C in a CO ₂ incubator for an appropriate length of time (15-20 hours). Note: do not agitate or move the plate during this incubation.
6.	Addition	Empty the wells and remove excess solution, and then add 100 μ L of Wash Buffer to every well.
7.	Incubation	Incubate the plate at 4 $^{\circ}$ C for 10 minutes.
8.	Wash	Empty the wells by flicking the plate over a sink & gently tapping on absorbent paper. Thoroughly wash the plate 3x with 100 μ L of Wash Buffer per well.
9.	Addition	Add 100 μ L of diluted Detection Antibody to every well.
10.	Incubation	Cover the plate and incubate at RT for 1 hour 30 minutes
11.	Wash	Empty the wells as previously done, and wash the plate 3x with 100 μ L of Wash Buffer.
12.	Addition	Add 100 μ L of diluted Streptavidin-AP Conjugate to every well.
13.	Incubation	Cover the plate and incubate at RT for 1 hour.
14.	Wash	Empty the wells by flicking the plate over a sink & gently tapping on absorbent paper. Thoroughly wash the plate 3x with 100 μ L of Wash Buffer per well.
15.	Wash	Peel off the plate bottom and wash both sides of the membrane 3x under running distilled water, once washing is complete, remove any excess solution by repeated tapping on absorbent paper.
16.	Addition	Add 100 μ L of ready-to-use BCIP/NBT buffer to every well.
17.	Development	Incubate the plate for 5-15 minutes, monitoring spot formation visually throughout the incubation period to assess sufficient color development.
18.	Wash	Empty the wells and rinse both sides of the membrane 3x under running distilled water. Completely remove any excess solution by gentle repeated tapping on absorbent paper.

Read Spots:

Allow the wells to dry and then read results. The frequency of colored spots corresponding to the cytokine producing cells can be determined using an appropriate ELISPOT reader and analysis software, or manually using a microscope.

Note: Spots may become sharper after overnight incubation at 2-8 $^{\circ}$ C. Plate should be stored at RT away from direct light, but please note color may fade over prolonged periods, so read results within 24 hours.

Human IL-13 ELISPOT Kit, Pre-coated Plates

10. PERFORMANCE CHARACTERISTICS

10.1. Reproducibility and Linearity

Intra-assay reproducibility and linearity were evaluated by measuring the spot development following the stimulation (LPS) of 5 different T cell concentrations, 12 repetitions. The data shows the mean spot number, range and CV for the five cell concentrations.

Cells/well	n	Mean # of Spots per well	Min.	Max	CV%
100000 (recommended)	12	294	268	323	7.1
50000	12	199	181	211	6.1
25000	12	104	76	133	15.2
12500	12	50	33	64	19.2
6250	12	22	12	38	30.9

NOT FOR HUMAN USE. FOR RESEARCH ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.

