



Rat matrix metallopeptidase 8 (neutrophil collagenase) (MMP8) ELISA kit

Product Code	CSB-E07406r
Abbreviation	MMP8
Protein Biological Process 1	Developmental Protein
Target Name	matrix metallopeptidase 8 (neutrophil collagenase)
Uniprot No.	O88766
Alias	CLG1, HNC, PMNL-CL, PMNL collagenase matrix metalloproteinase 8 matrix metalloproteinase 8 (neutrophil collagenase)
Product Type	ELISA Kit
Immunogen Species	Rattus norvegicus (Rat)
Protein Biological Process 3	Collagen degradation
Sample Types	serum, plasma, cell culture supernates, tissue homogenates
Detection Range	0.312 ng/mL-20 ng/mL
Sensitivity	0.078 ng/mL
Assay Time	1-5h
Sample Volume	50-100ul
Detection Wavelength	450 nm
Lead Time	3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx.
Research Area	Cancer
Gene Names	Mmp8
Tag Info	quantitative
Protein Description	Sandwich
Description	This Rat MMP8 ELISA Kit was designed for the quantitative measurement of Rat MMP8 protein in serum, plasma, cell culture supernates, tissue homogenates. It is a Sandwich ELISA kit, its detection range is 0.312 ng/mL-20 ng/mL and the sensitivity is 0.078 ng/mL.
Target Details	Proteins of the matrix metalloproteinase (MMP) family are involved in the breakdown of extracellular matrix in normal physiological processes, such as embryonic development, reproduction, and tissue remodeling, as well as in disease processes, such as arthritis and metastasis. Most MMP s are secreted



as inactive proproteins which are activated when cleaved by extracellular proteinases. However, the enzyme encoded by this gene is stored in secondary granules within neutrophils and is activated by autolytic cleavage. Its function is degradation of type I, II and III collagens. The gene is part of a cluster of MMP genes which localize to chromosome 11q22.3.

Product Precision

Intra-assay Precision (Precision within an assay): CV%<8%

Three samples of known concentration were tested twenty times on one plate to assess.

Inter-assay Precision (Precision between assays): CV%<10%

Three samples of known concentration were tested in twenty assays to assess.

Linearity

To assess the linearity of the assay, samples were spiked with high concentrations of rat MMP8 in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.

	Sample	Serum(n=4)
1:1	Average %	97
	Range %	92-101
1:2	Average %	90
	Range %	85-94
1:4	Average %	88
	Range %	84-92
1:8	Average %	102
	Range %	98-108


Recovery

The recovery of rat MMP8 spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

Sample Type	Average % Recovery	Range
Serum (n=5)	94	90-98
EDTA plasma (n=4)	84	80-88

Typical

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



ng/ml	OD1	OD2	Average	Corrected
20	2.620	2.459	2.540	2.363
10	2.056	2.146	2.101	1.924
5	1.383	1.432	1.408	1.231
2.5	0.825	0.801	0.813	0.636
1.25	0.509	0.518	0.514	0.337
0.625	0.417	0.405	0.411	0.234
0.312	0.251	0.257	0.254	0.077
0	0.180	0.173	0.177	

Msds

```
{
  "0": {
    "fileurl": "https://www.cusabio.com/uploadfile/msds/MSDS CSB-E07406r.pdf",
    "filename": "MSDS"
  }
}
```