



Human Peroxisome Proliferator-activated receptor γ , PPAR- γ ELISA Kit

| | |
|-------------------------------------|---|
| Product Code | CSB-E08623h |
| Abbreviation | PPARG |
| Protein Biological Process 1 | Transcription/Transcription regulation |
| Target Name | peroxisome proliferator-activated receptor gamma |
| Uniprot No. | P37231 |
| Alias | CIMT1, GLM1, NR1C3, PPARG1, PPARG2, PPARgamma, OTTHUMP00000185030 OTTHUMP00000185033 PPAR gamma nuclear receptor subfamily 1 group C member 3 peroxisome proliferative activated receptor gamma peroxi |
| Product Type | ELISA Kit |
| Immunogen Species | Homo sapiens (Human) |
| Protein Biological Process 3 | Transcription |
| Sample Types | serum, plasma |
| Detection Range | 78 pg/mL-5000 pg/mL |
| Sensitivity | 19.5 pg/mL |
| Assay Time | 1-5h |
| Sample Volume | 50-100ul |
| Detection Wavelength | 450 nm |
| Lead Time | 3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx. |
| Research Area | Metabolism |
| Gene Names | PPARG |
| Tag Info | quantitative |
| Protein Description | Sandwich |
| Description | This Human PPARG ELISA Kit was designed for the quantitative measurement of Human PPARG protein in serum, plasma. It is a Sandwich ELISA kit, its detection range is 78 pg/mL-5000 pg/mL and the sensitivity is 19.5 pg/mL. |
| Target Details | This gene encodes a member of the peroxisome proliferator-activated receptor (PPAR) subfamily of nuclear receptors. PPARs form heterodimers with retinoid X receptors (RXRs) and these heterodimers regulate transcription of various |



genes. Three subtypes of PPARs are known: PPAR-alpha, PPAR-delta, and PPAR-gamma. This protein is PPAR-gamma and is a regulator of adipocyte differentiation. Additionally, PPAR-gamma has been implicated in the pathology of numerous diseases including obesity, diabetes, atherosclerosis and cancer. Alternatively spliced transcript variants that encode different isoforms have been described.

Product Precision

Intra-assay Precision (Precision within an assay): CV%<8%

Three samples of known concentration were tested twenty times on one plate to assess.

Inter-assay Precision (Precision between assays): CV%<10%

Three samples of known concentration were tested in twenty assays to assess.

Linearity

To assess the linearity of the assay, samples were spiked with high concentrations of human PPAR- γ in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.

| ? | Sample | Serum(n=4) |
|-----|-----------|------------|
| 1:1 | Average % | 93 |
| | Range % | 86-99 |
| 1:2 | Average % | 94 |
| | Range % | 84-99 |
| 1:4 | Average % | 90 |
| | Range % | 85-96 |
| 1:8 | Average % | 86 |
| | Range % | 80-94 |


Recovery

The recovery of human PPAR- γ spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

| Sample Type | Average % Recovery | Range |
|-------------------|--------------------|--------|
| Serum (n=5) | 96 | 89-102 |
| EDTA plasma (n=4) | 89 | 84-92 |

Typical

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



| pg/ml | OD1 | OD2 | Average | Corrected |
|-------|-------|-------|---------|-----------|
| 5000 | 2.315 | 2.215 | 2.265 | 2.119 |
| 2500 | 1.817 | 1.717 | 1.767 | 1.621 |
| 1250 | 1.257 | 1.157 | 1.207 | 1.061 |
| 625 | 0.712 | 0.702 | 0.707 | 0.561 |
| 312 | 0.418 | 0.408 | 0.413 | 0.267 |
| 156 | 0.331 | 0.321 | 0.326 | 0.180 |
| 78 | 0.255 | 0.245 | 0.250 | 0.104 |
| 0 | 0.151 | 0.141 | 0.146 | ? |

Msds

```
{
  "0": {
    "fileurl": "https://www.cusabio.com/uploadfile/msds/MSDS CSB-E08623h.pdf",
    "filename": "MSDS"
  }
}
```