





# Human CX3C-chemokine receptor 1,CX3CR1 **ELISA Kit**

Product Code	CSB-E09942h
Abbreviation	CX3CR1
Protein Biological Process 1	Immunity
Target Name	chemokine (C-X3-C motif) receptor 1
Uniprot No.	P49238
Alias	CCRL1, CMKBRL1, CMKDR1, GPR13, GPRV28, V28, G protein-coupled receptor 13 chemokine (C-C) receptor-like 1 chemokine (C-X3-C) receptor 1
Product Type	ELISA Kit
Immunogen Species	Homo sapiens (Human)
Protein Biological Process 3	Host-virus interaction
Sample Types	serum, plasma, tissue homogenates
<b>Detection Range</b>	31.25 pg/mL-2000 pg/mL
Sensitivity	7.8 pg/mL
Assay Time	1-5h
Sample Volume	50-100ul
<b>Detection Wavelength</b>	450 nm
Lead Time	3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx.
Research Area	Immunology
Gene Names	CX3CR1
Tag Info	quantitative
<b>Protein Description</b>	Sandwich
Description	This Human CX3CR1 ELISA Kit was designed for the quantitative measurement of Human CX3CR1 protein in serum, plasma, tissue homogenates. It is a Sandwich ELISA kit, its detection range is 31.25 pg/mL-2000 pg/mL and the sensitivity is 7.8 pg/mL.
Product Precision	Intra-assay Precision (Precision within an assay): CV%<8% Three samples of known concentration were tested twenty times on one plate

Inter-assay Precision (Precision between assays): CV%<10%

**CUSABIO**® Your good partner in biology research

#### **CUSABIO TECHNOLOGY LLC**











Three samples of known concentration were tested in twenty assays to assess.

# Linearity

To assess the linearity of the assay, samples were spiked with high concentrations of human CX3CR1 in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.

?	Sample	Serum(n=4)
1:1	Average %	87
	Range %	82-92
1:2	Average %	92
	Range %	88-97
1:4	Average %	84
	Range %	80-88
1:8	Average %	101
	Range %	95-105

### Recovery

The recovery of human CX3CR1 spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

Sample Type	Average % Recovery	Range
Serum (n=5)	95	90-99
EDTA plasma (n=4)	97	92-101

## **Typical**


These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.











pg/ml OD1 OD2 Average Corrected

2000 2.764 2.647 2.706 2.549 1000 2.006 2.018 2.012 1.855 500 1.295 1.238 1.267 1.110 250 0.731 0.753 0.742 0.585 125 0.498 0.512 0.505 0.348 62.5 0.327 0.341 0.334 0.177 31.25 0.243 0.239 0.241 0.084 ? 0.159 0.154 0.157

**Msds** 

{"0":{"fileurl":"https://www.cusabio.com/uploadfile/msds/MSDS CSB-E09942h.pdf","filename":"MSDS"}}