



Mouse keyhole limpet hemocyanin (KLH) antibody (IgG) ELISA kit

| | | | | | | | | | | | | | | | | | | |
|----------------------|---|-------------|--|----------------|---------------|-------------|------------------|------|-------|------------------|------|-------|---------------|-------------|------|---------------|-------------|------|
| Product Code | CSB-E13374m | | | | | | | | | | | | | | | | | |
| Abbreviation | KLH Ab (IgG) | | | | | | | | | | | | | | | | | |
| Target Name | keyhole limpet hemocyanin (KLH) antibody (IgG) | | | | | | | | | | | | | | | | | |
| Product Type | ELISA Kit | | | | | | | | | | | | | | | | | |
| Immunogen Species | Mus musculus (Mouse) | | | | | | | | | | | | | | | | | |
| Sample Types | serum, plasma | | | | | | | | | | | | | | | | | |
| Detection Range | Request Information | | | | | | | | | | | | | | | | | |
| Sensitivity | Request Information | | | | | | | | | | | | | | | | | |
| Assay Time | 1-5h | | | | | | | | | | | | | | | | | |
| Sample Volume | 50-100ul | | | | | | | | | | | | | | | | | |
| Detection Wavelength | 450 nm | | | | | | | | | | | | | | | | | |
| Lead Time | 3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx. | | | | | | | | | | | | | | | | | |
| Research Area | Tags & Cell Markers | | | | | | | | | | | | | | | | | |
| Tag Info | qualitative | | | | | | | | | | | | | | | | | |
| Protein Description | Indirect | | | | | | | | | | | | | | | | | |
| Product Precision | Intra-assay Precision (Precision within an assay): CV%<15% Three samples of known concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV%<15% Three samples of known concentration were tested in twenty assays to assess. | | | | | | | | | | | | | | | | | |
| Typical | <table><tr><td>Test parameter</td><td>specification</td><td>test result</td></tr><tr><td>Positive control</td><td>>1.0</td><td>1.592</td></tr><tr><td>Negative control</td><td><0.2</td><td>0.115</td></tr><tr><td>Positive rate</td><td>10?Positive</td><td>100%</td></tr><tr><td>Negative rate</td><td>10?Negative</td><td>100%</td></tr></table> | | | Test parameter | specification | test result | Positive control | >1.0 | 1.592 | Negative control | <0.2 | 0.115 | Positive rate | 10?Positive | 100% | Negative rate | 10?Negative | 100% |
| Test parameter | specification | test result | | | | | | | | | | | | | | | | |
| Positive control | >1.0 | 1.592 | | | | | | | | | | | | | | | | |
| Negative control | <0.2 | 0.115 | | | | | | | | | | | | | | | | |
| Positive rate | 10?Positive | 100% | | | | | | | | | | | | | | | | |
| Negative rate | 10?Negative | 100% | | | | | | | | | | | | | | | | |