





Rat glucagon-like peptide-2,GLP-2 ELISA Kit

| Product Code | CSB-E14203r | | |
|-----------------------------|---|--|--|
| Abbreviation | Glp2 | | |
| Target Name | glucagon-like peptide-2,GLP-2 | | |
| Alias | N/A | | |
| Product Type | ELISA Kit | | |
| Immunogen Species | Rattus norvegicus (Rat) | | |
| Sample Types | serum, plasma, tissue homogenates | | |
| Detection Range | 15.6 pg/mL-1000 pg/mL | | |
| Sensitivity | 3.9 pg/mL | | |
| Assay Time | 1-5h | | |
| Sample Volume | 50-100ul | | |
| Detection Wavelength 450 nm | | | |
| Lead Time | 3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx. | | |
| Research Area | Metabolism | | |
| Gene Names | Gcg | | |
| Tag Info | quantitative | | |
| Protein Description | Sandwich | | |
| Description | This Rat Glp2 ELISA Kit was designed for the quantitative measurement of Rat Glp2 protein in serum, plasma, tissue homogenates. It is a Sandwich ELISA kit, its detection range is 15.6 pg/mL-1000 pg/mL and the sensitivity is 3.9 pg/mL. | | |
| Product Precision | Intra-assay Precision (Precision within an assay): CV%<8% Three samples of known concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV%<10% Three samples of known concentration were tested in twenty assays to assess. | | |
| Linearity | To assess the linearity of the assay, samples were spiked with high concentrations of rat GLP-2 in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay. ? Sample Serum(n=4) Average % 89 Range % 82-95 | | |

CUSABIO TECHNOLOGY LLC



| CUSABIO TECHNOLOGY LLC | | | | | |
|------------------------|-----------------------------|----------------------------|--------------------------|--|--|
| | Tel: +1-301-363-4651 | Email: cusabio@cusabio.com | Website: www.cusabio.com | | |

| 1:2 | Average % | 98 |
|-----|-----------|--------|
| 1.2 | Range % | 93-103 |
| 1:4 | Average % | 97 |
| 1.4 | Range % | 90-104 |
| 1.0 | Average % | 86 |
| 1:8 | Range % | 82-91 |


Recovery

The recovery of rat GLP-2 spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

| Sample Type | Average % Recovery | Range |
|-------------------|--------------------|--------|
| Serum (n=5) | 92 | 87-99 |
| EDTA plasma (n=4) | 99 | 96-104 |

Typical

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



pg/ml OD1 OD2 Average Corrected

| 1000 | 2.057 2.052 2.055 | 1.912 |
|------|-------------------|-------|
| 500 | 1.603 1.633 1.618 | 1.475 |
| 250 | 1.212 1.228 1.220 | 1.077 |
| 125 | 0.846 0.911 0.879 | 0.736 |
| 62.5 | 0.559 0.543 0.551 | 0.408 |
| 31.2 | 0.332 0.351 0.342 | 0.199 |
| 15.6 | 0.252 0.265 0.259 | 0.116 |
| 0 | 0.142 0.144 0.143 | ? |

Msds

 $\label{eq:complex} $$ \{"0": {"fileurl": "https://www.cusabio.com/uploadfile/msds/MSDS CSB-E14203r.pdf", "filename": "MSDS"}\} $$$