



Rabbit Matrix Metalloproteinase 13,MMP-13 ELISA Kit

Product Code	CSB-E14220Rb
Abbreviation	MMP13
Protein Biological Process 1	Developmental Protein
Target Name	matrix metalloproteinase 13 (collagenase 3)
Uniprot No.	O62806
Alias	CLG3, MANDP1, collagenase 3 matrix metalloproteinase 13 matrix metalloproteinase 13 (collagenase 3)
Product Type	ELISA Kit
Immunogen Species	Oryctolagus cuniculus (Rabbit)
Protein Biological Process 3	Collagen degradation
Sample Types	serum, plasma, tissue homogenates
Detection Range	93.75 pg/mL-6000 pg/mL
Sensitivity	23.44 pg/mL
Assay Time	1-5h
Sample Volume	50-100ul
Detection Wavelength	450 nm
Lead Time	3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx.
Research Area	Cancer
Gene Names	MMP13
Tag Info	quantitative
Protein Description	Sandwich

Description

This Rabbit MMP13 ELISA Kit was designed for the quantitative measurement of Rabbit MMP13 protein in serum, plasma, tissue homogenates. It is a Sandwich ELISA kit, its detection range is 93.75 pg/mL-6000 pg/mL and the sensitivity is 23.44 pg/mL.

Target Details

Proteins of the matrix metalloproteinase (MMP) family are involved in the breakdown of extracellular matrix in normal physiological processes, such as embryonic development, reproduction, and tissue remodeling, as well as in disease processes, such as arthritis and metastasis. Most MMP s are secreted



as inactive proproteins which are activated when cleaved by extracellular proteinases. This protein cleaves type II collagen more efficiently than types I and III. It may be involved in articular cartilage turnover and cartilage pathophysiology associated with osteoarthritis. The gene is part of a cluster of MMP genes which localize to chromosome 11q22.3.

Product Precision

Intra-assay Precision (Precision within an assay): CV%<8%

Three samples of known concentration were tested twenty times on one plate to assess.

Inter-assay Precision (Precision between assays): CV%<10%

Three samples of known concentration were tested in twenty assays to assess.

Linearity

To assess the linearity of the assay, samples were spiked with high concentrations of rabbit MMP-13 in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.

	Sample	Serum(n=4)
1:1	Average %	91
	Range %	86-95
1:2	Average %	105
	Range %	97-108
1:4	Average %	90
	Range %	85-95
1:8	Average %	95
	Range %	89-100


Recovery

The recovery of rabbit MMP-13 spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

Sample Type	Average % Recovery	Range
Serum (n=5)	90	83-95
EDTA plasma (n=4)	98	90-102

Typical

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



pg/ml	OD1	OD2	Average	Corrected
6000	2.554	2.444	2.499	2.405
3000	1.937	1.976	1.957	1.863
1500	1.409	1.422	1.416	1.322
750	0.815	0.789	0.802	0.708
375	0.422	0.434	0.428	0.334
187.5	0.302	0.289	0.296	0.202
93.75	0.207	0.198	0.203	0.109
0	0.093	0.095	0.094	

Msds

```
{
  "0": {
    "fileurl": "https://www.cusabio.com/uploadfile/msds/MSDS CSB-E14220Rb.pdf",
    "filename": "MSDS"
  }
}
```