



Mouse Ovalbumin Specific IgG(OVA sIgG) ELISA Kit

| Product Code | CSB-E14990m | | | | | | | | | | | | | | | | |
|-----------------------------|--|-------------|----------------|---------------|-------------|------------------|------|-------|------------------|------|-------|---------------|-------------|------|---------------|-------------|------|
| Abbreviation | OVA sIgG | | | | | | | | | | | | | | | | |
| Target Name | Ovalbumin Specific IgG(OVA sIgG) | | | | | | | | | | | | | | | | |
| Alias | N/A | | | | | | | | | | | | | | | | |
| Product Type | ELISA Kit | | | | | | | | | | | | | | | | |
| Immunogen Species | Mus musculus (Mouse) | | | | | | | | | | | | | | | | |
| Sample Types | serum, plasma | | | | | | | | | | | | | | | | |
| Detection Range | Request Information | | | | | | | | | | | | | | | | |
| Sensitivity | Request Information | | | | | | | | | | | | | | | | |
| Assay Time | 1-5h | | | | | | | | | | | | | | | | |
| Sample Volume | 50-100ul | | | | | | | | | | | | | | | | |
| Detection Wavelength | 450 nm | | | | | | | | | | | | | | | | |
| Lead Time | 3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx. | | | | | | | | | | | | | | | | |
| Research Area | Others | | | | | | | | | | | | | | | | |
| Tag Info | qualitative | | | | | | | | | | | | | | | | |
| Protein Description | Indirect | | | | | | | | | | | | | | | | |
| Product Precision | <p>Intra-assay Precision (Precision within an assay): CV%<15%</p> <p>Three samples of known concentration were tested twenty times on one plate to assess.</p> <p>Inter-assay Precision (Precision between assays): CV%<15%</p> <p>Three samples of known concentration were tested in twenty assays to assess.</p> | | | | | | | | | | | | | | | | |
| Typical | <table> <thead> <tr> <th>Test parameter</th><th>specification</th><th>test result</th></tr> </thead> <tbody> <tr> <td>Positive control</td><td>>0.8</td><td>1.663</td></tr> <tr> <td>Negative control</td><td><0.2</td><td>0.136</td></tr> <tr> <td>Positive rate</td><td>10?Positive</td><td>100%</td></tr> <tr> <td>Negative rate</td><td>10?Negative</td><td>100%</td></tr> </tbody> </table> | | Test parameter | specification | test result | Positive control | >0.8 | 1.663 | Negative control | <0.2 | 0.136 | Positive rate | 10?Positive | 100% | Negative rate | 10?Negative | 100% |
| Test parameter | specification | test result | | | | | | | | | | | | | | | |
| Positive control | >0.8 | 1.663 | | | | | | | | | | | | | | | |
| Negative control | <0.2 | 0.136 | | | | | | | | | | | | | | | |
| Positive rate | 10?Positive | 100% | | | | | | | | | | | | | | | |
| Negative rate | 10?Negative | 100% | | | | | | | | | | | | | | | |