



Mouse Eotaxin 3 ELISA Kit

Product Code	CSB-E15062m			
Abbreviation	Eotaxin 3			
Target Name	Eotaxin 3			
Product Type	ELISA Kit			
Immunogen Species	Mus musc	culus (Mouse)		
Sample Types	serum, pla	asma, cell culture sup	pernates, tissue homogenates	
Detection Range	6.25 pg/mL-400 pg/mL 1.56 pg/mL			
Sensitivity				
Assay Time	1-5h			
Sample Volume	50-100ul 450 nm 3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx. Immunology			
Detection Wavelength				
Lead Time				
Research Area				
Tag Info	quantitativ	re		
Protein Description	Sandwich			
Description	This Mouse Eotaxin 3 ELISA Kit was designed for the quantitative measurement of Mouse Eotaxin 3 protein in serum, plasma, cell culture supernates, tissue homogenates. It is a Sandwich ELISA kit, its detection range is 6.25 pg/mL-400 pg/mL and the sensitivity is 1.56 pg/mL.			
Product Precision	Intra-assay Precision (Precision within an assay): CV%<8% Three samples of known concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV%<10% Three samples of known concentration were tested in twenty assays to assess.			
Linearity	To assess the linearity of the assay, samples were spiked with high concentrations of mouse Eotaxin 3 in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.			
		Sample	Serum(n=4)	
	1:1	Average % Range %	92 90-93	
	4.0	Average %	96	
	1:2	Range %	92-100	
	1:4	Average % Range %	103 101-105	

CUSABIO® Your good partner in biology research







1:8	Average %	89
	Range %	84-94


Recovery

The recovery of mouse Eotaxin 3 spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

Sample Type	Average % Recovery	Range
Serum (n=5)	92	90-94
EDTA plasma (n=4)	89	85-94

Typical

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



pg/ml OD1 OD2 Average Corrected

1- 3,		9
400	2.927 2.874 2.901	2.745
200	2.015 2.135 2.075	1.919
100	1.303 1.403 1.353	1.197
50	0.801 0.822 0.812	0.656
25	0.532 0.524 0.528	0.372
12.5	0.319 0.332 0.326	0.170
6.25	0.244 0.254 0.249	0.093
0	0.151 0.160 0.156	

Msds

{"0":{"fileurl":"https://www.cusabio.com/uploadfile/msds/MSDS CSB-E15062m.pdf","filename":"MSDS"}}