



Human alpha-synuclein ?SNCA?ELISA kit

Product Code	CSB-E18033h
Abbreviation	SNCA
Target Name	synuclein, alpha (non A4 component of amyloid precursor)
Uniprot No.	P37840
Alias	MGC110988, NACP, PARK1, PARK4, PD1, OTTHUMP00000161559 OTTHUMP00000161561 alpha-synuclein alpha-synuclein, isoform NACP140 non A-beta component of AD amyloid
Product Type	ELISA Kit
Immunogen Species	Homo sapiens (Human)
Sample Types	serum, plasma
Detection Range	0.312 ng/mL-20 ng/mL
Sensitivity	0.078 ng/mL
Assay Time	1-5h
Sample Volume	50-100ul
Detection Wavelength	450 nm
Lead Time	3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx.
Research Area	Neuroscience
Quality Control	A microplate reader capable of measuring absorbance at 450 nm, with the correction wavelength set at 540 nm or 570 nm. An incubator can provide stable incubation conditions up to 37°C±5°C. Centrifuge Vortex Squirt bottle, manifold dispenser, or automated microplate washer Absorbent paper for blotting the microtiter plate 50-300ul multi-channel micropipette Pipette tips Single-channel micropipette with different ranges 100ml and 500ml graduated cylinders Deionized or distilled water Timer Test tubes for dilution
Gene Names	SNCA
Tag Info	quantitative
Protein Description	Sandwich
Component	A micro ELISA plate --- The 96-well plate has been pre-coated with an anti-human SNCA antibody. This dismountable microplate can be divided into 12 x 8



strip plates.

Two vials lyophilized standard ---Dilute a bottle of the standard at dilution series, read the OD values, and then draw a standard curve.

One vial Biotin-labeled SNCA antibody (100 x concentrate) (120 µl/bottle) ---Act as the detection antibody.

One vial HRP-avidin (100 x concentrate) (120 µl/bottle) ---Bind to the detection antibody and react with the TMB substrate to make the solution chromogenic.

One vial Biotin-antibody Diluent (15 ml/bottle) ---Dilute the Biotin-antibody.

One vial HRP-avidin Diluent (15 ml/bottle) ---Dilute the HRP-avidin solution.

One vial Sample Diluent (50 ml/bottle)---Dilute the sample to an appropriate concentration.

One vial Wash Buffer (25 x concentrate) (20 ml/bottle) ---Wash away unbound or free substances.

One vial TMB Substrate (10 ml/bottle) ---Act as the chromogenic agent. TMB interacts with HRP, eliciting the solution turns blue.

One vial Stop Solution (10 ml/bottle) ---Stop the color reaction. The solution color immediately turns from blue to yellow.

Four Adhesive Strips (For 96 wells) --- Cover the microplate when incubation. An instruction manual

Description

This Human SNCA ELISA Kit was designed for the quantitative measurement of Human SNCA protein in serum, plasma. It is a Sandwich ELISA kit, its detection range is 0.312 ng/mL-20 ng/mL and the sensitivity is 0.078 ng/mL.

Target Details

Alpha-synuclein is a member of the synuclein family, which also includes beta- and gamma-synuclein. Synucleins are abundantly expressed in the brain and alpha- and beta-synuclein inhibit phospholipase D2 selectively. SNCA may serve to integrate presynaptic signaling and membrane trafficking. Defects in SNCA have been implicated in the pathogenesis of Parkinson disease. SNCA peptides are a major component of amyloid plaques in the brains of patients with Alzheimer s disease. Four alternatively spliced transcripts encoding two different isoforms have been identified for this gene.

Product Precision

Intra-assay Precision (Precision within an assay): CV%<8%

Three samples of known concentration were tested twenty times on one plate to assess.

Inter-assay Precision (Precision between assays): CV%<10%

Three samples of known concentration were tested in twenty assays to assess.

Linearity

To assess the linearity of the assay, samples were spiked with high concentrations of human SNCA oligomer in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.

?	Sample	Serum(n=4)
1:1	Average %	87
	Range %	82-94
1:2	Average %	92
	Range %	87-99



1:4	Average %	96
	Range %	90-102
1:8	Average %	102
	Range %	96-108


Recovery

The recovery of human SNCA oligomer spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

Sample Type	Average % Recovery	Range
Serum (n=5)	90	86-97
EDTA plasma (n=4)	93	90-96

Typical

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



ng/ml	OD1	OD2	Average	Corrected
20	2.368	2.277	2.323	2.161
10	1.813	1.905	1.859	1.697
5	1.198	1.164	1.181	1.019
2.5	0.772	0.801	0.787	0.625
1.25	0.451	0.434	0.443	0.281
0.625	0.328	0.316	0.322	0.160
0.312	0.241	0.233	0.237	0.075
0	0.163	0.161	0.162	?

Msds

{\"0\":{\"fileurl\":\"https://www.cusabio.com/uploadfile/msds/MSDS CSB-E18033h.pdf\",\"filename\":\"MSDS\"}}