





Human Complement C1q-like protein 3(C1QL3) **ELISA** kit

Product Code	CSB-EL003644HU
Abbreviation	C1QL3
Target Name	complement component 1, q subcomponent-like 3
Uniprot No.	Q5VWW1
Alias	C1QTNF13, C1ql, K100
Product Type	ELISA Kit
Immunogen Species	Homo sapiens (Human)
Sample Types	serum, plasma, tissue homogenates
Detection Range	0.312 ng/mL-20 ng/mL
Sensitivity	0.078 ng/mL
Assay Time	1-5h
Sample Volume	50-100ul
Detection Wavelength	450 nm
Lead Time	3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx.
Research Area	Others
Gene Names	C1QL3
Tag Info	quantitative
Protein Description	Sandwich
Description	This Human C1QL3 ELISA Kit was designed for the quantitative measurement of Human C1QL3 protein in serum, plasma, tissue homogenates. It is a Sandwich ELISA kit, its detection range is 0.312 ng/mL-20 ng/mL and the sensitivity is 0.078 ng/mL.
Product Precision	Intra-assay Precision (Precision within an assay): CV%<8% Three samples of known concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV%<10% Three samples of known concentration were tested in twenty assays to assess.
Linearity	To assess the linearity of the assay, samples were spiked with high concentrations of human C1QL3 in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.

CUSABIO TECHNOLOGY LLC





	Sample	Serum(n=4)
1:1	Average %	105
	Range %	100-110
1:2	Average %	99
	Range %	95-107
1:4	Average %	89
	Range %	85-93
1:8	Average %	84
	Range %	80-88


Recovery

The recovery of human C1QL3 spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

Sample Type	Average % Recovery	Range
Serum (n=5)	95	90-100
EDTA plasma (n=4)	90	85-95

Typical

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



ng/ml OD1 OD2 Average Corrected

20	2.510 2.350 2.430	2.292
10	1.961 1.856 1.909	1.771
5	1.276 1.324 1.300	1.162
2.5	0.799 0.825 0.812	0.674
1.25	0.453 0.427 0.440	0.302
0.625	0.347 0.372 0.360	0.222
0.312	0.262 0.259 0.261	0.123
0	0.141 0.135 0.138	

Msds

{"0":{"fileurl":"https://www.cusabio.com/uploadfile/msds/MSDS CSB-EL003644HU.pdf", "filename": "MSDS"}}