



Human Interferon-induced 35 kDa protein(IFI35) ELISA kit

Product Code	CSB-EL011013HU
Abbreviation	IFI35
Target Name	interferon-induced protein 35
Uniprot No.	P80217
Alias	FLJ21753, IFP35
Product Type	ELISA Kit
Immunogen Species	Homo sapiens (Human)
Sample Types	serum, plasma, tissue homogenates, cell lysates
Detection Range	25 pg/mL-1600 pg/mL
Sensitivity	6.25 pg/mL
Assay Time	1-5h
Sample Volume	50-100ul
Detection Wavelength	450 nm
Lead Time	3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx.
Research Area	Cell Biology
Gene Names	IFI35
Tag Info	quantitative
Protein Description	Sandwich
Description	This Human IFI35 ELISA Kit was designed for the quantitative measurement of Human IFI35 protein in serum, plasma, tissue homogenates, cell lysates. It is a Sandwich ELISA kit, its detection range is 25 pg/mL-1600 pg/mL and the sensitivity is 6.25 pg/mL.
Product Precision	<p>Intra-assay Precision (Precision within an assay): CV%<8% Three samples of known concentration were tested twenty times on one plate to assess.</p> <p>Inter-assay Precision (Precision between assays): CV%<10% Three samples of known concentration were tested in twenty assays to assess.</p>
Linearity	To assess the linearity of the assay, samples were spiked with high concentrations of human IFI35 in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.



?	Sample	Serum(n=4)
1:1	Average %	98
	Range %	96-102
1:2	Average %	95
	Range %	89-100
1:4	Average %	94
	Range %	90-98
1:8	Average %	88
	Range %	86-95


Recovery

The recovery of human IFI35 spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

Sample Type	Average % Recovery	Range
Serum (n=5)	105	99-108
EDTA plasma (n=4)	93	90-98

Typical

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



pg/ml	OD1	OD2	Average	Corrected
1600	2.768	2.848	2.808	2.715
800	1.622	1.709	1.666	1.573
400	1.034	1.115	1.075	0.982
200	0.669	0.655	0.662	0.569
100	0.323	0.346	0.335	0.242
50	0.227	0.222	0.225	0.132
25	0.144	0.134	0.139	0.046
0	0.096	0.090	0.093	?

Msds

{"0":{"fileurl":"https://www.cusabio.com/uploadfile/msds/MSDS CSB-EL011013HU.pdf","filename":"MSDS"}}}