



Human Ly6/PLAUR domain-containing protein 1 (LYPD1) ELISA kit

Product Code	CSB-EL013261HU
Abbreviation	LYPD1
Target Name	LY6/PLAUR domain containing 1
Uniprot No.	Q8N2G4
Alias	FLJ41033, LYPDC1, MGC29643, PHTS
Product Type	ELISA Kit
Immunogen Species	Homo sapiens (Human)
Sample Types	serum, plasma, tissue homogenates, cell lysates
Detection Range	25 pg/mL-1600 pg/mL
Sensitivity	6.25 pg/mL
Assay Time	1-5h
Sample Volume	50-100ul
Detection Wavelength	450 nm
Lead Time	3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx.
Research Area	Cancer
Gene Names	LYPD1
Tag Info	quantitative
Protein Description	Sandwich
Description	This Human LYPD1 ELISA Kit was designed for the quantitative measurement of Human LYPD1 protein in serum, plasma, tissue homogenates, cell lysates. It is a Sandwich ELISA kit, its detection range is 25 pg/mL-1600 pg/mL and the sensitivity is 6.25 pg/mL.
Product Precision	<p>Intra-assay Precision (Precision within an assay): CV%<8%</p> <p>Three samples of known concentration were tested twenty times on one plate to assess.</p> <p>Inter-assay Precision (Precision between assays): CV%<10%</p> <p>Three samples of known concentration were tested in twenty assays to assess.</p>
Linearity	To assess the linearity of the assay, samples were spiked with high concentrations of human LYPD1 in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.



	Sample	Serum(n=4)
1:1	Average %	87
	Range %	83-91
1:2	Average %	97
	Range %	92-104
1:4	Average %	85
	Range %	82-90
1:8	Average %	93
	Range %	90-96


Recovery

The recovery of human LYPD1 spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

Sample Type	Average % Recovery	Range
Serum (n=5)	94	90-98
EDTA plasma (n=4)	98	92-106

Typical

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



pg/ml	OD1	OD2	Average	Corrected
1600	2.243	2.230	2.237	2.100
800	1.485	1.472	1.479	1.342
400	0.924	0.984	0.954	0.817
200	0.591	0.617	0.604	0.467
100	0.433	0.426	0.430	0.293
50	0.337	0.317	0.327	0.190
25	0.229	0.237	0.233	0.096
0	0.136	0.138	0.137	

Msds

{"0":{"fileurl":"https://www.cusabio.com/uploadfile/msds/MSDS CSB-EL013261HU.pdf","filename":"MSDS"}}