



Human Protein S100-A4(S100A4) ELISA kit

Product Code	CSB-EL020632HU
Abbreviation	S100A4
Target Name	S100 calcium binding protein A4
Uniprot No.	P26447
Alias	18A2, 42A, CAPL, FSP1, MTS1, P9KA, PEL98, OTTHUMP00000015469 OTTHUMP00000032895 S100 calcium binding protein A4 (calcium protein, calvasculin, metastasin, murine placental homolog) S100 calcium-bind
Product Type	ELISA Kit
Immunogen Species	Homo sapiens (Human)
Sample Types	serum, plasma, cell culture supernates, tissue homogenates
Detection Range	0.9 ng/mL-60 ng/mL
Sensitivity	0.225 ng/mL
Assay Time	1-5h
Sample Volume	50-100ul
Detection Wavelength	450 nm
Lead Time	3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx.
Research Area	Signal Transduction
Gene Names	S100A4
Tag Info	quantitative
Protein Description	Sandwich

Description

This Human S100A4 ELISA Kit was designed for the quantitative measurement of Human S100A4 protein in serum, plasma, cell culture supernates, tissue homogenates. It is a Sandwich ELISA kit, its detection range is 0.9 ng/mL-60 ng/mL and the sensitivity is 0.225 ng/mL.

Target Details

This protein is a member of the S100 family of proteins containing 2 EF-hand calcium-binding motifs. S100 proteins are localized in the cytoplasm and/or nucleus of a wide range of cells, and involved in the regulation of a number of cellular processes such as cell cycle progression and differentiation. S100 genes include at least 13 members which are located as a cluster on chromosome 1q21. This protein may function in motility, invasion, and tubulin polymerization. Chromosomal rearrangements and altered expression of this gene have been implicated in tumor metastasis. Multiple alternatively spliced variants, encoding the same protein, have been identified.



Product Precision

Intra-assay Precision (Precision within an assay): CV%<8%

Three samples of known concentration were tested twenty times on one plate to assess.

Inter-assay Precision (Precision between assays): CV%<10%

Three samples of known concentration were tested in twenty assays to assess.

Linearity

To assess the linearity of the assay, samples were spiked with high concentrations of human S100A4 in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.

?	Sample	Serum(n=4)
1:1	Average %	85
	Range %	80-96
1:2	Average %	93
	Range %	82-101
1:4	Average %	97
	Range %	93-104
1:8	Average %	93
	Range %	86-98


Recovery

The recovery of human S100A4 spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

Sample Type	Average % Recovery	Range
Serum (n=5)	98	92-105
EDTA plasma (n=4)	100	95-106

Typical

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



ng/ml	OD1	OD2	Average	Corrected
60	2.654	2.587	2.621	2.471
30	2.216	2.115	2.166	2.016
15	1.369	1.406	1.388	1.238
7.5	0.689	0.715	0.702	0.552
3.75	0.385	0.367	0.376	0.226
1.8	0.288	0.279	0.284	0.134
0.9	0.216	0.206	0.211	0.061
0	0.156	0.144	0.150	?

Msds

```
{
  "0": {
    "fileurl": "https://www.cusabio.com/uploadfile/msds/MSDS CSB-EL020632HU.pdf",
    "filename": "MSDS"
  }
}
```