





Mouse serum amyloid A2 (SAA2) ELISA kit

Product Code	CSB-EL020657MO
Abbreviation	SAA2
Protein Biological Process 1	Immunity
Target Name	serum amyloid A2
Uniprot No.	P05367
Alias	N/A
Product Type	ELISA Kit
Immunogen Species	Mus musculus (Mouse)
Protein Biological Process 3	Acute phase
Sample Types	serum, plasma, tissue homogenates
Detection Range	7.8 ng/mL-500 ng/mL
Sensitivity	1.95 ng/mL
Assay Time	1-5h
Sample Volume	50-100ul
Detection Wavelength	450 nm
Lead Time	3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx.
Research Area	Cardiovascular
Gene Names	Saa2
Tag Info	quantitative
Protein Description	Sandwich
Description	This Mouse SAA2 ELISA Kit was designed for the quantitative measurement of Mouse SAA2 protein in serum, plasma, tissue homogenates. It is a Sandwich ELISA kit, its detection range is 7.8 ng/mL-500 ng/mL and the sensitivity is 1.95 ng/mL.
Product Precision	Intra-assay Precision (Precision within an assay): CV%<8% Three samples of known concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV%<10%

Three samples of known concentration were tested in twenty assays to

assess.







Linearity

To assess the linearity of the assay, samples were spiked with high concentrations of mouse SAA2 in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.

?	Sample	Serum(n=4)
1:1	Average %	108
	Range %	98-112
1:2	Average %	97
	Range %	88-101
1:4	Average %	90
	Range %	83-94
1:8	Average %	100
	Range %	90-104


Recovery

The recovery of mouse SAA2 spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

Sample Type	Average % Recovery	Range
Serum (n=5)	102	93-106
EDTA plasma (n=4)	93	85-97

Typical

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



ng/ml OD1 OD2 Average Corrected

500	1.723 1.835 1.779	1.591	
250	1.277 1.189 1.233	1.045	
125	0.864 0.887 0.876	0.688	
62.5	0.685 0.674 0.680	0.492	
31.2	0.463 0.456 0.460	0.272	
15.6	0.384 0.367 0.376	0.188	
7.8	0.286 0.291 0.289	0.101	
0	0.190 0.186 0.188	?	

Msds

{"0":{"fileurl":"https://www.cusabio.com/uploadfile/msds/MSDS CSB-EL020657MO.pdf","filename":"MSDS"}}