



# Human Diamine acetyltransferase 1(SAT1) ELISA kit

<b>Product Code</b>	CSB-EL020717HU
<b>Abbreviation</b>	SAT1
<b>Target Name</b>	spermidine/spermine N1-acetyltransferase 1
<b>Uniprot No.</b>	P21673
<b>Alias</b>	DC21, KFSD, KFSDX, SAT, SSAT, SSAT-1, diamine N-acetyltransferase 1 diamine acetyltransferase 1 spermidine/spermine N1-acetyltransferase alpha
<b>Product Type</b>	ELISA Kit
<b>Immunogen Species</b>	Homo sapiens (Human)
<b>Sample Types</b>	serum, plasma, tissue homogenates
<b>Detection Range</b>	0.156 ng/mL-10 ng/mL
<b>Sensitivity</b>	0.039 ng/mL
<b>Assay Time</b>	1-5h
<b>Sample Volume</b>	50-100ul
<b>Detection Wavelength</b>	450 nm
<b>Lead Time</b>	3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx.
<b>Research Area</b>	Metabolism
<b>Gene Names</b>	SAT1
<b>Tag Info</b>	quantitative
<b>Protein Description</b>	Sandwich
<b>Description</b>	This Human SAT1 ELISA Kit was designed for the quantitative measurement of Human SAT1 protein in serum, plasma, tissue homogenates. It is a Sandwich ELISA kit, its detection range is 0.156 ng/mL-10 ng/mL and the sensitivity is 0.039 ng/mL.
<b>Product Precision</b>	<p>Intra-assay Precision (Precision within an assay): CV%&lt;8%</p> <p>Three samples of known concentration were tested twenty times on one plate to assess.</p> <p>Inter-assay Precision (Precision between assays): CV%&lt;10%</p> <p>Three samples of known concentration were tested in twenty assays to assess.</p>
<b>Linearity</b>	To assess the linearity of the assay, samples were spiked with high concentrations of human SAT1 in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.



	Sample	Serum(n=4)
1:1	Average %	99
	Range %	93-102
1:2	Average %	99
	Range %	92-102
1:4	Average %	89
	Range %	82-92
1:8	Average %	94
	Range %	90-98


## Recovery

The recovery of human SAT1 spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

Sample Type	Average % Recovery	Range
Serum (n=5)	88	82-92
EDTA plasma (n=4)	100	93-103

## Typical

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



ng/ml	OD1	OD2	Average	Corrected
10	2.560	2.513	2.537	2.408
5	1.623	1.587	1.605	1.476
2.5	1.094	1.068	1.081	0.952
1.25	0.658	0.648	0.653	0.524
0.625	0.437	0.451	0.444	0.315
0.312	0.273	0.266	0.270	0.141
0.156	0.198	0.188	0.193	0.064
0	0.128	0.130	0.129	

## Msds

{"0":{"fileurl":"https://www.cusabio.com/uploadfile/msds/MSDS CSB-EL020717HU.pdf","filename":"MSDS"}}