





Rat C-C motif chemokine 6 (CCL6) ELISA kit

Product Code	CSB-EL027415RA
Abbreviation	CCL6
Target Name	C-C motif chemokine 6 (CCL6)
Uniprot No.	Q68FP3
Alias	Small-inducible cytokine A6
Product Type	ELISA kit
Immunogen Species	Rattus norvegicus (Rat)
Sample Types	serum, plasma, cell culture supernates, tissue homogenates
Detection Range	0.156 ng/mL-10 ng/mL
Sensitivity	0.039 ng/mL
Assay Time	1-5h
Sample Volume	50-100ul
Detection Wavelength	450 nm
Lead Time	3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx.
Research Area	Immunology
Gene Names	Ccl6
Tag Info	quantitative
Protein Description	Sandwich
Description	This Rat CCL6 ELISA Kit was designed for the quantitative measurement of Rat CCL6 protein in serum, plasma, cell culture supernates, tissue homogenates. It is a Sandwich ELISA kit, its detection range is 0.156 ng/mL-10 ng/mL and the sensitivity is 0.039 ng/mL.
Product Precision	Intra-assay Precision (Precision within an assay): CV%<8% Three samples of known concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV%<10% Three samples of known concentration were tested in twenty assays to assess.
Linearity	To assess the linearity of the assay, samples were spiked with high concentrations of rat CCL6 in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay. ? Sample Serum(n=4)

CUSABIO TECHNOLOGY LLC



Tel: +1-301-363-4651	
1011 - 1 001 000 1001	





1:1	Average %	94
1.1	Range %	87-98
1:2	Average %	98
1.2	Range %	91-104
1:4	Average %	91
1.4	Range %	85-97
1:8	Average %	97
	Range %	92-103


Recovery

The recovery of rat CCL6 spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

Sample Type	Average % Recovery	Range
Serum (n=5)	92	87-96
EDTA plasma (n=4)	98	93-104

Typical

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



ng/ml OD1 OD2 Average Corrected

10	2.634	2.782	2.708	2.604
5	2.178	2.358	2.268	2.164
2.5	1.572	1.556	1.564	1.460
1.25	0.852	0.860	0.856	0.752
0.625	0.413	0.426	0.420	0.316
0.312	0.317	0.301	0.309	0.205
0.156	0.204	0.200	0.202	0.098
0	0.103	0.105	0.104	?

Msds

{"0":{"fileurl":"https://www.cusabio.com/uploadfile/msds/MSDS CSB-EL027415RA.pdf", "filename": "MSDS"}}