



**Sino Biological Inc.**  
Biological Solution Specialist

---

# **Immunoprecipitation/IP Kit- Anti-HA Immunomagnetic Beads**

**Catalog Number: MB11068-T60**

Please read this instruction manual carefully before using the product

## Product Contents

| Contents                                    | Package 1                                    | Package 2                                  | Storage             |
|---|--|--|---------------------|
| Anti-HA Immunomagnetic Beads <sup>1 3</sup> | 1 mL   | 5 mL                                       | 2-8°C for 12 months |
| NP40 Cell Lysis Buffer                      | 4 mL   | 22 mL                                      | -20°C for 12 months |
| 5×TBST (pH7.4)                              | Required but not supplied                    |  |                     |
| 1×TBST (pH7.4)                              | Required but not supplied                    |  |                     |
| ddH <sub>2</sub> O                          | Required but not supplied                    |  |                     |
| Alkaline Elution Buffer                     | 3 mL   | 15 mL                                      | 2-8°C for 12 months |
| Acidity Elution Buffer                      | 3 mL   | 15 mL                                      | 2-8°C for 12 months |
| Neutralization Buffer                       | 2 mL   | 8 mL                                       | 2-8°C for 12 months |
| Magnetic Separator                          | Not included (refer related product MAGS001) | One MAGS001 included in China <sup>2</sup> |                     |

[1] The IP KIT contains anti-HA Immunomagnetic Beads(2 mg/mL) in phosphate buffered saline (PBS, pH 7.4) with sodium azide (0.1%).

[2] The Magnetic Separator cannot be included for oversea customers due to shipment prohibition.

[3] Immunomagnetic Beads kits are shipped at ambient temperature in which immunomagnetic beads are provided in liquid buffer.

## Product Description

The Anti-HA Immunomagnetic Beads, conjugated with Anti-HA antibody, are used for immunoprecipitation (IP) of HA proteins which expressed in vitro expression systems and bacterial and mammalian cell lysates.

For IP, the beads are added to a sample containing HA proteins to form a bead-protein complex. The complex is removed from the solution manually using a Magnetic Separator. The bound HA proteins are dissociated from the Immunomagnetic Beads using an elution buffer.

## Antibody Information

**Antibody:** Influenza A H1N1 (A/Brevig Mission/1/1918) Hemagglutinin / HA Antibody, Rabbit PAb, Antigen Affinity Purified(11068-T60)

**Immunogen:** Recombinant Influenza A H1N1 (A/Brevig Mission/1/1918) Hemagglutinin / HA Protein (Catalog#11068-V08H)

**Clone ID:**

**Isotype:** Rabbit IgG

**Specificity:** Influenza A H1N1 (A/Brevig Mission/1/1918) Hemagglutinin / HA

**Guaranteed Applications:** IP, Minimum Protein Purification

**Preparation:** Produced in rabbits immunized with purified, recombinant Influenza A H1N1 (A/Brevig Mission/1/1918) Hemagglutinin / HA (Catalog#11068-V08H; AAD17229.1; Met1-Gln529). Influenza A H1N1 (A/Brevig Mission/1/1918) Hemagglutinin / HA specific IgG was purified by Influenza A H1N1 (A/Brevig Mission/1/1918) Hemagglutinin / HA affinity chromatography.




Fig. 1 Immunoprecipitation (IP) Protocol

## Protocol

The protocol (Fig. 1) uses 50  $\mu\text{L}$  Anti-HAImmunomagnetic Beads, but this can be scaled up or down as required.

### Cell Lysis

Cells may be lysed using any standard cell lysis protocol in accordance with your starting materials. We suggest using NP40 Cell Lysis Buffer (supplied with kit).

### Immunoprecipitate Target Antigen

1. Add 50  $\mu\text{L}$  of Immunomagnetic Beads into a 1.5 mL microcentrifuge tube.
2. Add 150  $\mu\text{L}$  of 1 $\times$  TBST buffer to the Immunomagnetic Beads and gently vortex to mix.
3. Place the tube into a Magnetic Separator to collect the beads against the wall side of the tube. Remove and discard the supernatant.
4. Add 1 mL of 1 $\times$  TBST buffer to the tube. Invert the tube several times or gently vortex to mix for 1 min. Collect Immunomagnetic Beads with a Magnetic Separator. Remove and discard the supernatant.
5. Add the sample containing target protein (~100  $\mu\text{g}$  of protein in 100  $\mu\text{L}$ ) to the pre-washed Immunomagnetic Beads, add 400  $\mu\text{L}$  of 1 $\times$  TBST buffer and incubate at room temperature for 30 min with mixing.
6. Collect the Immunomagnetic Beads with a Magnetic Separator, remove the unbounded sample and save for analysis.
7. Add 300  $\mu\text{L}$  of 5 $\times$  TBST buffer to the tube and gently mix. Collect the Immunomagnetic Beads and discard the supernatant. Repeat this wash twice.
8. Add 300  $\mu\text{L}$  of ddH<sub>2</sub>O to the tube and gently mix. Collect the Immunomagnetic Beads on a Magnetic Separator and discard the supernatant.

### Elute Target Antigen.

#### A. Alkaline Elution Protocols

1. Add 100  $\mu\text{L}$  of Alkaline Elution buffer to the tube.
2. Gently vortex to mix and incubate the sample at room temperature on a rotator for 5 min.
3. Magnetically separate the Immunomagnetic Beads and save the supernatant containing the target antigen.
4. To neutralize the sample, add 50  $\mu\text{L}$  of Neutralization Buffer for each 100  $\mu\text{L}$  of eluate.

#### B. Acidity Elution

1. Add 100  $\mu\text{L}$  Acidity Elution Buffer.
2. Gently vortex to mix and incubate the sample at room temperature on a rotator for 5-10 min.
3. Magnetically separate the Immunomagnetic Beads and save the supernatant containing the target antigen.

4. To neutralize the low pH, add 15  $\mu\text{L}$  of Neutralization Buffer for each 100  $\mu\text{L}$  of eluate.

#### C. Elution Using Sample Buffer

1. Add 100  $\mu\text{L}$  of SDS-PAGE sample buffer to the tube.
2. Gently vortex to mix and incubate the sample at 95-100°C for 5-10 min.
3. Magnetically separate the Immunomagnetic Beads and save the supernatant containing the antigen.

## Reference Information

### Related Products

| Products   | Cat No.  |
|--|----------|
| Magnetic Separator-1.5 (2 tubes)   | MAGS001  |
| Immunoprecipitation Kit<br>-Immunomagnetic Beads Protein A Kit                   | BA10600  |
| Immunoprecipitation Kit<br>-Immunomagnetic Beads Protein G Kit                   | BG13103  |
| Immunoprecipitation Kit<br>-Immunomagnetic Beads Protein L Kit                   | BL11044  |
| Immunoprecipitation Kit<br>-Anti-DYKDDDDK(Flag®) Tag<br>Immunomagnetic Beads Kit | TB101274 |
| Immunoprecipitation Kit<br>-Anti-GFP Tag Immunomagnetic Beads Kit                | TB13105  |
| Immunoprecipitation Kit<br>-Anti-Myc Tag Immunomagnetic Beads Kit                | TB100029 |
| Immunoprecipitation Kit<br>-Anti-HA Tag Immunomagnetic Beads Kit                 | TB100028 |
| Immunoprecipitation Kit<br>-Anti-V5 Tag Immunomagnetic Beads Kit                 | TB100378 |

### Trouble Shooting

| Problem                          | Possible Cause                      | Solution  |
|----------------------------------|-------------------------------------|---|
| Little or no protein is detected | Protein degraded                    | Include protease inhibitors (PMSF) in the lysis buffer          |
|                                  |                                     | Use new lysate or lysate stored at -80° C                       |
|                                  | No or minimal protein was expressed | Verify protein expression by SDS-PAGE or Western blot           |
|                                  |                                     | Analysis of the lysate using an positive control as a reference |
|                                  |                                     | Increase the amount of lysate used for IP/Co-IP                 |
|                                  |                                     | Use a more sensitive detection system                           |

| Problem  | Possible Cause   | Solution  |
|--|--|---|
| Magnetic Beads aggregated  | Magnetic Beads were frozen or centrifuged                                | Handle the Beads as directed in the instructions        |
|  | Buffer was incompatible with magnetic beads                              |   |
|  | Detergent was not added to the wash and bind solutions                   |   |
| Failure to co-IP interacting protein   | Wash conditions were too stringent for the weak or transient interaction | Reduce the number of washes                             |
|  |  | Lower the ionic strength of the wash buffer             |
|  | Interacting protein was expressed at a low level                         | Apply additional protein sample                         |
|  |  | Use a more sensitive detection system                   |
|  | Buffer system was not optimal for the protein: protein interaction       | Optimize the co-IP buffer                               |
|  | Insufficient sample was loaded on the gel for Western blot detection     | Elute sample in 30% acetonitrile 0.5% formic acid, then |
| Bring the sample back up in SDS-PAGE Sample Buffer and load entire elution fraction on |  |   |