

# Human CD208 / DC-LAMP / LAMP3 ELISA Pair Set

Catalog Number: SEK10527

To achieve the best assay results, this manual must be read carefully before using this product and the assay is run as summarized in the General ELISA protocol.

#### BACKGROUND

Dendritic cell-lysosomal associated membrane protein (DC-LAMP)/CD208, also known as LAMP3, is a member of the lysosomal associated membrane protein (LAMP) family, which is specifically expressed by human dendritic cells (DCs) upon activation and therefore serves as marker of human DC maturation. Confocal and immunoelectron microscopy showed that mouse DC-LAMP protein co-localizes with lbm180, a specific marker for the limiting membrane of lamellar bodies that contain surfactant protein B. The present study demonstrates that DC-LAMP is constitutively expressed by mouse, sheep, and human type II pneumocytes. DC-LAMP is constitutively expressed in normal type II pneumocytes. DC-LAMP is detected first in activated human DC within MHC class II molecules-containing compartments just before the translocation of MHC class II-peptide complexes to the cell surface, suggesting a possible involvement in this process. Furthermore, overexpression of LAMP3 is actively involved in tumor invasion through increased migration into lymph-vascular spaces.


#### PRINCIPLE OF THE TEST

The Sino Biological ELISA Pair Set is a solid phase sandwich ELISA (Enzyme-Linked Immunosorbent Assay). It utilizes a monoclonal antibody specific for CD208 / DC-LAMP / LAMP3 coated on a 96-well plate. Standards and samples are added to the wells, and any CD208 / DC-LAMP / LAMP3 present binds to the immobilized antibody. The wells are washed and a horseradish peroxidase conjugated rabbit anti-CD208 / DC-LAMP / LAMP3 monoclonal antibody is then added, producing an antibody-antigenantibody "sandwich". The wells are again washed and TMB substrate solution is loaded, which produces color in proportion to the amount of CD208 / DC-LAMP / LAMP3 present in the sample. To end the enzyme reaction, the stop solution is added and absorbances of the microwell are read at 450 nm.

#### INTENDED USE

- ◆The Human CD208 / DC-LAMP / LAMP3 ELISA Pair Set is for the quantitative determination of Human CD208 / DC-LAMP / LAMP3.
- ◆This ELISA Pair Set contains the basic components required for the development of sandwich ELISAs.

#### ASSAY PROCEDURE SUMMARY



This Pair Set has been configured for research use only and is not to be used in diagnostic procedures.

# **MATERIALS PROVIDED**

Bring all reagents to room temperature before use.

**Capture Antibody** – 1 mg/mL of rabbit anti-LAMP3 monoclonal antibody, Dilute to a working concentration of 2.0 µg/mL in CBS before coating. (Catalog: # 10527-R020)

**Detection Antibody** - 0.5 mg/mL rabbit anti-LAMP3 monoclonal antibody conjugated to horseradish-peroxidase (HRP). Dilute to working concentration of 0.1 µg/mL in detection antibody dilution buffer before use. (Catalog: # 10527-R002)

**Standard** – Each vial contains 50 ng of recombinant LAMP3. Reconstitute standard powder with 1 mL detection antibody dilution buffer. After reconstitution, store at -20°C to -80°C in a manual defrost freezer. A seven-point standard curve usi ng 2-fold serial dilutions in sample dilution buffer, and a high standard of 2000 pg/mL is recommended.

Standard reconstitution tips: Add dilution buffer, gently mix it up and down 3~5 times. Avoid violent and long-time shock.

#### SOLUTIONS REQUIRED

CBS - 0.05M Na<sub>2</sub>CO<sub>3</sub>, 0.05M NaHCO<sub>3</sub>, pH 9.6,  $0.2 \mu m$  filtered

TBS - 20 mM Tris, 150 mM NaCl, pH 7.4

Wash Buffer - 0.05% Tween20 in TBS, pH 7.2 - 7.4

Blocking Buffer - 2% BSA in Wash Buffer

Sample dilution buffer - 0.1% BSA in wash buffer, pH 7.2 - 7.4, 0.2 µm filtered

Detection antibody dilution buffer - 0.5% BSA in wash buffer, pH 7.2 - 7.4, 0.2 µm filtered

Substrate Solution: To achieve best assay results, fresh substrate solution is recommended

Substrate stock solution - 10mg / ml TMB ( Tetramethylbenzidine ) in DMSO

Substrate dilution buffer - 0.05M Na<sub>2</sub>HPO<sub>4</sub> and 0.025M citric acid; adjust pH to 5.5

Substrate working solution - For each plate dilute 250  $\mu$ l substrate stock solution in 25ml substrate dilution buffer and then add 80  $\mu$ l 0.75%  $H_2O_2$ , mix it well

Stop Solution - 2 N H<sub>2</sub>SO<sub>4</sub>

## **PRECAUTION**

The Stop Solution suggested for use with this Pair Set is an acid solution. Wear eye, hand, face, and clothing protection when using this material.

#### STORAGE

**Capture Antibody**: Aliquot and store at -20°C to -80°C for up to 6 months from date of receipt. Avoid repeated freeze-thaw cycles.

**Detection Antibody**: Protect it from prolonged exposure to light. Aliquot and store at -20°C to -80°C and for up to 6 months from date of receipt. Avoid repeated freeze-thaw cycles.

**Standard**: Store lyophilized standard at -20°C to -80°C for up to 6 months from date of receipt. Aliquot and store the reconstituted standard at -80°C for up to 1 month. Avoid repeated freeze-thaw cycles.

#### GENERAL ELISA PROTOCOL

#### **Plate Preparation**

- 1.Dilute the capture antibody to the working concentration in CBS. Immediately coat a 96-well microplate with 100µL per well of the diluted capture antibody. Seal the plate and incubate overnight at 4°C.
- 2.Aspirate each well and wash with at least  $300\mu\text{I}$  wash buffer, repeating the process two times for a total of three washes. Complete removal of liquid at each step is essential for good performance. After the last wash, remove any remaining wash buffer by inverting the plate and blotting it against clean paper towels. 3.Block plates by adding  $300~\mu\text{L}$  of blocking buffer to each well. Incubate at room temperature for a minimum of 1 hour.
- 4.Repeat the aspiration/wash as in step 2. The plates are now ready for sample addition.

#### **Assay Procedure**


- $1.Add\ 100\ \mu L$  of sample or standards in sample dilution buffer per well. Seal the plate and incubate 2 hours at room temperature.
- 2. Repeat the aspiration/wash as in step 2 of plate preparation.
- $3.Add\ 100\ \mu L$  of the detection antibody, diluted in antibody dilution buffer, to each well. Seal the plate and incubate 1 hour at room temperature.
- 4. Repeat the aspiration/wash as in step 2 of plate preparation.
- $5.Add\ 200\ \mu L$  of substrate solution to each well. Incubate for 20 minutes at room temperature ( if substrate solution is not as requested, the incubation time should be optimized ). Avoid placing the plate in direct light.
- 6.Add 50 µL of stop solution to each well. Gently tap the plate to ensure thorough mixing.
- 7.Determine the optical density of each well immediately, using a microplate reader set to 450 nm.

#### **CALCULATION OF RESULTS**

- •Calculate the mean absorbance for each set of duplicate standards, controls and samples. Subtract the mean zero standard absorbance from each.
- •Construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph.
- •To determine the concentration of the unknowns, find the unknowns' mean absorbance value on the y-axis and draw a horizontal line to the standard curve. At the point of intersection, draw a vertical line to the x-axis and read the concentration. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.
- •Alternatively, computer-based curve-fitting statistical software may also be employed to calculate the concentration of the sample.

#### TYPICAL DATA

This standard curve is only for demonstration purposes. A standard curve should be generated for each assay.



Concentration (pg/ml)	Zero standard subtracted OD
0	0.000
31.25	0.034
62.5	0.069
125	0.164
250	0.369
500	0.779
1000	1.308
2000	2.827

## PERFORMANCE CHARACTERISTIC

#### **SENSITIVITY**

The minimum detectable dose of Human CD208 / DC-LAMP / LAMP3 was determined to be approximately 31.25 pg/ml. This is defined as at least three times standard deviations above the mean optical density of 10 replicates of the zero standard.

# TROUBLE SHOOTING

Problems	Possible Sources	Solutions
No signal	Incorrect or no Detection Antibody was added	Add appropriate Detection Antibody and continue
	Substrate solution was not added	Add substrate solution and continue
	Incorrect storage condition	Check if the kit is stored at recommended condition and used before expiration date
Poor Standard Curve	Standard was incompletely reconstituted or was inappropriately stored	Aliquot reconstituted standard and store at -80 °C
	Imprecise / inaccurate pipetting	Check / calibrate pipettes
	Incubations done at inappropriate temperature, timing or agitation	Follow the general ELISA protocol
	Background wells were contaminated	Avoid cross contamination by using the sealer appropriately
Poor detection value	The concentration of antigen in samples was too low	Enriching samples to increase the concentration of antigen
	Samples were ineffective	Check if the samples are stored at cold environment. Detect samples in timely manner
High Background	Insufficient washes	Use multichannel pipettes without touching the reagents on the plate
		Increase cycles of washes and soaking time between washes
	TMB Substrate Solution was contaminated	TMB Substrate Solution should be clear and colorless prior to addition to wells
	Materials were contaminated.	Use clean plates, tubes and pipettes tips
Non-specificity	Samples were contaminated	Avoid cross contamination of samples
	The concentration of samples was too high	Try higher dilution rate of samples

U

ഥ

I

ш

Fax:+86-10-58628220

# Human CD208 / DC-LAMP / LAMP3 ELISA Pair Set Notes